

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
 Data File : BN017291.D
 Acq On : 04 Nov 2021 00:58
 Operator : CG/JU
 Sample : M4364-05
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG336

Quant Time: Nov 08 01:07:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 03 10:42:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.514	152	5390	0.400	ng/ul	0.00
4) Naphthalene-d8	10.284	136	21851	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.166	164	13890	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.924	188	29539	0.400	ng/ul	0.00
17) Chrysene-d12	21.133	240	26304	0.400	ng/ul	0.00
23) Perylene-d12	23.331	264	28796	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.038	96	26622	4.442	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.890	152	11400	0.357	ng/ul	0.00
18) Fluoranthene-d10	18.962	212	35625	0.517	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
Data File : BN017291.D
Acq On : 04 Nov 2021 00:58
Operator : CG/JU
Sample : M4364-05
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG336

Quant Time: Nov 08 01:07:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 03 10:42:03 2021
Response via : Initial Calibration

