


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
 Data File : BN017266.D
 Acq On : 03 Nov 2021 09:10
 Operator : CG/JU
 Sample : DFTPP213
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP213

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 08 11:43:44 2021



AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2441

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	209728	PASS
68	69	0.00	2	1.8	3589	PASS
69	198	0.00	100	34.5	197739	PASS
70	69	0.00	2	0.6	1153	PASS
127	198	10	80	45.2	258688	PASS
197	198	0.00	2	0.4	2434	PASS
198	198	100	100	100.0	572565	PASS
199	198	5	9	6.8	38752	PASS
275	198	10	60	26.7	152965	PASS
365	198	1	100	3.3	18696	PASS
441	198	0.01	100	14.9	85128	PASS
442	198	50	100	90.1	515627	PASS
443	442	15	24	21.9	112995	PASS

m/z	Abundance
40.00	995.0
43.90	1436.0
48.90	610.0
73.00	591.0
84.00	406.0
207.00	1952.0
208.00	339.0

Instrument :

BNA_N

ClientSampleId :

DFTPP213

m/z	Abundance
36.00	314.0
37.00	1007.0
38.00	2151.0
39.00	14519.0
40.00	1768.0
41.10	569.0
44.00	1385.0
44.90	388.0
49.00	1981.0
50.00	50456.0
51.00	194560.0
52.00	10902.0
53.00	417.0
54.90	1472.0
56.00	5875.0
57.00	13389.0
58.00	504.0
61.00	2644.0
62.00	2782.0
63.00	7314.0
64.00	1165.0
65.00	3867.0
66.00	397.0
67.00	346.0
68.00	3388.0
69.00	182656.0
70.00	1035.0
73.00	2051.0
74.00	19392.0
75.00	31072.0
76.00	10677.0
77.00	212544.0
78.00	14779.0
79.00	14277.0
80.00	10078.0
81.00	14759.0
82.00	3485.0
83.00	3820.0
84.00	772.0
85.00	4094.0
86.00	4714.0
86.90	1907.0
88.00	950.0
89.00	345.0
91.00	3290.0
92.00	3843.0
93.00	23128.0
94.00	1552.0
95.10	338.0
96.00	1207.0
97.00	537.0
98.00	18192.0
99.00	16640.0
100.00	1652.0
101.00	7666.0
101.90	463.0
103.00	2703.0
104.00	4951.0
105.00	4902.0
106.00	1703.0
107.00	63864.0
108.00	9776.0
109.00	2129.0
110.00	119416.0
111.00	18032.0
112.00	1925.0
112.90	745.0
115.00	434.0
116.00	3513.0
117.00	53032.0
118.00	4507.0
119.00	598.0
119.90	835.0
120.90	394.0
122.00	4115.0
123.00	6876.0
124.00	2673.0
125.00	2681.0
127.00	230272.0
128.00	17624.0
129.00	83360.0
130.00	7270.0
131.00	1507.0
131.90	777.0
133.00	426.0
134.00	2328.0
135.00	7246.0
136.00	2668.0
137.00	4001.0
137.80	767.0
138.90	514.0
140.00	1015.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

141.00	10806.0
142.00	3703.0
143.00	2644.0
143.90	737.0
145.00	633.0
146.00	1815.0
147.00	6222.0
147.90	12304.0
149.00	3211.0
150.00	957.0
150.90	1490.0
151.90	888.0
153.00	3801.0
154.00	2468.0
155.00	6433.0
156.00	8866.0
157.00	1613.0
158.00	1999.0
159.00	1787.0
160.00	2918.0
161.00	4869.0
162.00	1525.0
163.10	435.0
164.10	610.0
165.00	4224.0
166.00	3173.0
167.00	22632.0
168.00	17096.0
169.00	2293.0
170.00	793.0
171.00	775.0
171.90	1854.0
173.00	2456.0
174.00	4090.0
175.00	8027.0
176.00	2808.0
177.00	3800.0
178.10	1222.0
179.00	15872.0
180.00	12149.0
181.00	5391.0
182.00	802.0
182.90	445.0
184.00	1481.0
185.00	7603.0
186.00	61520.0
187.00	17192.0
188.00	2424.0
189.00	3647.0
190.00	864.0
191.00	2130.0
192.00	5179.0
193.00	5423.0
193.90	1509.0
195.00	682.0
196.00	13116.0
198.00	458624.0
199.00	30992.0
200.00	2193.0
201.50	2279.0
203.00	3080.0
204.00	14430.0
205.00	25656.0
206.00	114200.0
207.00	16400.0
208.00	3817.0
209.00	1343.0
210.10	1680.0
211.00	4639.0
211.90	1009.0
213.00	384.0
215.00	1507.0
216.00	2112.0
217.00	28888.0
218.00	3536.0
218.90	345.0
221.00	19472.0
223.00	6455.0
224.00	61024.0
225.00	15978.0
226.00	2256.0
227.00	24888.0
228.00	3301.0
229.00	5495.0
230.00	708.0
231.00	2242.0
231.90	510.0
232.90	425.0
234.00	1569.0
235.00	1876.0
236.00	1249.0
237.00	3125.0
237.90	408.0
239.00	1059.0
240.00	862.0
241.00	1244.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

242.00	3044.0
243.00	3782.0
244.00	50464.0
245.00	7529.0
246.00	8524.0
247.00	2005.0
248.00	458.0
249.00	1674.0
249.90	309.0
250.80	326.0
251.90	495.0
253.00	951.0
255.00	246848.0
256.00	37728.0
257.00	3474.0
258.00	14257.0
259.00	2502.0
260.00	409.0
261.00	400.0
264.00	325.0
265.00	5501.0
265.90	714.0
271.00	578.0
272.00	668.0
273.00	6989.0
274.00	19664.0
275.00	106768.0
276.00	15097.0
277.00	7536.0
278.00	2048.0
279.00	308.0
281.10	385.0
282.90	888.0
283.90	464.0
285.00	1699.0
286.00	327.0
288.90	358.0
292.00	510.0
293.00	2062.0
294.00	674.0
295.00	722.0
296.00	28576.0
297.00	3909.0
297.90	313.0
300.90	413.0
302.00	431.0
303.00	3088.0
304.00	1066.0
308.00	319.0
310.10	308.0
314.00	1213.0
315.00	2966.0
316.00	1705.0
317.10	465.0
321.00	958.0
322.10	532.0
323.00	9057.0
324.00	1784.0
326.90	1924.0
328.00	1066.0
331.90	714.0
333.00	865.0
334.00	5275.0
335.00	1602.0
341.00	1102.0
341.90	343.0
345.90	1736.0
346.90	475.0
352.00	2919.0
353.00	2103.0
354.00	2715.0
355.10	735.0
364.90	11943.0
366.00	1638.0
370.90	711.0
372.00	5130.0
373.00	1290.0
383.00	1014.0
384.00	492.0
390.00	612.0
391.00	444.0
392.00	303.0
400.90	334.0
402.00	1794.0
402.90	2433.0
404.00	1117.0
421.00	2349.0
422.00	2710.0
423.00	17264.0
424.00	6692.0
424.90	969.0
441.00	44952.0
442.00	307008.0
443.00	62000.0
444.00	6979.0
445.00	387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

m/z	Abundance
36.00	375.0
37.00	1119.0
38.00	3021.0
39.00	17080.0
40.00	1621.0
41.00	664.0
43.10	329.0
44.00	1193.0
45.00	383.0
49.10	2169.0
50.00	63352.0
51.00	251584.0
52.00	12654.0
53.00	497.0
55.00	1389.0
56.00	7487.0
57.00	16150.0
58.00	639.0
61.00	3325.0
62.00	3312.0
63.00	9710.0
64.00	1137.0
65.00	5170.0
66.00	469.0
66.90	376.0
68.00	4103.0
69.00	237696.0
70.00	1393.0
73.00	2017.0
74.00	24896.0
75.00	38096.0
76.10	14198.0
77.00	285696.0
78.10	20416.0
79.00	18888.0
80.00	14012.0
81.00	19312.0
82.00	4685.0
83.00	4403.0
83.90	1013.0
85.00	5546.0
86.00	5732.0
87.00	2539.0
88.00	1371.0
89.00	502.0
91.00	4050.0
92.00	4775.0
93.00	31696.0
94.00	2236.0
94.90	596.0
96.00	1600.0
96.90	550.0
98.00	23544.0
99.00	21208.0
100.00	1745.0
101.00	10341.0
102.00	503.0
103.00	3941.0
104.00	7325.0
105.00	6163.0
106.00	2619.0
107.00	84672.0
108.00	12848.0
109.10	2540.0
110.00	162624.0
111.00	24264.0
112.00	2887.0
113.00	1118.0
114.90	426.0
116.00	4688.0
117.00	71616.0
118.00	5503.0
119.00	753.0
120.00	1199.0
120.90	395.0
122.00	5516.0
123.00	8687.0
123.90	3781.0
125.00	3601.0
126.20	1850.0
127.00	307584.0
128.00	23224.0
129.00	114064.0
130.00	9979.0
131.00	2201.0
132.00	1039.0
133.00	629.0
134.00	3551.0
135.00	10174.0
136.00	3837.0
137.00	6110.0
137.80	1048.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

139.00	662.0
140.00	1670.0
141.00	15172.0
142.00	4835.0
143.00	3580.0
144.00	1012.0
145.00	674.0
146.00	2621.0
147.00	7257.0
148.00	18744.0
149.00	4500.0
150.10	1206.0
151.10	2107.0
151.90	1124.0
153.00	4662.0
154.00	3354.0
155.00	8723.0
156.00	13369.0
157.10	3031.0
157.90	2704.0
159.00	2297.0
160.00	4932.0
161.00	7572.0
162.00	2081.0
162.90	575.0
164.10	928.0
165.00	5679.0
166.00	4651.0
167.00	32600.0
168.00	20240.0
169.00	2970.0
169.90	976.0
170.90	1402.0
172.00	3006.0
173.00	3438.0
174.00	6513.0
175.00	12882.0
176.00	3999.0
177.00	5759.0
178.00	1960.0
179.00	23216.0
180.00	17144.0
181.00	7825.0
182.10	1454.0
182.90	652.0
184.00	1707.0
185.00	10927.0
186.00	86992.0
187.00	26648.0
188.00	2926.0
189.00	5195.0
190.00	891.0
191.00	2661.0
192.00	7672.0
193.00	8651.0
194.00	1746.0
195.00	1109.0
196.00	20344.0
197.10	3673.0
198.00	691776.0
199.00	45552.0
200.00	3281.0
201.50	2992.0
203.00	4061.0
204.00	24352.0
205.00	40184.0
206.00	171968.0
207.00	25512.0
208.00	5165.0
209.00	2069.0
211.00	7005.0
212.90	371.0
215.00	1913.0
216.00	3881.0
217.00	41472.0
218.00	5782.0
219.00	508.0
221.00	29880.0
223.00	9906.0
224.00	94944.0
225.00	25288.0
226.00	2672.0
227.00	36296.0
228.00	5623.0
229.00	8242.0
230.00	1044.0
231.00	3535.0
232.10	576.0
233.00	641.0
234.00	2506.0
234.90	2567.0
236.00	1824.0
237.00	3797.0
237.90	590.0
239.00	1522.0
240.00	1499.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

241.00	2120.0
242.00	5095.0
243.00	5622.0
244.00	79760.0
245.00	10852.0
246.00	14398.0
247.00	2981.0
247.90	799.0
249.00	2699.0
250.00	610.0
250.90	588.0
252.00	745.0
253.00	1442.0
255.00	393600.0
256.00	56224.0
257.00	4651.0
258.00	21520.0
259.00	3860.0
259.90	656.0
261.00	797.0
264.00	505.0
265.00	8593.0
266.00	1188.0
269.90	368.0
271.00	814.0
272.00	860.0
273.00	12320.0
274.00	31008.0
275.00	182400.0
276.00	25320.0
277.00	13787.0
278.00	2508.0
279.00	686.0
280.90	311.0
282.10	569.0
283.00	1397.0
284.10	918.0
285.00	2855.0
286.00	569.0
288.90	638.0
289.90	462.0
290.90	416.0
292.00	613.0
293.00	3337.0
294.00	868.0
295.10	1314.0
296.00	42160.0
297.00	7189.0
297.90	498.0
300.90	690.0
302.00	995.0
303.00	5588.0
304.00	1453.0
305.00	305.0
308.00	682.0
308.90	430.0
309.90	643.0
313.00	402.0
314.00	2366.0
314.90	5012.0
316.00	2813.0
317.00	478.0
321.00	1443.0
322.00	803.0
323.00	15152.0
324.10	2809.0
326.00	335.0
326.90	2975.0
328.00	1370.0
328.90	405.0
332.00	1188.0
333.00	1416.0
334.00	10390.0
334.90	2921.0
339.00	314.0
341.00	1978.0
341.90	729.0
346.00	3266.0
347.00	626.0
352.00	5478.0
353.00	3552.0
354.00	5336.0
355.00	1331.0
359.00	431.0
365.00	20632.0
365.90	3238.0
369.90	516.0
371.10	1405.0
372.00	9363.0
373.10	2249.0
383.00	2530.0
384.10	621.0
389.90	1253.0
391.00	771.0
392.00	830.0
400.90	508.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

402.00	3693.0
403.00	4793.0
404.00	2111.0
405.20	458.0
421.00	4646.0
422.00	4822.0
423.10	33120.0
424.00	9700.0
425.00	1334.0
441.00	99456.0
442.10	586112.0
443.00	129784.0
444.00	12313.0
444.90	637.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

m/z	Abundance
36.00	327.0
37.00	744.0
38.00	2002.0
39.10	12729.0
40.00	1581.0
41.00	436.0
43.00	444.0
43.90	1038.0
44.90	384.0
49.00	2008.0
50.00	47096.0
51.00	183040.0
52.00	9282.0
53.00	325.0
55.00	1411.0
56.00	4935.0
57.00	12276.0
57.90	593.0
61.00	2021.0
62.00	2656.0
63.00	6942.0
63.90	1021.0
65.00	3541.0
67.10	379.0
68.00	3275.0
69.00	172864.0
70.00	1031.0
73.00	1767.0
74.00	18528.0
75.00	27168.0
76.10	10184.0
77.00	210432.0
78.00	14119.0
79.00	12498.0
80.00	9894.0
81.00	14187.0
82.00	3376.0
83.00	3171.0
83.90	666.0
85.00	4064.0
86.00	4257.0
87.00	2246.0
88.00	922.0
89.00	395.0
91.00	3136.0
92.00	3822.0
93.00	23872.0
94.00	1673.0
95.00	554.0
96.00	1322.0
97.00	651.0
98.00	19032.0
99.00	15912.0
99.90	1516.0
101.00	8130.0
101.90	471.0
103.00	2792.0
104.00	5107.0
105.00	4957.0
106.10	2025.0
107.00	65000.0
108.00	10708.0
109.10	1786.0
110.00	124992.0
111.00	18256.0
112.00	2071.0
112.90	763.0
114.80	318.0
116.00	3704.0
117.00	58016.0
118.00	4357.0
119.00	737.0
120.00	933.0
121.00	433.0
122.00	4474.0
123.00	7329.0
124.00	3198.0
125.00	2978.0
126.20	1586.0
127.00	238208.0
128.00	18328.0
129.00	92504.0
130.00	7555.0
131.00	1291.0
132.10	719.0
134.00	2873.0
135.00	7495.0
136.00	3286.0
137.00	4365.0
137.90	1041.0
138.90	603.0
140.00	1096.0

Instrument :

BNA_N

ClientSampleId :

DFTPP213

141.00	11729.0
142.00	4146.0
143.00	2850.0
144.00	742.0
145.00	773.0
145.90	1925.0
147.00	6598.0
148.00	15366.0
149.00	3181.0
150.00	1000.0
151.00	1664.0
152.10	867.0
153.00	3919.0
154.00	3091.0
155.00	7145.0
156.00	10772.0
157.10	1972.0
157.90	2097.0
158.90	2066.0
160.00	3631.0
161.00	5656.0
162.00	1493.0
163.00	437.0
164.00	540.0
165.00	4865.0
166.00	3993.0
167.00	26488.0
168.00	16157.0
169.00	2400.0
170.00	905.0
170.90	1171.0
172.00	2200.0
173.00	2970.0
174.00	4817.0
175.00	10129.0
176.00	2921.0
177.00	4668.0
178.00	1432.0
179.00	19840.0
180.00	14326.0
181.00	5858.0
182.10	922.0
182.90	654.0
184.00	1698.0
185.00	10199.0
186.00	78280.0
187.00	21608.0
188.00	2525.0
189.00	4624.0
189.90	699.0
191.00	2530.0
191.90	6028.0
193.00	7485.0
194.00	1756.0
195.00	1224.0
196.00	17824.0
196.90	3630.0
198.00	567296.0
199.00	39712.0
200.00	2899.0
201.40	2503.0
203.00	3860.0
204.00	20640.0
205.00	34416.0
206.10	146176.0
207.00	21168.0
208.00	5086.0
209.00	1753.0
210.10	2482.0
211.00	5763.0
212.00	1129.0
212.90	423.0
215.00	1663.0
216.00	2886.0
217.00	38816.0
218.00	4724.0
218.90	419.0
221.00	27448.0
223.00	8604.0
224.00	82624.0
225.00	20976.0
226.10	2599.0
227.00	31584.0
228.00	4959.0
229.00	6886.0
230.00	949.0
231.00	2939.0
232.10	693.0
232.90	546.0
234.00	2431.0
235.00	2450.0
236.00	1506.0
237.00	3283.0
238.00	387.0
238.90	1229.0
239.90	1026.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

241.00	2028.0
242.10	4257.0
243.00	5134.0
244.00	68840.0
245.00	9517.0
246.00	13120.0
246.90	2331.0
247.90	539.0
249.00	2820.0
249.90	490.0
251.00	504.0
252.00	494.0
253.00	1357.0
255.00	350720.0
256.00	51192.0
257.00	4332.0
258.00	18424.0
259.00	3054.0
260.00	573.0
260.90	740.0
264.00	479.0
265.00	7966.0
266.00	930.0
269.80	392.0
270.90	758.0
272.10	712.0
273.00	10664.0
274.00	29288.0
275.00	169728.0
276.00	21944.0
277.00	13982.0
278.00	2855.0
278.90	365.0
280.80	403.0
282.10	351.0
283.00	1697.0
284.00	982.0
285.00	2373.0
286.00	502.0
289.00	654.0
290.00	530.0
290.90	414.0
292.00	681.0
293.00	2767.0
294.00	836.0
296.00	47608.0
297.00	6434.0
300.90	643.0
302.00	1061.0
303.00	5515.0
304.00	1582.0
304.90	305.0
307.90	590.0
308.80	358.0
310.00	815.0
313.00	501.0
314.00	2183.0
315.00	5106.0
316.00	2694.0
317.00	544.0
321.00	1604.0
322.10	828.0
323.00	16092.0
324.00	3112.0
325.90	310.0
327.00	2728.0
327.90	1590.0
329.00	361.0
332.00	1181.0
333.00	1453.0
334.00	10934.0
335.00	3236.0
339.00	320.0
341.00	2098.0
342.00	409.0
346.00	3973.0
346.90	795.0
352.00	5243.0
353.00	3681.0
354.00	5286.0
355.10	1029.0
359.00	353.0
365.00	23512.0
365.90	3211.0
369.80	465.0
371.00	1304.0
372.00	9636.0
373.10	2325.0
382.90	2715.0
383.90	690.0
384.90	350.0
390.00	1536.0
391.00	742.0
392.00	466.0
400.90	578.0
401.90	4257.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213

403.00	5132.0
404.00	2146.0
405.00	352.0
420.90	4977.0
422.00	5176.0
423.00	39680.0
424.00	8909.0
425.00	1122.0
441.00	110976.0
442.10	653760.0
443.00	147200.0
444.00	12167.0
445.00	795.0

Instrument :
BNA_N
ClientSampleId :
DFTPP213