

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110322\
 Data File : BN022482.D
 Acq On : 03 Nov 2022 17:17
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN110322

Quant Time: Nov 04 03:30:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 04 03:28:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.601	0.657	-9.3	92	0.00
3	n-Nitrosodimethylamine	0.624	0.647	-3.7	92	0.00
4 S	2-Fluorophenol	1.130	1.289	-14.1	95	0.00
5 S	Phenol-d6	1.424	1.578	-10.8	93	0.00
6	bis(2-Chloroethyl)ether	1.275	1.331	-4.4	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.361	0.383	-6.1	108	0.00
9	Naphthalene	1.212	1.167	3.7	94	0.01
10	Hexachlorobutadiene	0.209	0.201	3.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.671	0.628	6.4	94	0.00
12	2-Methylnaphthalene	0.288	0.257	10.8	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.210	0.197	6.2	92	0.00
15 S	2-Fluorobiphenyl	1.667	1.708	-2.5	93	0.00
16	Acenaphthylene	2.030	1.862	8.3	93	0.00
17	Acenaphthene	1.308	1.260	3.7	94	0.00
18	Fluorene	1.834	1.763	3.9	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.065	0.051	21.5	99	0.00
21	4-Bromophenyl-phenylether	0.247	0.232	6.1	93	0.00
22	Hexachlorobenzene	0.282	0.276	2.1	95	0.00
23	Atrazine	0.223	0.201	9.9	92	0.00
24	Pentachlorophenol	0.107	0.088	17.8	94	0.00
25	Phenanthrene	1.337	1.264	5.5	94	0.00
26	Anthracene	1.176	1.086	7.7	94	0.00
27 SURR	Fluoranthene-d10	1.093	1.066	2.5	94	0.00
28	Fluoranthene	1.441	1.405	2.5	93	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
30	Pyrene	1.727	1.587	8.1	82	0.00
31 S	Terphenyl-d14	1.247	1.243	0.3	86	0.00
32	Benzo(a)anthracene	1.608	1.511	6.0	82	0.00
33	Chrysene	1.593	1.559	2.1	84	0.00
34	Bis(2-ethylhexyl)phthalate	1.355	1.129	16.7	76	0.00
35 I	Perylene-d12	1.000	1.000	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	1.926	1.855	3.7	93	0.00
37	Benzo(b)fluoranthene	1.573	1.536	2.4	93	0.00
38	Benzo(k)fluoranthene	1.604	1.576	1.7	94	0.00
39 C	Benzo(a)pyrene	1.347	1.314	2.4	93	0.00
40	Dibenzo(a,h)anthracene	1.532	1.475	3.7	93	0.00
41	Benzo(g,h,i)perylene	1.628	1.572	3.4	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN110322