

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110420\
 Data File : BN012498.D
 Acq On : 04 Nov 2020 21:21
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN110420

Quant Time: Nov 05 05:22:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:18:04 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.626	0.554	11.5	92	0.00
3	n-Nitrosodimethylamine	1.709	1.554	9.1	99	0.00
4 S	2-Fluorophenol	1.398	1.229	12.1	92	0.00
5 S	Phenol-d6	1.636	1.446	11.6	92	0.00
6	bis(2-Chloroethyl)ether	1.115	0.917	17.8	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.473	0.425	10.1	98	0.00
9	Naphthalene	1.122	1.003	10.6	97	0.00
10	Hexachlorobutadiene	0.245	0.220	10.2	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.580	10.6	99	0.00
12	2-Methylnaphthalene	0.724	0.636	12.2	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.290	0.263	9.3	100	0.00
15 S	2-Fluorobiphenyl	1.560	1.420	9.0	96	0.00
16	Acenaphthylene	2.005	1.800	10.2	99	0.00
17	Acenaphthene	1.276	1.183	7.3	102	0.00
18	Fluorene	1.627	1.504	7.6	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.097	0.083	14.4	102	0.00
21	4-Bromophenyl-phenylether	0.231	0.203	12.1	99	0.00
22	Hexachlorobenzene	0.301	0.277	8.0	100	-0.01
23	Atrazine	0.240	0.217	9.6	102	0.00
24	Pentachlorophenol	0.141	0.131	7.1	111	0.00
25	Phenanthrene	1.184	1.076	9.1	101	0.00
26	Anthracene	1.096	0.959	12.5	99	0.00
27 SURR	Fluoranthene-d10	1.240	1.130	8.9	103	0.00
28	Fluoranthene	1.413	1.264	10.5	101	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
30	Pyrene	1.344	1.227	8.7	104	0.00
31 S	Terphenyl-d14	0.954	0.873	8.5	103	0.00
32	Benzo(a)anthracene	1.273	1.158	9.0	103	0.00
33	Chrysene	1.333	1.187	11.0	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.864	0.796	7.9	108	0.00
35	Indeno(1,2,3-cd)pyrene	2.006	1.951	2.7	103	0.00
36 I	Perylene-d12	1.000	1.000	0.0	102	0.00
37	Benzo(b)fluoranthene	1.276	1.120	12.2	102	0.00
38	Benzo(k)fluoranthene	1.360	1.239	8.9	104	0.00
39 C	Benzo(a)pyrene	1.249	1.130	9.5	104	0.00
40	Dibenzo(a,h)anthracene	1.416	1.322	6.6	101	0.00
41	Benzo(g,h,i)perylene	1.458	1.366	6.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN110420