

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110420\
 Data File : BN012497.D
 Acq On : 04 Nov 2020 20:08
 Operator : CG/JU
 Sample : SSTDIC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC5.0

Quant Time: Nov 05 05:08:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 04 18:15:52 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.490	152	704	0.40	ng	0.00
7) Naphthalene-d8	10.250	136	2858	0.40	ng	# 0.00
13) Acenaphthene-d10	14.128	164	1599	0.40	ng	-0.01
19) Phenanthrene-d10	16.883	188	3355	0.40	ng	# 0.00
29) Chrysene-d12	21.089	240	3823	0.40	ng	#-0.01
36) Perylene-d12	23.220	264	4173	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.098	112	11136	4.50	ng	0.00
5) Phenol-d6	6.668	99	14431	4.74	ng	0.00
8) Nitrobenzene-d5	8.628	82	17892	5.43	ng	-0.01
11) 2-Methylnaphthalene-d10	11.851	152	23106	4.93	ng	0.00
14) 2,4,6-Tribromophenol	15.638	330	5724	5.30	ng	0.00
15) 2-Fluorobiphenyl	12.746	172	30110	5.01	ng	-0.01
27) Fluoranthene-d10	18.923	212	52175	5.02	ng	0.00
31) Terphenyl-d14	19.539	244	42401	4.77	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.044	88	4841	4.09	ng	# 80
3) n-Nitrosodimethylamine	3.358	42	14474	4.98	ng	# 66
6) bis(2-Chloroethyl)ether	6.926	93	10460	5.26	ng	# 87
9) Naphthalene	10.301	128	38572	4.78	ng	95
10) Hexachlorobutadiene	10.592	225	7969	4.76	ng	# 100
12) 2-Methylnaphthalene	11.922	142	26253	5.02	ng	# 88
16) Acenaphthylene	13.849	152	39638	4.98	ng	100
17) Acenaphthene	14.192	154	24724	4.86	ng	92
18) Fluorene	15.194	166	31926	5.04	ng	99
20) 4,6-Dinitro-2-methylph...	15.283	198	4677	6.75	ng	# 11
21) 4-Bromophenyl-phenylether	16.080	248	9733	4.77	ng	# 64
22) Hexachlorobenzene	16.189	284	11981	4.68	ng	# 84
23) Atrazine	16.360	200	9917	5.08	ng	97
24) Pentachlorophenol	16.542	266	6139	5.30	ng	98
25) Phenanthrene	16.919	178	49704	4.97	ng	98
26) Anthracene	17.017	178	48038	5.17	ng	99
28) Fluoranthene	18.953	202	59423	5.08	ng	100
30) Pyrene	19.320	202	60705	4.65	ng	100
32) Benzo(a)anthracene	21.078	228	61228	5.14	ng	98
33) Chrysene	21.131	228	61005	4.69	ng	98
34) Bis(2-ethylhexyl)phtha...	21.025	149	39839	5.12	ng	# 90
35) Indeno(1,2,3-cd)pyrene	25.315	276	87466	4.60	ng	98
37) Benzo(b)fluoranthene	22.590	252	69202	5.14	ng	# 94
38) Benzo(k)fluoranthene	22.631	252	67816	4.81	ng	# 95
39) Benzo(a)pyrene	23.128	252	64352	5.01	ng	# 92
40) Dibenzo(a,h)anthracene	25.329	278	72949	5.02	ng	95
41) Benzo(g,h,i)perylene	25.955	276	71780	4.73	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110420\
Data File : BN012497.D
Acq On : 04 Nov 2020 20:08
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Nov 05 05:08:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 04 18:15:52 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.490 min Scan# 560
 Delta R.T. 0.001 min
 Lab File: BN012497.D
 Acq: 04 Nov 2020 20:08

Instrument :
 BNA_N
 ClientSampled :
 SSTDICC5.0

Tgt Ion: 152 Resp: 704




Ion	Ratio	Lower	Upper
152	100		
150	156.0	116.6	174.8
115	77.7	52.9	79.3




#2
 1,4-Dioxane
 Concen: 4.09 ng
 RT: 3.044 min Scan# 8
 Delta R.T. -0.016 min
 Lab File: BN012497.D
 Acq: 04 Nov 2020 20:08

Tgt Ion: 88 Resp: 4841

Ion	Ratio	Lower	Upper
88	100		
58	82.6	59.4	89.2
43	65.0	32.3	48.5







Abundance Scan 50 (3.382 min): BN012493.D\data.ms (-46)



#3
n-Nitrosodimethylamine
Concen: 4.98 ng
RT: 3.358 min Scan# 47
Delta R.T. -0.024 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Tgt Ion: 42 Resp: 14474
Ion Ratio Lower Upper
42 100
74 49.3 64.4 96.6#
44 3.3 2.8 4.2


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



#4
2-Fluorophenol
Concen: 4.50 ng
RT: 5.098 min Scan# 263
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion: 112 Resp: 11136
Ion Ratio Lower Upper
112 100
64 73.6 49.5 74.3
63 57.2 32.0 48.0#









#5
Phenol-d6
Concen: 4.74 ng
RT: 6.668 min Scan# 458
Delta R.T. -0.008 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 14431
Ion Ratio Lower Upper
99 100
42 50.4 22.4 33.6#
71 65.0 37.8 56.6#



#6
bis(2-Chloroethyl)ether
Concen: 5.26 ng
RT: 6.926 min Scan# 490
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion: 93 Resp: 10460
Ion Ratio Lower Upper
93 100
63 95.5 63.1 94.7#
95 32.6 26.1 39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.250 min Scan# 815
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	2858
Ion	Ratio	Lower	Upper
136	100		
137	12.7	10.6	15.8
54	11.7	8.6	12.8
68	5.7	5.7	8.5#



#8
Nitrobenzene-d5
Concen: 5.43 ng
RT: 8.628 min Scan# 687
Delta R.T. -0.013 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:	82	Resp:	17892
Ion	Ratio	Lower	Upper
82	100		
128	30.0	30.6	46.0#
54	70.2	48.3	72.5









#9
Naphthalene
Concen: 4.78 ng
RT: 10.301 min Scan# 819
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	128	Resp:	38572
Ion Ratio	Lower	Upper	
128	100		
129	10.9	9.8	14.8
127	13.1	12.2	18.4



#10
Hexachlorobutadiene
Concen: 4.76 ng
RT: 10.592 min Scan# 842
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:	225	Resp:	7969
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.8	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 4.93 ng
RT: 11.851 min Scan# 990
Delta R.T. -0.004 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 23106
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 5.02 ng
RT: 11.922 min Scan# 1006
Delta R.T. -0.009 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:142 Resp: 26253
Ion Ratio Lower Upper
142 100
141 91.7 69.4 104.2
115 60.2 35.7 53.5#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.128 min Scan# 1265
Delta R.T. -0.013 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 1599
Ion Ratio Lower Upper
164 100
162 105.1 80.6 120.8
160 51.5 39.5 59.3



#14
2,4,6-Tribromophenol
Concen: 5.30 ng
RT: 15.638 min Scan# 1384
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:330 Resp: 5724
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 53.9 34.4 51.6#









#15
2-Fluorobiphenyl
Concen: 5.01 ng
RT: 12.746 min Scan# 1156
Delta R.T. -0.013 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172	Resp:	30110
Ion Ratio	Lower	Upper
172	100	
171	35.7	28.2 42.4
170	24.7	20.0 30.0



#16
Acenaphthylene
Concen: 4.98 ng
RT: 13.849 min Scan# 1243
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:152	Resp:	39638
Ion Ratio	Lower	Upper
152	100	
151	19.9	15.8 23.8
153	13.0	10.6 16.0









#17
Acenaphthene
Concen: 4.86 ng
RT: 14.192 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 24724
Ion Ratio Lower Upper
154 100
153 115.7 83.6 125.4
152 56.5 43.9 65.9



#18
Fluorene
Concen: 5.04 ng
RT: 15.194 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:166 Resp: 31926
Ion Ratio Lower Upper
166 100
165 99.8 78.6 118.0
167 13.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.883 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188	Resp:	3355
Ion	Ratio	Lower Upper
188	100	
94	0.0	0.0 0.0
80	8.0	5.0 7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 6.75 ng
RT: 15.283 min Scan# 1356
Delta R.T. -0.013 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:198	Resp:	4677
Ion	Ratio	Lower Upper
198	100	
51	45.3	152.0 228.0#
105	30.3	80.0 120.0#









#21
4-Bromophenyl-phenylether
Concen: 4.77 ng
RT: 16.080 min Scan# 1419
Delta R.T. -0.012 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 9733
Ion Ratio Lower Upper
248 100
250 87.9 77.3 115.9
141 131.7 57.8 86.8#



#22
Hexachlorobenzene
Concen: 4.68 ng
RT: 16.189 min Scan# 1428
Delta R.T. -0.012 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:284 Resp: 11981
Ion Ratio Lower Upper
284 100
142 56.0 32.0 48.0#
249 30.7 27.0 40.6









#23
Atrazine
Concen: 5.08 ng
RT: 16.360 min Scan# 1442
Delta R.T. -0.012 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	200	Resp:	9917
Ion	Ratio	Lower	Upper
200	100		
173	29.7	22.8	34.2
215	44.9	38.1	57.1



#24
Pentachlorophenol
Concen: 5.30 ng
RT: 16.542 min Scan# 1457
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:	266	Resp:	6139
Ion	Ratio	Lower	Upper
266	100		
264	62.1	50.9	76.3
268	63.5	51.5	77.3









#25
Phenanthrene
Concen: 4.97 ng
RT: 16.919 min Scan# 1488
Delta R.T. -0.012 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 49704
Ion Ratio Lower Upper
178 100
176 19.7 14.8 22.2
179 15.2 12.4 18.6



#26
Anthracene
Concen: 5.17 ng
RT: 17.017 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:178 Resp: 48038
Ion Ratio Lower Upper
178 100
176 19.0 14.8 22.2
179 15.2 12.5 18.7









#27
Fluoranthene-d10
Concen: 5.02 ng
RT: 18.923 min Scan# 1761
Delta R.T. -0.005 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 52175
Ion Ratio Lower Upper
212 100
106 7.8 6.8 10.2
104 5.0 4.1 6.1



#28
Fluoranthene
Concen: 5.08 ng
RT: 18.953 min Scan# 1767
Delta R.T. -0.005 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:202 Resp: 59423
Ion Ratio Lower Upper
202 100
101 7.2 5.8 8.6
203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.089 min Scan# 2081
Delta R.T. -0.011 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 3823
Ion Ratio Lower Upper
240 100
120 10.0 5.3 7.9#
236 27.4 21.8 32.8



#30
Pyrene
Concen: 4.65 ng
RT: 19.320 min Scan# 1841
Delta R.T. -0.005 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:202 Resp: 60705
Ion Ratio Lower Upper
202 100
200 21.0 16.6 24.8
203 17.4 13.8 20.8









#31
 Terphenyl-d14
 Concen: 4.77 ng
 RT: 19.539 min Scan# 1885
 Delta R.T. -0.005 min
 Lab File: BN012497.D
 Acq: 04 Nov 2020 20:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244	Resp:	42401
Ion Ratio	Lower	Upper
244	100	
212	9.3	7.3 10.9
122	9.2	7.3 10.9



#32
 Benzo(a)anthracene
 Concen: 5.14 ng
 RT: 21.078 min Scan# 2080
 Delta R.T. 0.000 min
 Lab File: BN012497.D
 Acq: 04 Nov 2020 20:08

Tgt Ion:228	Resp:	61228
Ion Ratio	Lower	Upper
228	100	
226	26.4	22.3 33.5
229	20.1	15.7 23.5









#33
Chrysene
Concen: 4.69 ng
RT: 21.131 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	228	Resp:	61005
Ion	Ratio	Lower	Upper
228	100		
226	29.0	24.2	36.2
229	19.8	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 5.12 ng
RT: 21.025 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:	149	Resp:	39839
Ion	Ratio	Lower	Upper
149	100		
167	23.0	23.2	34.8#
279	4.0	3.9	5.9









#35
Indeno(1,2,3-cd)pyrene
Concen: 4.60 ng
RT: 25.315 min Scan# 3055
Delta R.T. -0.007 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 87466
Ion Ratio Lower Upper
276 100
138 15.1 13.3 19.9
277 24.8 20.2 30.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.220 min Scan# 2443
Delta R.T. -0.003 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:264 Resp: 4173
Ion Ratio Lower Upper
264 100
260 25.2 19.8 29.6
265 61.1 51.4 77.0









#37
Benzo(b)fluoranthene
Concen: 5.14 ng
RT: 22.590 min Scan# 2259
Delta R.T. -0.003 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 69202
Ion Ratio Lower Upper
252 100
253 22.4 20.1 30.1
125 6.9 7.2 10.8#



#38
Benzo(k)fluoranthene
Concen: 4.81 ng
RT: 22.631 min Scan# 2271
Delta R.T. -0.003 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:252 Resp: 67816
Ion Ratio Lower Upper
252 100
253 22.3 19.8 29.8
125 6.8 7.1 10.7#









#39
Benzo(a)pyrene
Concen: 5.01 ng
RT: 23.128 min Scan# 2416
Delta R.T. -0.003 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 64352
Ion Ratio Lower Upper
252 100
253 22.4 21.3 31.9
125 8.2 8.3 12.5#



#40
Dibenzo(a,h)anthracene
Concen: 5.02 ng
RT: 25.329 min Scan# 3059
Delta R.T. -0.007 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08


Tgt Ion:278 Resp: 72949
Ion Ratio Lower Upper
278 100
139 10.3 9.3 13.9
279 23.8 21.6 32.4





#41
Benzo(g,h,i)perylene
Concen: 4.73 ng
RT: 25.955 min Scan# 3242
Delta R.T. -0.007 min
Lab File: BN012497.D
Acq: 04 Nov 2020 20:08

Tgt Ion:	276	Resp:	71780
Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.8	29.8
138	13.3	11.7	17.5



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0