

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110422\
 Data File : BN022511.D
 Acq On : 04 Nov 2022 16:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 05 01:57:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 04 03:28:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	0.00
2	1,4-Dioxane	0.400	0.349	12.8	108	0.00
3	n-Nitrosodimethylamine	0.400	0.366	8.5	118	0.00
4 S	2-Fluorophenol	0.400	0.477	-19.2	145	0.00
5 S	Phenol-d6	0.400	0.480	-20.0	147	0.00
6	bis(2-Chloroethyl)ether	0.400	0.436	-9.0	136	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.400	0.500	-25.0	136	-0.01
9	Naphthalene	0.400	0.355	11.3	93	0.00
10	Hexachlorobutadiene	0.400	0.335	16.3	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.416	-4.0	112	-0.01
12	2-Methylnaphthalene	0.400	0.400	0.0	115	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.405	-1.3	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.414	-3.5	109	-0.01
16	Acenaphthylene	0.400	0.376	6.0	110	-0.01
17	Acenaphthene	0.400	0.385	3.8	108	-0.01
18	Fluorene	0.400	0.376	6.0	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.382	4.5	84	0.00
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	106	-0.01
22	Hexachlorobenzene	0.400	0.386	3.5	107	0.00
23	Atrazine	0.400	0.369	7.8	108	0.00
24	Pentachlorophenol	0.400	0.463	-15.8	129	-0.01
25	Phenanthrene	0.400	0.377	5.8	107	-0.01
26	Anthracene	0.400	0.373	6.8	108	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	107	0.00
28	Fluoranthene	0.400	0.390	2.5	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.405	-1.3	106	0.00
31 S	Terphenyl-d14	0.400	0.414	-3.5	104	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	99	0.00
33	Chrysene	0.400	0.386	3.5	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	94	0.00
35 I	Perylene-d12	0.400	0.400	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	92	-0.01
37	Benzo(b)fluoranthene	0.400	0.394	1.5	96	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	95	0.00
39 C	Benzo(a)pyrene	0.400	0.393	1.8	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.369	7.8	92	-0.01
41	Benzo(g,h,i)perylene	0.400	0.361	9.8	89	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4