

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110424\
 Data File : BN034842.D
 Acq On : 04 Nov 2024 16:43
 Operator : RC/JU
 Sample : P4652-01DL 2X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW5-SP100-50-20241030DL

Quant Time: Nov 04 17:45:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN103024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 30 13:28:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.590	152	2324	0.400	ng	-0.01
7) Naphthalene-d8	10.362	136	7023	0.400	ng	-0.01
13) Acenaphthene-d10	14.219	164	3514	0.400	ng	-0.01
19) Phenanthrene-d10	16.964	188	7039	0.400	ng	#-0.02
29) Chrysene-d12	21.158	240	3897	0.400	ng	#-0.02
35) Perylene-d12	23.342	264	3011	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.213	112	641	0.091	ng	0.00
5) Phenol-d6	6.773	99	560	0.061	ng	0.00
8) Nitrobenzene-d5	8.728	82	983	0.178	ng	-0.01
11) 2-Methylnaphthalene-d10	11.954	152	1633	0.161	ng	-0.01
14) 2,4,6-Tribromophenol	15.723	330	202	0.116	ng	0.00
15) 2-Fluorobiphenyl	12.839	172	2495	0.180	ng	-0.02
27) Fluoranthene-d10	19.004	212	3248	0.197	ng	-0.01
31) Terphenyl-d14	19.612	244	1920	0.229	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.198	88	10305	4.045	ng	93
28) Fluoranthene	19.032	202	457	0.020	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110424\
Data File : BN034842.D
Acq On : 04 Nov 2024 16:43
Operator : RC/JU
Sample : P4652-01DL 2X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW5-SP100-50-20241030DL

Quant Time: Nov 04 17:45:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN103024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 30 13:28:24 2024
Response via : Initial Calibration

