

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038132.D
 Acq On : 04 Nov 2025 12:19
 Operator : RC/JU
 Sample : PB170381BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170381BS

Quant Time: Nov 04 12:44:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	11523	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	31801	0.400	ng	# 0.00
13) Acenaphthene-d10	14.441	164	16154	0.400	ng	0.01
19) Phenanthrene-d10	17.186	188	30417	0.400	ng	0.00
29) Chrysene-d12	21.375	240	16510	0.400	ng	0.00
35) Perylene-d12	23.686	264	13194	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	9248	0.341	ng	0.00
5) Phenol-d6	6.937	99	10615	0.321	ng	0.00
8) Nitrobenzene-d5	8.929	82	9450	0.368	ng	0.00
11) 2-Methylnaphthalene-d10	12.177	152	15982	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1687	0.253	ng	0.00
15) 2-Fluorobiphenyl	13.062	172	26882	0.415	ng	0.01
27) Fluoranthene-d10	19.220	212	23204	0.303	ng	0.00
31) Terphenyl-d14	19.824	244	17337	0.483	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.261	88	3965	0.333	ng	# 62
3) n-Nitrosodimethylamine	3.564	42	5606	0.366	ng	# 94
6) bis(2-Chloroethyl)ether	7.197	93	11700	0.360	ng	99
9) Naphthalene	10.626	128	31176	0.356	ng	100
10) Hexachlorobutadiene	10.925	225	5678	0.384	ng	# 99
12) 2-Methylnaphthalene	12.248	142	17976	0.308	ng	99
16) Acenaphthylene	14.152	152	28630	0.391	ng	99
17) Acenaphthene	14.494	154	17838	0.349	ng	98
18) Fluorene	15.489	166	24190	0.361	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	1040	0.363	ng	95
21) 4-Bromophenyl-phenylether	16.379	248	7282	0.365	ng	97
22) Hexachlorobenzene	16.503	284	8817	0.389	ng	99
23) Atrazine	16.652	200	4704	0.322	ng	99
24) Pentachlorophenol	16.838	266	4384	0.449	ng	99
25) Phenanthrene	17.223	178	35147	0.364	ng	100
26) Anthracene	17.322	178	30413	0.360	ng	99
28) Fluoranthene	19.252	202	32822	0.304	ng	100
30) Pyrene	19.615	202	31678	0.425	ng	100
32) Benzo(a)anthracene	21.357	228	20604	0.350	ng	99
33) Chrysene	21.411	228	24309	0.380	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	8897	0.287	ng	98
36) Indeno(1,2,3-cd)pyrene	26.036	276	23949	0.442	ng	98
37) Benzo(b)fluoranthene	22.987	252	20712	0.367	ng	98
38) Benzo(k)fluoranthene	23.031	252	21547	0.379	ng	98
39) Benzo(a)pyrene	23.584	252	17494	0.387	ng	98
40) Dibenzo(a,h)anthracene	26.054	278	17865	0.428	ng	99
41) Benzo(g,h,i)perylene	26.756	276	21276	0.447	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038132.D
Acq On : 04 Nov 2025 12:19
Operator : RC/JU
Sample : PB170381BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170381BS

Quant Time: Nov 04 12:44:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

