

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
 Data File : BN038135.D
 Acq On : 04 Nov 2025 14:08
 Operator : RC/JU
 Sample : Q3525-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW5-SP100-20251030

Quant Time: Nov 04 14:32:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	8620	0.400	ng	0.00
7) Naphthalene-d8	10.573	136	25357	0.400	ng	# 0.00
13) Acenaphthene-d10	14.430	164	13158	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	22343	0.400	ng	0.00
29) Chrysene-d12	21.375	240	15115	0.400	ng	# 0.00
35) Perylene-d12	23.683	264	14719	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2495	0.123	ng	0.00
5) Phenol-d6	6.937	99	1619	0.065	ng	0.00
8) Nitrobenzene-d5	8.929	82	6316	0.309	ng	0.00
11) 2-Methylnaphthalene-d10	12.172	152	9507	0.263	ng	0.00
14) 2,4,6-Tribromophenol	15.932	330	1331	0.246	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	16049	0.305	ng	0.00
27) Fluoranthene-d10	19.220	212	19109	0.339	ng	0.00
31) Terphenyl-d14	19.819	244	15576	0.474	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	53034	5.954	ng	99
6) bis(2-Chloroethyl)ether	7.197	93	930	0.038	ng	93
34) Bis(2-ethylhexyl)phtha...	21.295	149	8915	0.314	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110425\
Data File : BN038135.D
Acq On : 04 Nov 2025 14:08
Operator : RC/JU
Sample : Q3525-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW5-SP100-20251030

Quant Time: Nov 04 14:32:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 29 13:09:46 2025
Response via : Initial Calibration

