

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110520\  
 Data File : BN012502.D  
 Acq On : 05 Nov 2020 15:49  
 Operator : CG/JU  
 Sample : PB132704BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 PB132704BL

Quant Time: Nov 05 16:38:38 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN110420.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 05 05:46:47 2020  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|--------|------|----------|------|-------|----------|
| -----                       |        |      |          |      |       |          |
| Internal Standards          |        |      |          |      |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.482  | 152  | 864      | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8           | 10.250 | 136  | 3275     | 0.40 | ng    | # 0.00   |
| 13) Acenaphthene-d10        | 14.129 | 164  | 1960     | 0.40 | ng    | -0.01    |
| 19) Phenanthrene-d10        | 16.883 | 188  | 4067     | 0.40 | ng    | # 0.00   |
| 29) Chrysene-d12            | 21.099 | 240  | 4355     | 0.40 | ng    | 0.00     |
| 36) Perylene-d12            | 23.224 | 264  | 5044     | 0.40 | ng    | # 0.00   |
| System Monitoring Compounds |        |      |          |      |       |          |
| 4) 2-Fluorophenol           | 5.090  | 112  | 991      | 0.33 | ng    | 0.00     |
| 5) Phenol-d6                | 6.668  | 99   | 1115     | 0.32 | ng    | 0.00     |
| 8) Nitrobenzene-d5          | 8.641  | 82   | 1427     | 0.37 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10 | 11.855 | 152  | 1851     | 0.35 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol    | 15.651 | 330  | 441      | 0.31 | ng    | 0.01     |
| 15) 2-Fluorobiphenyl        | 12.758 | 172  | 2870     | 0.38 | ng    | 0.00     |
| 27) Fluoranthene-d10        | 18.928 | 212  | 4237     | 0.34 | ng    | 0.00     |
| 31) Terphenyl-d14           | 19.539 | 244  | 4888     | 0.47 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110520\  
Data File : BN012502.D  
Acq On : 05 Nov 2020 15:49  
Operator : CG/JU  
Sample : PB132704BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
PB132704BL

Quant Time: Nov 05 16:38:38 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN110420.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 05 05:46:47 2020  
Response via : Initial Calibration

