

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110520\
 Data File : BN012501.D
 Acq On : 05 Nov 2020 15:08
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 05 16:38:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.626	0.529	15.5	85	0.00
3	n-Nitrosodimethylamine	1.709	1.571	8.1	97	0.00
4 S	2-Fluorophenol	1.398	1.267	9.4	92	0.00
5 S	Phenol-d6	1.636	1.518	7.2	93	0.00
6	bis(2-Chloroethyl)ether	1.115	1.081	3.0	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.473	0.445	5.9	101	-0.01
9	Naphthalene	1.122	1.027	8.5	98	0.00
10	Hexachlorobutadiene	0.245	0.242	1.2	104	-0.01
11 SURR	2-Methylnaphthalene-d10	0.649	0.600	7.6	101	0.00
12	2-Methylnaphthalene	0.724	0.680	6.1	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	-0.01
14 S	2,4,6-Tribromophenol	0.290	0.269	7.2	105	0.00
15 S	2-Fluorobiphenyl	1.560	1.425	8.7	99	-0.01
16	Acenaphthylene	2.005	1.804	10.0	102	0.00
17	Acenaphthene	1.276	1.144	10.3	101	0.00
18	Fluorene	1.627	1.450	10.9	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.097	0.083	14.4	106	0.00
21	4-Bromophenyl-phenylether	0.231	0.207	10.4	105	0.00
22	Hexachlorobenzene	0.301	0.275	8.6	103	-0.01
23	Atrazine	0.240	0.209	12.9	102	0.00
24	Pentachlorophenol	0.141	0.124	12.1	108	0.00
25	Phenanthrene	1.184	1.036	12.5	101	-0.01
26	Anthracene	1.096	0.968	11.7	104	0.00
27 SURR	Fluoranthene-d10	1.240	1.110	10.5	104	0.00
28	Fluoranthene	1.413	1.250	11.5	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
30	Pyrene	1.344	1.212	9.8	105	0.00
31 S	Terphenyl-d14	0.954	0.883	7.4	106	0.00
32	Benzo(a)anthracene	1.273	1.184	7.0	108	0.00
33	Chrysene	1.333	1.185	11.1	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.864	0.738	14.6	102	0.00
35	Indeno(1,2,3-cd)pyrene	2.006	1.891	5.7	102	0.00
36 I	Perylene-d12	1.000	1.000	0.0	103	0.00
37	Benzo(b)fluoranthene	1.276	1.135	11.1	105	0.00
38	Benzo(k)fluoranthene	1.360	1.252	7.9	106	0.00
39 C	Benzo(a)pyrene	1.249	1.134	9.2	106	0.00
40	Dibenzo(a,h)anthracene	1.416	1.310	7.5	102	0.00
41	Benzo(g,h,i)perylene	1.458	1.338	8.2	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4