

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022539.D
 Acq On : 07 Nov 2022 23:31
 Operator : CG/JU
 Sample : N5408-04
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y4

Quant Time: Nov 08 00:17:03 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.915	152	2138	0.400	ng/ul	0.00
4) Naphthalene-d8	10.739	136	7524	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.595	164	4815	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.353	188	10940	0.400	ng/ul	0.00
17) Chrysene-d12	21.555	240	8551	0.400	ng/ul	0.00
23) Perylene-d12	24.002	264	7046	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	10415	3.748	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.339	152	3493	0.307	ng/ul	0.00
18) Fluoranthene-d10	19.389	212	10273	0.358	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.261	88	505	0.177	ng/ul#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022539.D
Acq On : 07 Nov 2022 23:31
Operator : CG/JU
Sample : N5408-04
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y4

Quant Time: Nov 08 00:17:03 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

