

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110722\  
 Data File : BN022541.D  
 Acq On : 08 Nov 2022 00:46  
 Operator : CG/JU  
 Sample : N5408-05  
 Misc :  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 EW4Y5

Quant Time: Nov 08 02:06:35 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN102022.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Nov 07 23:39:33 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.915  | 152  | 1867     | 0.400 | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.739 | 136  | 6630     | 0.400 | ng/ul  | 0.00     |
| 9) Acenaphthene-d10         | 14.591 | 164  | 4319     | 0.400 | ng/ul  | 0.00     |
| 13) Phenanthrene-d10        | 17.349 | 188  | 9761     | 0.400 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.553 | 240  | 8113     | 0.400 | ng/ul  | 0.00     |
| 23) Perylene-d12            | 23.999 | 264  | 6845     | 0.400 | ng/ul  | 0.00     |
| System Monitoring Compounds |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8           | 3.223  | 96   | 9255     | 3.814 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.339 | 152  | 3621     | 0.361 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.389 | 212  | 10407    | 0.382 | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |       |        |          |
|                             |        |      |          |       | Qvalue |          |
| 2) 1,4-Dioxane              | 3.261  | 88   | 488      | 0.196 | ng/ul# | 70       |
| -----                       |        |      |          |       |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN110722\  
Data File : BN022541.D  
Acq On : 08 Nov 2022 00:46  
Operator : CG/JU  
Sample : N5408-05  
Misc :  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
EW4Y5

Quant Time: Nov 08 02:06:35 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN102022.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Nov 07 23:39:33 2022  
Response via : Initial Calibration

