

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022547.D
 Acq On : 08 Nov 2022 04:29
 Operator : CG/JU
 Sample : N5408-08
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y8

Quant Time: Nov 08 05:44:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1836	0.400	ng/ul	0.00
4) Naphthalene-d8	10.738	136	6464	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.590	164	4135	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.348	188	9225	0.400	ng/ul	0.00
17) Chrysene-d12	21.553	240	7484	0.400	ng/ul	0.00
23) Perylene-d12	24.000	264	6285	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.224	96	10468	4.387	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.338	152	3797	0.388	ng/ul	0.00
18) Fluoranthene-d10	19.388	212	10881	0.433	ng/ul	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	407	0.166	ng/ul#	Qvalue 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022547.D
Acq On : 08 Nov 2022 04:29
Operator : CG/JU
Sample : N5408-08
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y8

Quant Time: Nov 08 05:44:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

