

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022548.D
 Acq On : 08 Nov 2022 05:06
 Operator : CG/JU
 Sample : N5407-10
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y9

Quant Time: Nov 08 05:45:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1966	0.400	ng/u1	0.00
4) Naphthalene-d8	10.739	136	6805	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.591	164	4373	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.349	188	9821	0.400	ng/u1	0.00
17) Chrysene-d12	21.553	240	7804	0.400	ng/u1	0.00
23) Perylene-d12	23.996	264	6469	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.224	96	12819	5.017	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.340	152	4084	0.397	ng/u1	0.00
18) Fluoranthene-d10	19.384	212	11623	0.444	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022548.D
Acq On : 08 Nov 2022 05:06
Operator : CG/JU
Sample : N5407-10
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y9

Quant Time: Nov 08 05:45:29 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

