

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022552.D
 Acq On : 08 Nov 2022 07:34
 Operator : CG/JU
 Sample : PB148827BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS827

Quant Time: Nov 08 08:04:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1787	0.400	ng/u1	0.00
4) Naphthalene-d8	10.739	136	6308	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.591	164	4218	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.349	188	9160	0.400	ng/u1	0.00
17) Chrysene-d12	21.553	240	8020	0.400	ng/u1	0.00
23) Perylene-d12	23.999	264	6506	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	1494	0.643	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.334	152	3833	0.402	ng/u1	-0.01
18) Fluoranthene-d10	19.384	212	9669	0.359	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.257	88	3731	1.565	ng/u1	90
5) Naphthalene	10.789	128	6475	0.347	ng/u1	98
7) 2-Methylnaphthalene	12.411	142	4337	0.369	ng/u1	99
8) 1-Methylnaphthalene	12.631	142	4755	0.394	ng/u1	100
10) Acenaphthylene	14.309	152	6952	0.367	ng/u1	98
11) Acenaphthene	14.651	153	5412	0.339	ng/u1	99
12) Fluorene	15.641	166	6155	0.332	ng/u1	98
14) Pentachlorophenol	17.007	266	1424	0.654	ng/u1	99
15) Phenanthrene	17.391	178	10335	0.336	ng/u1	99
16) Anthracene	17.484	178	9385	0.362	ng/u1	99
19) Fluoranthene	19.417	202	12003	0.327	ng/u1	96
20) Pyrene	19.779	202	12476	0.338	ng/u1	99
21) Benzo(a)anthracene	21.538	228	11270	0.375	ng/u1	99
22) Chrysene	21.591	228	11241	0.353	ng/u1	99
24) Benzo(b)fluoranthene	23.248	252	10675	0.327	ng/u1	95
25) Benzo(k)fluoranthene	23.298	252	10345	0.326	ng/u1	96
26) Benzo(a)pyrene	23.891	252	9018	0.346	ng/u1	94
27) Indeno(1,2,3-cd)pyrene	26.568	276	10633	0.328	ng/u1	95
28) Dibenzo(a,h)anthracene	26.585	278	8567	0.331	ng/u1	96
29) Benzo(g,h,i)perylene	27.356	276	8655	0.328	ng/u1	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022552.D
Acq On : 08 Nov 2022 07:34
Operator : CG/JU
Sample : PB148827BS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS827

Quant Time: Nov 08 08:04:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

