

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022562.D
 Acq On : 08 Nov 2022 16:00
 Operator : CG/JU
 Sample : N5407-12
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Z1

Quant Time: Nov 09 01:22:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1580	0.400	ng/u1	0.00
4) Naphthalene-d8	10.739	136	5740	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.591	164	3618	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.345	188	7959	0.400	ng/u1	0.00
17) Chrysene-d12	21.553	240	6455	0.400	ng/u1	0.00
23) Perylene-d12	23.996	264	5394	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.219	96	9104	4.434	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.334	152	2665	0.307	ng/u1	-0.01
18) Fluoranthene-d10	19.384	212	8003	0.369	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022562.D
Acq On : 08 Nov 2022 16:00
Operator : CG/JU
Sample : N5407-12
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Z1

Quant Time: Nov 09 01:22:18 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

