

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022544.D
 Acq On : 08 Nov 2022 02:38
 Operator : CG/JU
 Sample : N5407-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y7

Quant Time: Nov 08 03:28:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	2170	0.400	ng/u1	0.00
4) Naphthalene-d8	10.739	136	7875	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.591	164	5971	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.349	188	12678	0.400	ng/u1	0.00
17) Chrysene-d12	21.553	240	10804	0.400	ng/u1	0.00
23) Perylene-d12	23.999	264	9013	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.224	96	11861	4.206	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.340	152	5424	0.455	ng/u1	0.00
18) Fluoranthene-d10	19.384	212	14381	0.396	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022544.D
Acq On : 08 Nov 2022 02:38
Operator : CG/JU
Sample : N5407-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y7

Quant Time: Nov 08 03:28:04 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

