

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022547.D
 Acq On : 08 Nov 2022 04:29
 Operator : CG/JU
 Sample : N5408-08
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y8

Quant Time: Nov 08 05:44:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1836	0.400	ng/ul	0.00
4) Naphthalene-d8	10.738	136	6464	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.590	164	4135	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.348	188	9225	0.400	ng/ul	0.00
17) Chrysene-d12	21.553	240	7484	0.400	ng/ul	0.00
23) Perylene-d12	24.000	264	6285	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.224	96	10468	4.387	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.338	152	3797	0.388	ng/ul	0.00
18) Fluoranthene-d10	19.388	212	10881	0.433	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.261	88	407	0.166	ng/ul#	73


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022547.D
Acq On : 08 Nov 2022 04:29
Operator : CG/JU
Sample : N5408-08
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y8


Quant Time: Nov 08 05:44:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.166 ng/ul
 RT: 3.261 min Scan# 60
 Delta R.T. 0.000 min
 Lab File: BN022547.D
 Acq: 08 Nov 2022 04:29

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y8



Tgt Ion: 88 Resp: 407

Ion	Ratio	Lower	Upper
88	100		
43	63.5	27.4	41.2#
58	49.9	48.2	72.2

