

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
 Data File : BN022560.D
 Acq On : 08 Nov 2022 14:44
 Operator : CG/JU
 Sample : N5407-06
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4Y6

Quant Time: Nov 09 01:22:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 23:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	1557	0.400	ng/u1	0.00
4) Naphthalene-d8	10.739	136	5292	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.591	164	3515	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.345	188	7706	0.400	ng/u1	0.00
17) Chrysene-d12	21.553	240	6439	0.400	ng/u1	0.00
23) Perylene-d12	23.996	264	5296	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.219	96	4959	2.451	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.339	152	2310	0.288	ng/u1	0.00
18) Fluoranthene-d10	19.384	212	5895	0.273	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110722\
Data File : BN022560.D
Acq On : 08 Nov 2022 14:44
Operator : CG/JU
Sample : N5407-06
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4Y6

Quant Time: Nov 09 01:22:00 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 23:39:33 2022
Response via : Initial Calibration

