

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110724\
 Data File : BN034896.D
 Acq On : 07 Nov 2024 16:58
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Nov 07 17:29:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 15:02:36 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	5700	0.400	ng	0.00
7) Naphthalene-d8	10.340	136	16694	0.400	ng	0.00
13) Acenaphthene-d10	14.211	164	8018	0.400	ng	0.00
19) Phenanthrene-d10	16.957	188	15858	0.400	ng	0.00
29) Chrysene-d12	21.151	240	9873	0.400	ng	0.00
35) Perylene-d12	23.317	264	8029	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.191	112	5771	0.363	ng	0.00
5) Phenol-d6	6.751	99	7419	0.352	ng	0.00
8) Nitrobenzene-d5	8.706	82	4642	0.357	ng	0.00
11) 2-Methylnaphthalene-d10	11.935	152	8263	0.363	ng	0.00
14) 2,4,6-Tribromophenol	15.703	330	737	0.354	ng	0.00
15) 2-Fluorobiphenyl	12.832	172	12014	0.355	ng	0.00
27) Fluoranthene-d10	18.987	212	13428	0.376	ng	0.00
31) Terphenyl-d14	19.600	244	7180	0.388	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.184	88	2683	0.373	ng	99
3) n-Nitrosodimethylamine	3.480	42	3658	0.376	ng	97
6) bis(2-Chloroethyl)ether	7.011	93	6667	0.366	ng	100
9) Naphthalene	10.383	128	17260	0.373	ng	100
10) Hexachlorobutadiene	10.682	225	2796	0.379	ng	# 99
12) 2-Methylnaphthalene	12.010	142	10259	0.362	ng	99
16) Acenaphthylene	13.923	152	13313	0.344	ng	100
17) Acenaphthene	14.265	154	9340	0.349	ng	100
18) Fluorene	15.259	166	11939	0.358	ng	100
20) 4,6-Dinitro-2-methylph...	15.345	198	498	0.358	ng	# 63
21) 4-Bromophenyl-phenylether	16.163	248	3084	0.365	ng	99
22) Hexachlorobenzene	16.262	284	3901	0.383	ng	98
23) Atrazine	16.436	200	2146	0.350	ng	99
24) Pentachlorophenol	16.610	266	911	0.370	ng	96
25) Phenanthrene	16.994	178	18315	0.377	ng	100
26) Anthracene	17.081	178	15397	0.367	ng	100
28) Fluoranthene	19.020	202	19441	0.380	ng	100
30) Pyrene	19.382	202	19401	0.388	ng	100
32) Benzo(a)anthracene	21.133	228	14648	0.381	ng	99
33) Chrysene	21.178	228	16118	0.396	ng	97
34) Bis(2-ethylhexyl)phtha...	21.089	149	7086	0.321	ng	99
36) Indeno(1,2,3-cd)pyrene	25.472	276	12724	0.356	ng	99
37) Benzo(b)fluoranthene	22.671	252	14283	0.405	ng	99
38) Benzo(k)fluoranthene	22.712	252	14541	0.396	ng	98
39) Benzo(a)pyrene	23.221	252	10527	0.376	ng	97
40) Dibenzo(a,h)anthracene	25.490	278	9662	0.349	ng	99
41) Benzo(g,h,i)perylene	26.124	276	10828	0.369	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110724\
Data File : BN034896.D
Acq On : 07 Nov 2024 16:58
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Nov 07 17:29:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 07 15:02:36 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.575 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN034896.D
 Acq: 07 Nov 2024 16:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 5700
 Ion Ratio Lower Upper
 152 100
 150 154.8 124.4 186.6
 115 62.8 50.5 75.7



#2
 1,4-Dioxane
 Concen: 0.373 ng
 RT: 3.184 min Scan# 42
 Delta R.T. -0.000 min
 Lab File: BN034896.D
 Acq: 07 Nov 2024 16:58

Tgt Ion: 88 Resp: 2683
 Ion Ratio Lower Upper
 88 100
 43 37.4 28.2 42.2
 58 83.9 67.1 100.7









#3
n-Nitrosodimethylamine
Concen: 0.376 ng
RT: 3.480 min Scan# 81
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 3658
Ion Ratio Lower Upper
42 100
74 101.3 83.4 125.2
44 9.1 8.6 12.8



#4
2-Fluorophenol
Concen: 0.363 ng
RT: 5.191 min Scan# 320
Delta R.T. -0.007 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58


Tgt Ion: 112 Resp: 5771
Ion Ratio Lower Upper
112 100
64 62.3 49.6 74.4
63 32.8 26.3 39.5








#5
Phenol-d6
Concen: 0.352 ng
RT: 6.751 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:	99	Resp:	7419
Ion	Ratio	Lower	Upper
99	100		
42	26.1	20.2	30.2
71	31.1	25.4	38.0



#6
bis(2-Chloroethyl)ether
Concen: 0.366 ng
RT: 7.011 min Scan# 572
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:	93	Resp:	6667
Ion	Ratio	Lower	Upper
93	100		
63	84.7	67.5	101.3
95	32.5	25.7	38.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.340 min Scan# 94
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	16694
Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.9	13.3
54	8.4	6.9	10.3
68	6.5	5.4	8.0



#8
Nitrobenzene-d5
Concen: 0.357 ng
RT: 8.706 min Scan# 789
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:	82	Resp:	4642
Ion	Ratio	Lower	Upper
82	100		
128	35.3	28.1	42.1
54	60.1	49.8	74.6









#9
Naphthalene
Concen: 0.373 ng
RT: 10.383 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	17260
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.4
127	13.6	10.8	16.2



#10
Hexachlorobutadiene
Concen: 0.379 ng
RT: 10.682 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:	225	Resp:	2796
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	52.0	78.0









#11
2-Methylnaphthalene-d10
Concen: 0.363 ng
RT: 11.935 min Scan# 1165
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 8263
Ion Ratio Lower Upper
152 100
151 21.6 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.362 ng
RT: 12.010 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:142 Resp: 10259
Ion Ratio Lower Upper
142 100
141 89.2 70.5 105.7
115 36.8 29.4 44.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.211 min Scan# 1493
Delta R.T. 0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 8018
Ion Ratio Lower Upper
164 100
162 101.8 81.9 122.9
160 54.3 43.5 65.3



#14
2,4,6-Tribromophenol
Concen: 0.354 ng
RT: 15.703 min Scan# 1630
Delta R.T. -0.001 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:330 Resp: 737
Ion Ratio Lower Upper
330 100
332 95.9 77.1 115.7
141 63.1 54.1 81.1









#15
2-Fluorobiphenyl
Concen: 0.355 ng
RT: 12.832 min Scan# 1364
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 12014
Ion Ratio Lower Upper
172 100
171 33.5 27.9 41.9
170 23.9 19.0 28.4



#16
Acenaphthylene
Concen: 0.344 ng
RT: 13.923 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:152 Resp: 13313
Ion Ratio Lower Upper
152 100
151 18.9 15.2 22.8
153 12.8 10.4 15.6





#17

Acenaphthene

Concen: 0.349 ng

RT: 14.265 min Scan# 1498

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:154 Resp: 9340

Ion Ratio Lower Upper

154 100

153 114.8 92.2 138.2

152 64.2 51.1 76.7



#18

Fluorene

Concen: 0.358 ng

RT: 15.259 min Scan# 1591

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58




Tgt Ion:166 Resp: 11939


Ion Ratio Lower Upper

166 100

165 99.7 79.5 119.3

167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.957 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 15858
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.4 8.6 12.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.358 ng
RT: 15.345 min Scan# 1599
Delta R.T. 0.011 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:198 Resp: 498
Ion Ratio Lower Upper
198 100
51 124.5 141.8 212.8#
105 59.6 75.6 113.4#









#21
4-Bromophenyl-phenylether
Concen: 0.365 ng
RT: 16.163 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3084
Ion Ratio Lower Upper
248 100
250 101.4 82.2 123.4
141 44.8 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.383 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:284 Resp: 3901
Ion Ratio Lower Upper
284 100
142 52.8 43.4 65.2
249 33.8 25.8 38.6





#23

Atrazine

Concen: 0.350 ng

RT: 16.436 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	200	Resp:	2146
Ion	Ratio	Lower	Upper
200	100		
173	29.4	23.4	35.2
215	44.9	35.4	53.0



#24

Pentachlorophenol

Concen: 0.370 ng




RT: 16.610 min Scan# 1703


Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58

Tgt Ion:	266	Resp:	911
Ion	Ratio	Lower	Upper
266	100		
264	61.0	51.3	76.9
268	62.7	53.0	79.6









#25
Phenanthrene
Concen: 0.377 ng
RT: 16.994 min Scan# 1734
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 18315
Ion Ratio Lower Upper
178 100
176 19.5 15.5 23.3
179 15.4 12.2 18.2



#26
Anthracene
Concen: 0.367 ng
RT: 17.081 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:178 Resp: 15397
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.2 12.1 18.1









#27
Fluoranthene-d10
Concen: 0.376 ng
RT: 18.987 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
212	100		
106	23.2	18.2	27.4
104	13.3	10.6	15.8



#28
Fluoranthene
Concen: 0.380 ng
RT: 19.020 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion	Ratio	Lower	Upper
202	100		
101	16.0	12.7	19.1
203	16.9	13.7	20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.151 min Scan# 2410
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 9873
Ion Ratio Lower Upper
240 100
120 15.8 13.8 20.8
236 28.8 23.8 35.6



#30
Pyrene
Concen: 0.388 ng
RT: 19.382 min Scan# 2044
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:202 Resp: 19401
Ion Ratio Lower Upper
202 100
200 21.0 16.8 25.2
203 17.4 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.388 ng
 RT: 19.600 min Scan# 2091
 Delta R.T. -0.000 min
 Lab File: BN034896.D
 Acq: 07 Nov 2024 16:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 7180
 Ion Ratio Lower Upper
 244 100
 212 11.8 9.4 14.0
 122 29.5 23.0 34.4



#32
 Benzo(a)anthracene
 Concen: 0.381 ng
 RT: 21.133 min Scan# 2408
 Delta R.T. -0.000 min
 Lab File: BN034896.D
 Acq: 07 Nov 2024 16:58

Tgt Ion:228 Resp: 14648
 Ion Ratio Lower Upper
 228 100
 226 27.3 22.2 33.2
 229 19.7 16.0 24.0





#33

Chrysene

Concen: 0.396 ng

RT: 21.178 min Scan# 2413

Delta R.T. -0.009 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 16118

Ion Ratio Lower Upper

228 100

226 31.2 23.7 35.5

229 19.1 16.3 24.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.321 ng

RT: 21.089 min Scan# 2403

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58




Tgt Ion:149 Resp: 7086


Ion Ratio Lower Upper

149 100

167 23.0 18.1 27.1

279 1.7 1.2 1.8





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.317 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:264 Resp: 8029

Ion Ratio Lower Upper

264 100

260 27.3 22.2 33.2

265 65.2 60.9 91.3



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.356 ng

RT: 25.472 min Scan# 3599

Delta R.T. -0.000 min

Lab File: BN034896.D

Acq: 07 Nov 2024 16:58




Tgt Ion:276 Resp: 12724


Ion Ratio Lower Upper

276 100

138 47.4 38.4 57.6


277 24.0 19.4 29.2







#37
Benzo(b)fluoranthene
Concen: 0.405 ng
RT: 22.671 min Scan# 2641
Delta R.T. 0.003 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 14283
Ion Ratio Lower Upper
252 100
253 24.5 19.4 29.2
125 25.7 21.4 32.2



#38
Benzo(k)fluoranthene
Concen: 0.396 ng
RT: 22.712 min Scan# 2655
Delta R.T. -0.000 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:252 Resp: 14541
Ion Ratio Lower Upper
252 100
253 24.3 19.8 29.8
125 26.2 22.6 33.8









#39
Benzo(a)pyrene
Concen: 0.376 ng
RT: 23.221 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 10527
Ion Ratio Lower Upper
252 100
253 25.8 21.4 32.2
125 32.2 27.8 41.6



#40
Dibenzo(a,h)anthracene
Concen: 0.349 ng
RT: 25.490 min Scan# 3605
Delta R.T. 0.003 min
Lab File: BN034896.D
Acq: 07 Nov 2024 16:58

Tgt Ion:278 Resp: 9662
Ion Ratio Lower Upper
278 100
139 34.3 27.2 40.8
279 26.4 21.4 32.0





#41
 Benzo(g,h,i)perylene
 Concen: 0.369 ng
 RT: 26.124 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN034896.D
 Acq: 07 Nov 2024 16:58

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	10828
Ion Ratio	Lower	Upper	
276	100		
277	25.0	20.2	30.2
138	43.1	33.9	50.9

