




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110921\
Data File : BN017359.D
Acq On : 09 Nov 2021 21:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 17:10:53 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 01 19:02:51 2021
Response via : Initial Calibration






TIC: BN017359.D\data.ms

(70) Pentachlorophenol (C)**16.563min (-0.012) 25302.55 ng****response 449558**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.50
264.00	61.60	64.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110921\
Data File : BN017359.D
Acq On : 09 Nov 2021 21:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 17:10:53 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 01 19:02:51 2021
Response via : Initial Calibration



TIC: BN017359.D\data.ms

(77) Benzidine**19.186min (-0.012) 0.00 ng****response 2931838**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.87
183.00	13.20	11.92
0.00	0.00	0.00