

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110921\
 Data File : BN017348.D
 Acq On : 09 Nov 2021 14:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN110921

Quant Time: Nov 09 15:17:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 09 14:15:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.224	0.238	-6.2	104	0.00
3	n-Nitrosodimethylamine	0.344	0.351	-2.0	101	0.00
4 S	2-Fluorophenol	0.922	0.957	-3.8	103	0.00
5 S	Phenol-d6	1.016	1.053	-3.6	104	0.00
6	bis(2-Chloroethyl)ether	1.070	1.131	-5.7	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.01
8 S	Nitrobenzene-d5	0.240	0.250	-4.2	109	0.01
9	Naphthalene	1.054	1.085	-2.9	99	0.00
10	Hexachlorobutadiene	0.207	0.212	-2.4	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.589	0.597	-1.4	99	0.00
12	2-Methylnaphthalene	0.694	0.711	-2.4	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.149	0.127	14.8	97	0.00
15 S	2-Fluorobiphenyl	1.522	1.483	2.6	90	0.01
16	Acenaphthylene	1.631	1.663	-2.0	99	0.00
17	Acenaphthene	1.123	1.168	-4.0	98	0.00
18	Fluorene	1.341	1.393	-3.9	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4,6-Dinitro-2-methylphenol	0.041	0.024	41.5#	86	0.00
21	4-Bromophenyl-phenylether	0.224	0.233	-4.0	98	0.01
22	Hexachlorobenzene	0.251	0.265	-5.6	97	0.00
23	Atrazine	0.151	0.144	4.6	99	0.00
24	Pentachlorophenol	0.089	0.063	29.2#	86	0.00
25	Phenanthrene	1.146	1.196	-4.4	96	0.00
26	Anthracene	0.955	0.951	0.4	97	0.00
27 SURR	Fluoranthene-d10	1.020	1.041	-2.1	96	0.00
28	Fluoranthene	1.236	1.311	-6.1	97	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	97	0.01
30	Pyrene	1.292	1.318	-2.0	96	0.00
31 S	Terphenyl-d14	0.891	0.935	-4.9	96	0.00
32	Benzo(a)anthracene	1.214	1.184	2.5	95	0.00
33	Chrysene	1.292	1.324	-2.5	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.428	0.390	8.9	94	0.00
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	1.400	1.321	5.6	92	0.00
37	Benzo(b)fluoranthene	1.308	1.370	-4.7	94	0.00
38	Benzo(k)fluoranthene	1.308	1.385	-5.9	99	0.00
39 C	Benzo(a)pyrene	1.158	1.191	-2.8	100	0.00
40	Dibenzo(a,h)anthracene	1.126	1.065	5.4	92	0.00
41	Benzo(g,h,i)perylene	1.215	1.241	-2.1	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN110921