

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110921\
 Data File : BN017358.D
 Acq On : 09 Nov 2021 20:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 10 11:21:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 10 11:00:44 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.224	0.250	-11.6	125	0.00
3	n-Nitrosodimethylamine	0.344	0.391	-13.7	130	0.00
4 S	2-Fluorophenol	0.922	0.988	-7.2	122	0.00
5 S	Phenol-d6	1.016	1.104	-8.7	125	0.00
6	bis(2-Chloroethyl)ether	1.070	1.153	-7.8	118	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.01
8 S	Nitrobenzene-d5	0.240	0.268	-11.7	134	0.01
9	Naphthalene	1.054	1.093	-3.7	114	0.00
10	Hexachlorobutadiene	0.207	0.211	-1.9	114	0.00
11 SURR	2-Methylnaphthalene-d10	0.589	0.599	-1.7	114	0.00
12	2-Methylnaphthalene	0.694	0.716	-3.2	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.149	0.144	3.4	128	0.00
15 S	2-Fluorobiphenyl	1.522	1.610	-5.8	114	0.01
16	Acenaphthylene	1.631	1.703	-4.4	118	0.00
17	Acenaphthene	1.123	1.168	-4.0	115	0.00
18	Fluorene	1.341	1.405	-4.8	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.041	0.014	65.9#	62	0.00
21	4-Bromophenyl-phenylether	0.224	0.232	-3.6	116	0.01
22	Hexachlorobenzene	0.251	0.260	-3.6	113	0.00
23	Atrazine	0.151	0.158	-4.6	128	0.00
24	Pentachlorophenol	0.089	0.076	14.6	123	0.00
25	Phenanthrene	1.146	1.199	-4.6	114	0.00
26	Anthracene	0.955	0.992	-3.9	120	0.00
27 SURR	Fluoranthene-d10	1.020	1.067	-4.6	117	0.00
28	Fluoranthene	1.236	1.329	-7.5	116	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	115	0.01
30	Pyrene	1.292	1.348	-4.3	116	0.00
31 S	Terphenyl-d14	0.891	0.948	-6.4	116	0.00
32	Benzo(a)anthracene	1.214	1.263	-4.0	120	0.00
33	Chrysene	1.292	1.343	-3.9	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.428	0.499	-16.6	143	0.00
35 I	Perylene-d12	1.000	1.000	0.0	124	0.00
36	Indeno(1,2,3-cd)pyrene	1.400	1.402	-0.1	121	0.00
37	Benzo(b)fluoranthene	1.308	1.365	-4.4	115	0.00
38	Benzo(k)fluoranthene	1.308	1.366	-4.4	121	0.00
39 C	Benzo(a)pyrene	1.158	1.202	-3.8	124	0.00
40	Dibenzo(a,h)anthracene	1.126	1.127	-0.1	121	0.00
41	Benzo(g,h,i)perylene	1.215	1.186	2.4	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4