




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110922\
Data File : BN022571.D
Acq On : 09 Nov 2022 09:47
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 07:57:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 07:57:47 2022
Response via : Initial Calibration



TIC: BN022571.D\data.ms

(70) Pentachlorophenol (C)




16.998min (0.000) 21639.13 ng

response 189857

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.85
264.00	61.60	60.50
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110922\
Data File : BN022571.D
Acq On : 09 Nov 2022 09:47
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 07:57:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 07:57:47 2022
Response via : Initial Calibration



TIC: BN022571.D\data.ms

(77) Benzidine

19.610min (0.000) 0.00 ng

response 951680

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.70
183.00	13.20	11.49
0.00	0.00	0.00