

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110924\
 Data File : BN034931.D
 Acq On : 09 Nov 2024 08:40
 Operator : RC/JU
 Sample : P4744-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCP60

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/10/2024
 Supervised By :mohammad ahmed 11/12/2024

Quant Time: Nov 09 22:51:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 09 00:09:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.571	152	2520	0.400	ng/ul	0.00
4) Naphthalene-d8	10.333	136	6786	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.203	164	5393	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.954	188	13286m	0.400	ng/ul	0.00
17) Chrysene-d12	21.145	240	13571	0.400	ng/ul	0.00
23) Perylene-d12	23.314	264	13007m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.149	96	12217	5.095	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.934	152	4081	0.393	ng/ul	0.00
18) Fluoranthene-d10	18.986	212	16278	0.467	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.182	88	237	0.090	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110924\
Data File : BN034931.D
Acq On : 09 Nov 2024 08:40
Operator : RC/JU
Sample : P4744-06
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
GCP60

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/10/2024
Supervised By :mohammad ahmed 11/12/2024


Quant Time: Nov 09 22:51:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 09 00:09:12 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.090 ng/ul
 RT: 3.182 min Scan# 71
 Delta R.T. -0.004 min
 Lab File: BN034931.D
 Acq: 09 Nov 2024 08:40

Instrument :
 BNA_N
 ClientSampleId :
 GCP60



Tgt Ion: 88 Resp: 23
 Ion Ratio Lower Upper
 88 100
 43 53.2 23.3 34.9
 58 42.9 35.4 53.2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/10/2024
 Supervised By :mohammad ahmed 11/12/2024

