

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110924\
 Data File : BN034937.D
 Acq On : 09 Nov 2024 12:15
 Operator : RC/JU
 Sample : P4744-11
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCP65

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/10/2024
 Supervised By :mohammad ahmed 11/12/2024

Quant Time: Nov 09 22:52:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 09 00:09:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.571	152	2601	0.400	ng/u1	0.00
4) Naphthalene-d8	10.334	136	6866	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.208	164	5486	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.954	188	13911m	0.400	ng/u1	0.00
17) Chrysene-d12	21.145	240	14713	0.400	ng/u1	0.00
23) Perylene-d12	23.311	264	13595m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.149	96	12396	5.009	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.928	152	4126	0.392	ng/u1	0.00
18) Fluoranthene-d10	18.981	212	17541	0.464	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110924\
Data File : BN034937.D
Acq On : 09 Nov 2024 12:15
Operator : RC/JU
Sample : P4744-11
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCP65

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/10/2024
Supervised By :mohammad ahmed 11/12/2024

Quant Time: Nov 09 22:52:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 09 00:09:12 2024
Response via : Initial Calibration

