


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110924\
Data File : BN034912.D
Acq On : 08 Nov 2024 20:01
Operator : RC/JU
Sample : DFTPP239
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP239

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Nov 09 00:09:12 2024



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.5	37579	PASS
68	69	0.00	2	1.7	879	PASS
69	198	0.00	100	39.1	51592	PASS
70	69	0.00	2	0.7	338	PASS
127	198	10	80	39.0	51469	PASS
197	198	0.00	2	0.7	883	PASS
198	198	100	100	100.0	131973	PASS
199	198	5	9	6.9	9102	PASS
275	198	10	60	30.9	40736	PASS
365	198	1	100	5.8	7712	PASS
441	198	0.01	100	8.9	11783	PASS
442	198	50	100	57.6	76013	PASS
443	442	15	24	18.2	13829	PASS

m/z	Abundance
44.00	226.0
206.90	802.0
281.00	197.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

m/z	Abundance
36.90	217.0
38.00	550.0
39.00	3576.0
40.00	296.0
43.90	205.0
44.90	160.0
50.00	14199.0
51.00	48624.0
52.00	2565.0
54.80	290.0
55.90	1424.0
57.00	2951.0
57.90	174.0
61.00	676.0
62.00	980.0
63.00	2478.0
63.90	326.0
65.00	912.0
68.00	1174.0
69.00	65120.0
70.00	477.0
73.00	1001.0
74.00	9328.0
75.00	12194.0
76.00	4380.0
77.00	72944.0
78.00	5110.0
79.00	5977.0
80.00	5022.0
81.00	5486.0
82.00	1516.0
82.90	834.0
83.90	175.0
85.00	885.0
86.00	1908.0
87.00	889.0
88.00	181.0
90.90	1555.0
91.90	1362.0
93.00	12741.0
94.00	643.0
94.80	151.0
96.00	667.0
96.90	239.0
97.90	9998.0
98.90	7046.0
100.00	560.0
100.90	3067.0
101.90	208.0
103.00	679.0
104.00	2190.0
104.90	2123.0
105.90	678.0
107.00	25800.0
108.00	3410.0
108.90	825.0
110.00	33192.0
111.00	5827.0
111.90	772.0
112.90	366.0
116.00	1482.0
116.90	37360.0
118.00	2907.0
118.80	194.0
119.80	280.0
121.90	1958.0
123.00	2670.0
123.90	1395.0
125.00	946.0
127.00	62848.0
128.00	5256.0
129.00	29832.0
129.90	2247.0
130.90	542.0
132.10	246.0
133.90	930.0
134.90	2461.0
136.00	1017.0
137.00	1692.0
137.80	292.0
138.80	205.0
140.00	529.0
141.00	4503.0
141.90	1379.0
142.80	935.0
144.00	207.0
145.20	224.0
146.00	1154.0
146.90	2448.0
147.90	8481.0
148.90	1450.0
150.10	294.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

150.90	492.0
151.90	378.0
152.90	1842.0
153.90	1272.0
155.00	3153.0
156.00	3952.0
156.90	691.0
157.90	832.0
158.90	725.0
159.90	1593.0
161.00	2345.0
162.00	699.0
163.00	228.0
165.00	1889.0
166.00	1847.0
166.90	13116.0
168.00	6980.0
169.00	997.0
169.90	306.0
170.80	347.0
171.90	1226.0
173.00	1556.0
173.90	2050.0
175.00	3755.0
175.90	1027.0
176.90	2104.0
177.80	743.0
178.90	8554.0
179.90	5520.0
181.00	2177.0
182.00	423.0
182.90	191.0
183.80	690.0
184.90	4169.0
186.00	26136.0
187.00	8191.0
188.00	980.0
188.90	2300.0
189.80	387.0
191.00	1149.0
191.90	2426.0
193.00	3127.0
193.90	915.0
195.00	212.0
196.00	4438.0
196.90	1070.0
197.90	151744.0
198.90	10853.0
200.00	982.0
201.30	575.0
202.90	1800.0
203.90	8143.0
205.00	12757.0
206.00	50056.0
207.00	8063.0
208.00	2069.0
208.90	781.0
209.90	962.0
210.90	2157.0
211.70	418.0
212.80	205.0
214.90	723.0
216.00	1135.0
216.90	17456.0
217.90	2130.0
218.80	222.0
219.90	158.0
221.00	3825.0
221.80	2077.0
222.90	3284.0
224.00	31496.0
225.00	7984.0
226.10	719.0
226.90	17120.0
227.90	2037.0
229.00	2985.0
230.00	342.0
231.00	1104.0
232.00	193.0
232.90	201.0
233.90	923.0
234.80	1137.0
235.90	832.0
236.80	884.0
237.90	219.0
238.90	534.0
240.00	436.0
240.80	817.0
241.90	1392.0
243.00	1647.0
244.00	20488.0
245.00	2931.0
245.90	6742.0
246.90	1152.0
247.80	336.0
248.90	497.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

250.00	198.0
250.80	195.0
251.80	160.0
252.90	623.0
254.00	980.0
255.00	112136.0
256.00	16408.0
257.00	1325.0
257.90	8861.0
258.90	1213.0
260.00	272.0
260.90	185.0
263.70	303.0
264.90	3417.0
265.80	403.0
271.00	254.0
272.10	321.0
272.90	3159.0
273.90	8268.0
274.90	44344.0
276.00	5612.0
276.90	5271.0
278.00	927.0
281.20	268.0
282.90	596.0
283.90	393.0
284.90	744.0
292.00	273.0
292.90	684.0
293.80	281.0
294.90	402.0
295.90	19184.0
296.90	2504.0
297.90	155.0
300.90	204.0
302.00	221.0
303.00	1693.0
303.90	457.0
307.80	185.0
312.90	173.0
314.00	633.0
314.90	2184.0
315.90	800.0
320.90	289.0
322.00	248.0
322.90	4231.0
324.00	832.0
326.80	804.0
327.90	391.0
331.90	385.0
333.10	424.0
333.90	2807.0
335.00	713.0
340.90	506.0
341.90	173.0
345.80	1023.0
352.00	1213.0
353.00	815.0
353.90	1151.0
354.90	228.0
364.90	7137.0
365.90	1008.0
371.00	337.0
372.00	1885.0
373.00	415.0
383.00	426.0
389.90	256.0
400.90	169.0
401.90	721.0
402.90	867.0
403.90	471.0
420.90	731.0
422.00	626.0
422.90	3977.0
424.00	1316.0
441.00	9232.0
442.00	66456.0
443.00	13293.0
443.90	1047.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

m/z	Abundance
36.90	224.0
38.10	548.0
39.00	3123.0
40.10	309.0
40.90	163.0
43.90	198.0
49.10	471.0
50.00	13723.0
51.00	44536.0
52.00	1917.0
55.00	389.0
56.00	1483.0
57.00	2747.0
61.00	643.0
62.00	903.0
63.00	2408.0
63.90	231.0
65.00	1072.0
67.90	862.0
69.00	60536.0
69.90	329.0
72.90	770.0
74.00	8842.0
75.00	10541.0
76.10	4294.0
77.00	67176.0
78.00	4403.0
79.00	6214.0
80.00	4538.0
81.00	6122.0
81.90	1378.0
83.00	940.0
83.90	194.0
85.00	869.0
86.00	2094.0
86.90	969.0
88.00	277.0
89.00	157.0
90.90	1157.0
91.90	1531.0
93.00	13250.0
93.90	798.0
95.90	341.0
97.00	189.0
98.00	9578.0
99.00	6933.0
99.80	441.0
101.00	2512.0
102.90	881.0
104.00	1885.0
104.90	2290.0
106.00	752.0
107.00	24200.0
107.90	3180.0
109.00	722.0
110.00	30784.0
111.00	5768.0
111.90	941.0
113.00	252.0
116.00	1287.0
116.90	38424.0
117.90	2537.0
118.90	213.0
120.00	350.0
121.90	1816.0
123.00	2512.0
123.90	1110.0
124.90	772.0
127.00	60856.0
128.00	4798.0
128.90	29896.0
129.90	2504.0
131.00	402.0
132.00	250.0
133.90	927.0
135.00	2917.0
135.90	1156.0
136.90	1455.0
138.10	264.0
139.00	192.0
139.80	656.0
140.90	4756.0
142.00	1337.0
142.90	981.0
144.00	272.0
144.90	211.0
146.00	1046.0
147.00	2960.0
147.90	9205.0
148.90	1746.0
149.90	270.0
150.90	667.0

Instrument :

BNA_N

ClientSampleId :

DFTPP239

152.00	338.0
152.90	1811.0
153.90	1196.0
155.00	2979.0
156.00	3070.0
157.00	506.0
157.90	1096.0
158.90	865.0
160.00	1406.0
161.00	2464.0
161.90	551.0
162.70	165.0
164.90	2052.0
165.90	1473.0
166.90	13944.0
167.90	6773.0
169.00	1033.0
169.90	384.0
170.90	411.0
171.90	729.0
172.90	1409.0
174.00	2045.0
175.00	4319.0
175.90	845.0
176.90	1833.0
177.90	676.0
178.90	9347.0
179.90	5292.0
181.00	2375.0
181.90	458.0
183.80	875.0
185.00	3867.0
186.00	26968.0
187.00	8014.0
187.90	1017.0
188.90	1954.0
189.70	319.0
190.90	986.0
191.90	2986.0
193.00	3408.0
194.00	865.0
194.90	303.0
196.00	4878.0
196.90	1186.0
197.90	159168.0
198.90	10311.0
199.90	1062.0
201.50	522.0
202.90	1788.0
204.00	9073.0
205.00	14165.0
206.00	52360.0
207.00	7207.0
207.90	2668.0
209.00	927.0
209.90	1169.0
210.90	2483.0
214.90	917.0
215.90	1765.0
216.90	16928.0
217.90	2125.0
218.80	179.0
219.80	391.0
221.00	5094.0
221.80	2264.0
222.90	4082.0
224.00	34928.0
225.00	8574.0
225.90	766.0
226.90	17088.0
227.90	2663.0
228.90	3243.0
230.10	258.0
230.90	854.0
232.00	204.0
232.80	252.0
233.90	1092.0
234.90	1018.0
235.90	1029.0
236.90	1133.0
238.00	197.0
238.90	522.0
239.90	356.0
240.90	1056.0
241.90	1700.0
243.00	2000.0
244.00	22400.0
245.00	3080.0
245.90	7079.0
246.90	1338.0
247.90	287.0
248.90	668.0
249.90	160.0
250.90	188.0
251.90	374.0
252.80	694.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

253.90	736.0
255.00	125720.0
256.00	18760.0
256.90	1307.0
257.90	9203.0
258.90	1506.0
260.00	158.0
260.90	225.0
264.00	205.0
264.90	4324.0
265.90	713.0
269.80	168.0
270.80	357.0
271.90	471.0
273.00	3876.0
274.00	9792.0
275.00	49376.0
276.00	6161.0
276.90	6598.0
278.00	1014.0
278.80	163.0
282.00	158.0
282.90	517.0
284.00	430.0
285.00	810.0
289.90	216.0
291.00	184.0
291.90	268.0
292.90	1532.0
293.80	519.0
295.90	21808.0
296.90	2813.0
297.80	159.0
300.90	305.0
301.90	269.0
303.00	1878.0
303.90	736.0
308.00	225.0
309.90	235.0
313.90	750.0
314.90	2664.0
316.00	1026.0
316.90	240.0
320.80	466.0
321.90	229.0
323.00	4536.0
323.90	928.0
326.90	1107.0
327.90	350.0
332.00	609.0
332.90	535.0
333.90	3935.0
334.90	902.0
340.90	590.0
345.90	1069.0
347.00	240.0
352.00	1404.0
352.90	981.0
354.00	1400.0
354.90	279.0
364.90	9747.0
365.90	1299.0
366.80	158.0
369.90	231.0
371.10	416.0
372.00	2504.0
373.00	594.0
382.90	630.0
383.90	230.0
390.00	231.0
390.80	171.0
401.90	1087.0
402.90	1297.0
403.90	551.0
420.90	962.0
421.90	732.0
423.00	5529.0
423.90	1061.0
441.00	14811.0
442.00	93000.0
443.00	16265.0
444.00	1732.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

m/z	Abundance
37.90	344.0
39.00	1595.0
40.10	165.0
44.00	255.0
49.00	295.0
50.00	6646.0
51.00	19576.0
52.00	1188.0
55.90	748.0
56.90	1080.0
61.00	390.0
62.00	323.0
62.90	1004.0
64.00	180.0
64.90	566.0
68.00	601.0
69.00	29120.0
70.00	208.0
73.00	407.0
74.00	3804.0
75.00	5303.0
76.00	2140.0
77.00	33192.0
78.00	2334.0
78.90	3080.0
80.00	2047.0
81.00	2918.0
81.90	665.0
82.80	454.0
85.00	449.0
85.90	1115.0
87.00	334.0
87.90	162.0
90.90	572.0
92.00	900.0
92.90	5989.0
94.00	254.0
96.00	234.0
98.00	5685.0
99.00	3058.0
99.80	185.0
100.90	1486.0
102.80	466.0
103.90	1083.0
104.90	955.0
105.80	264.0
107.00	12331.0
108.00	1825.0
108.90	297.0
110.00	16632.0
111.00	2952.0
111.90	383.0
113.00	182.0
115.90	852.0
116.90	19000.0
118.00	1351.0
118.90	168.0
119.90	152.0
121.90	865.0
123.00	1611.0
124.00	714.0
124.90	499.0
127.00	30704.0
128.00	2749.0
129.00	15304.0
130.00	1258.0
130.90	165.0
131.80	156.0
134.00	672.0
134.90	1295.0
135.90	597.0
136.90	785.0
140.00	311.0
141.00	2848.0
141.90	955.0
143.00	612.0
145.90	578.0
147.00	1289.0
148.00	4394.0
148.90	729.0
149.90	157.0
151.10	275.0
152.00	201.0
152.90	853.0
153.90	500.0
154.90	1369.0
155.90	2145.0
156.90	283.0
158.00	594.0
158.90	343.0
160.00	1021.0
161.00	1042.0

Instrument :

BNA_N

ClientSampleId :

DFTPP239

162.10	339.0
162.90	188.0
164.90	1205.0
166.10	719.0
167.00	7159.0
167.90	4038.0
168.90	460.0
169.90	240.0
171.00	291.0
171.90	604.0
173.00	668.0
174.00	1081.0
175.00	2255.0
176.00	565.0
176.90	1224.0
177.80	397.0
178.90	4822.0
180.00	2841.0
180.90	1386.0
183.80	414.0
185.00	2107.0
186.00	13512.0
187.00	5030.0
187.90	459.0
188.90	1212.0
189.80	194.0
190.90	633.0
191.90	1681.0
192.90	1818.0
193.90	438.0
195.90	2785.0
196.90	392.0
197.90	85008.0
198.90	6143.0
199.90	465.0
201.40	299.0
201.80	305.0
202.90	957.0
204.00	4726.0
205.00	7200.0
206.00	28888.0
207.00	4127.0
207.90	1048.0
208.90	514.0
210.10	592.0
210.90	1293.0
211.90	203.0
214.80	443.0
215.90	1172.0
216.90	9768.0
217.90	1240.0
219.10	191.0
220.00	188.0
221.00	3367.0
221.80	1136.0
222.90	1989.0
224.00	19184.0
224.90	4524.0
226.00	463.0
226.90	11214.0
228.00	1564.0
228.90	1960.0
230.00	228.0
230.90	565.0
232.90	202.0
233.80	666.0
234.90	542.0
235.90	603.0
236.90	745.0
238.90	365.0
239.90	358.0
241.00	494.0
242.00	1232.0
243.00	1149.0
244.00	14013.0
245.00	1859.0
245.90	3789.0
246.90	487.0
248.90	499.0
251.10	186.0
253.00	413.0
253.90	651.0
254.90	73920.0
256.00	9655.0
256.90	746.0
257.90	5756.0
258.90	801.0
259.90	303.0
264.90	2897.0
265.90	360.0
270.90	244.0
271.90	325.0
272.90	2272.0
273.90	6102.0
275.00	28488.0
276.00	3705.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239

276.90	3776.0
277.90	405.0
281.00	168.0
282.90	545.0
283.80	189.0
285.00	553.0
293.00	816.0
294.90	361.0
295.90	12902.0
296.90	1496.0
300.90	202.0
301.80	160.0
303.00	1258.0
304.10	339.0
307.90	189.0
313.90	615.0
314.80	1465.0
315.90	637.0
320.90	406.0
321.90	220.0
322.90	2802.0
324.00	499.0
326.90	739.0
327.90	297.0
331.90	272.0
332.90	286.0
333.90	2040.0
335.00	534.0
340.80	301.0
345.90	867.0
346.90	222.0
351.90	1136.0
353.00	550.0
354.00	790.0
365.00	6252.0
365.90	899.0
370.80	249.0
372.00	1592.0
372.90	393.0
383.00	404.0
390.00	252.0
390.90	176.0
401.90	661.0
403.00	856.0
403.80	314.0
420.90	1006.0
421.90	663.0
423.00	4474.0
423.90	768.0
440.90	11305.0
442.00	68584.0
442.90	11929.0
443.90	1230.0

Instrument :
BNA_N
ClientSampleId :
DFTPP239