

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111120\
 Data File : BN012548.D
 Acq On : 12 Nov 2020 02:01
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 12 04:17:26 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	67	-0.02
2	1,4-Dioxane	0.400	0.332	17.0	59	-0.02
3	n-Nitrosodimethylamine	0.400	0.456	-14.0	85	-0.02
4 S	2-Fluorophenol	0.400	0.331	17.3	59	-0.02
5 S	Phenol-d6	0.400	0.338	15.5	60	-0.03
6	bis(2-Chloroethyl)ether	0.400	0.360	10.0	69	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	69	-0.03
8 S	Nitrobenzene-d5	0.400	0.389	2.8	75	-0.03
9	Naphthalene	0.400	0.373	6.8	72	-0.03
10	Hexachlorobutadiene	0.400	0.396	1.0	75	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	75	-0.03
12	2-Methylnaphthalene	0.400	0.380	5.0	74	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	74	-0.03
14 S	2,4,6-Tribromophenol	0.400	0.372	7.0	80	-0.03
15 S	2-Fluorobiphenyl	0.400	0.364	9.0	74	-0.03
16	Acenaphthylene	0.400	0.355	11.3	76	-0.03
17	Acenaphthene	0.400	0.354	11.5	75	-0.01
18	Fluorene	0.400	0.361	9.8	77	-0.03
19 I	Phenanthrene-d10	0.400	0.400	0.0	74	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.367	8.3	83	-0.01
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	84	-0.02
22	Hexachlorobenzene	0.400	0.394	1.5	81	-0.02
23	Atrazine	0.400	0.396	1.0	84	-0.02
24	Pentachlorophenol	0.400	0.416	-4.0	93	-0.01
25	Phenanthrene	0.400	0.373	6.8	78	-0.02
26	Anthracene	0.400	0.366	8.5	78	-0.02
27 SURR	Fluoranthene-d10	0.400	0.393	1.8	83	-0.02
28	Fluoranthene	0.400	0.390	2.5	83	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	85	-0.03
30	Pyrene	0.400	0.354	11.5	84	-0.02
31 S	Terphenyl-d14	0.400	0.375	6.3	88	-0.02
32	Benzo(a)anthracene	0.400	0.369	7.8	87	-0.02
33	Chrysene	0.400	0.382	4.5	90	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.375	6.3	92	-0.02
35	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	89	-0.05
36 I	Perylene-d12	0.400	0.400	0.0	87	-0.04
37	Benzo(b)fluoranthene	0.400	0.362	9.5	90	-0.03
38	Benzo(k)fluoranthene	0.400	0.380	5.0	92	-0.03
39 C	Benzo(a)pyrene	0.400	0.372	7.0	91	-0.03
40	Dibenzo(a,h)anthracene	0.400	0.392	2.0	90	-0.05
41	Benzo(g,h,i)perylene	0.400	0.381	4.8	88	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4