

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111120\
 Data File : BN012548.D
 Acq On : 12 Nov 2020 02:01
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

Quant Time: Nov 12 04:17:26 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:46:47 2020
 Response via : Initial Calibration

11/12/2020 11:38:24 AM

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.465	152	418	0.40	ng	#-0.02
7) Naphthalene-d8	10.225	136	1736	0.40	ng	#-0.03
13) Acenaphthene-d10	14.116	164	1026	0.40	ng	-0.03
19) Phenanthrene-d10	16.858	188	2111	0.40	ng	#-0.02
29) Chrysene-d12	21.067	240	2506	0.40	ng	#-0.03
36) Perylene-d12	23.186	264	3074	0.40	ng	-0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.073	112	483	0.33	ng	-0.02
5) Phenol-d6	6.644	99	578	0.34	ng	-0.03
8) Nitrobenzene-d5	8.615	82	799	0.39	ng	-0.03
11) 2-Methylnaphthalene-d10	11.829	152	1073	0.38	ng	-0.03
14) 2,4,6-Tribromophenol	15.613	330	277	0.37	ng	-0.03
15) 2-Fluorobiphenyl	12.733	172	1458	0.36	ng	-0.03
27) Fluoranthene-d10	18.903	212	2570	0.39	ng	-0.02
31) Terphenyl-d14	19.519	244	2243	0.38	ng	-0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.036	88	217	0.33	ng	# 73
3) n-Nitrosodimethylamine	3.358	42	814	0.46	ng	# 52
6) bis(2-Chloroethyl)ether	6.901	93	420	0.36	ng	# 81
9) Naphthalene	10.275	128	1817	0.37	ng	# 94
10) Hexachlorobutadiene	10.567	225	421	0.40	ng	# 98
12) 2-Methylnaphthalene	11.904	142	1193	0.38	ng	# 82
16) Acenaphthylene	13.824	152	1824	0.35	ng	100
17) Acenaphthene	14.179	154	1159	0.35	ng	88
18) Fluorene	15.169	166	1505	0.36	ng	97
20) 4,6-Dinitro-2-methylph...	15.283	198	188	0.37	ng	# 58
21) 4-Bromophenyl-phenylether	16.067	248	479	0.39	ng	# 90
22) Hexachlorobenzene	16.177	284	626	0.39	ng	# 77
23) Atrazine	16.347	200	502	0.40	ng	93
24) Pentachlorophenol	16.530	266	310	0.42	ng	94
25) Phenanthrene	16.907	178	2332	0.37	ng	97
26) Anthracene	16.992	178	2118	0.37	ng	97
28) Fluoranthene	18.933	202	2911	0.39	ng	98
30) Pyrene	19.300	202	2985	0.35	ng	99
32) Benzo(a)anthracene	21.057	228	2944	0.37	ng	99
33) Chrysene	21.110	228	3191	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.004	149	2031	0.38	ng	92
35) Indeno(1,2,3-cd)pyrene	25.274	276	5114	0.41	ng	# 93
37) Benzo(b)fluoranthene	22.563	252	3549	0.36	ng	94
38) Benzo(k)fluoranthene	22.604	252	3968m	0.38	ng	
39) Benzo(a)pyrene	23.097	252	3566	0.37	ng	# 85
40) Dibenzo(a,h)anthracene	25.287	278	4263	0.39	ng	94
41) Benzo(g,h,i)perylene	25.907	276	4273	0.38	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111120\
Data File : BN012548.D
Acq On : 12 Nov 2020 02:01
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

Quant Time: Nov 12 04:17:26 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 05:46:47 2020
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.465 min Scan# 557
 Delta R.T. -0.024 min
 Lab File: BN012548.D
 Acq: 12 Nov 2020 02:01





Tgt Ion: 152 Resp: 418
 Ion Ratio Lower Upper
 152 100
 150 145.6 116.6 174.8
 115 80.3 52.9 79.3#

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED




mohammad


11/12/2020 11:38:24 AM



#2
 1,4-Dioxane
 Concen: 0.33 ng
 RT: 3.036 min Scan# 7
 Delta R.T. -0.024 min
 Lab File: BN012548.D
 Acq: 12 Nov 2020 02:01

Tgt Ion: 88 Resp: 217
 Ion Ratio Lower Upper
 88 100
 58 79.3 59.4 89.2
 43 80.6 32.3 48.5#





#3
n-Nitrosodimethylamine
Concen: 0.46 ng
RT: 3.358 min Scan# 47
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01





Tgt Ion: 42 Resp: 814
Ion Ratio Lower Upper
42 100
74 36.0 64.4 96.6#
44 4.7 2.8 4.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


11/12/2020 11:38:24 AM



#4
2-Fluorophenol
Concen: 0.33 ng
RT: 5.073 min Scan# 260
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion: 112 Resp: 483
Ion Ratio Lower Upper
112 100
64 77.4 49.5 74.3#
63 67.5 32.0 48.0#









#5
Phenol-d6
Concen: 0.34 ng
RT: 6.644 min Scan# 455
Delta R.T. -0.032 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion	Ratio	Lower	Upper
99	100		
42	67.8	22.4	33.6#
71	75.4	37.8	56.6#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 6.901 min Scan# 487
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01


Tgt Ion	Ratio	Lower	Upper
93	100		
63	99.8	63.1	94.7#
95	36.4	26.1	39.1





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.225 min Scan# 813
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
136	100		
137	15.0	10.6	15.8
54	15.4	8.6	12.8#
68	7.3	5.7	8.5

Manual Integrations
APPROVED


mohammad





11/12/2020 11:38:24 AM




#8
Nitrobenzene-d5
Concen: 0.39 ng
RT: 8.615 min Scan# 686
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01




Tgt Ion	Ratio	Lower	Upper
82	100		
128	37.7	30.6	46.0
54	81.6	48.3	72.5#







#9
Naphthalene
Concen: 0.37 ng
RT: 10.275 min Scan# 817
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:128 Resp: 1817
Ion Ratio Lower Upper
128 100
129 15.2 9.8 14.8#
127 17.1 12.2 18.4




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.567 min Scan# 840
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:225 Resp: 421
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 65.1 51.0 76.6







#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 11.829 min Scan# 985
Delta R.T. -0.027 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:152 Resp: 1073
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







11/12/2020 11:38:24 AM




#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 11.904 min Scan# 1002
Delta R.T. -0.027 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:142 Resp: 1193
Ion Ratio Lower Upper
142 100
141 91.8 69.4 104.2
115 72.3 35.7 53.5#







#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.116 min Scan# 1264
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:164 Resp: 1026
Ion Ratio Lower Upper
164 100
162 103.4 80.6 120.8
160 49.5 39.5 59.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.613 min Scan# 1382
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01




Tgt Ion:330 Resp: 277
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 53.1 34.4 51.6#







#15
2-Fluorobiphenyl
Concen: 0.36 ng
RT: 12.733 min Scan# 1155
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:172 Resp: 1458
Ion Ratio Lower Upper
172 100
171 38.0 28.2 42.4
170 27.8 20.0 30.0




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#16
Acenaphthylene
Concen: 0.35 ng
RT: 13.824 min Scan# 1241
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01




Tgt Ion:152 Resp: 1824
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.5 10.6 16.0







#17
Acenaphthene
Concen: 0.35 ng
RT: 14.179 min Scan# 1269
Delta R.T. -0.013 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:154 Resp: 1159
Ion Ratio Lower Upper
154 100
153 119.7 83.6 125.4
152 60.4 43.9 65.9




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#18
Fluorene
Concen: 0.36 ng
RT: 15.169 min Scan# 1347
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:166 Resp: 1505
Ion Ratio Lower Upper
166 100
165 102.0 78.6 118.0
167 13.7 10.7 16.1







#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.858 min Scan# 1483
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	188	Resp:	2111
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.4	5.0	7.4#



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







11/12/2020 11:38:24 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.283 min Scan# 1356
Delta R.T. -0.013 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	198	Resp:	188
Ion Ratio	Lower	Upper	
198	100		
51	119.3	152.0	228.0#
105	69.7	80.0	120.0#









#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.067 min Scan# 1418
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01





Tgt Ion:248 Resp: 479
Ion Ratio Lower Upper
248 100
250 94.7 77.3 115.9
141 89.2 57.8 86.8#



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.177 min Scan# 1427
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:284 Resp: 626
Ion Ratio Lower Upper
284 100
142 63.6 32.0 48.0#
249 30.4 27.0 40.6




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

11/12/2020 11:38:24 AM







#23
Atrazine
Concen: 0.40 ng
RT: 16.347 min Scan# 1441
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	200	Resp:	502
Ion Ratio	Lower	Upper	
200	100		
173	32.3	22.8	34.2
215	52.5	38.1	57.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#24
Pentachlorophenol
Concen: 0.42 ng
RT: 16.530 min Scan# 1456
Delta R.T. -0.012 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	266	Resp:	310
Ion Ratio	Lower	Upper	
266	100		
264	54.5	50.9	76.3
268	63.9	51.5	77.3









#25
Phenanthrene
Concen: 0.37 ng
RT: 16.907 min Scan# 1487
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:178	Resp:	2332
Ion Ratio	Lower	Upper
178	100	
176	20.6	14.8 22.2
179	15.8	12.4 18.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#26
Anthracene
Concen: 0.37 ng
RT: 16.992 min Scan# 1494
Delta R.T. -0.024 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:178	Resp:	2118
Ion Ratio	Lower	Upper
178	100	
176	20.3	14.8 22.2
179	16.1	12.5 18.7









#27
Fluoranthene-d10
Concen: 0.39 ng
RT: 18.903 min Scan# 1757
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:212 Resp: 2570
Ion Ratio Lower Upper
212 100
106 5.3 6.8 10.2#
104 3.8 4.1 6.1#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#28
Fluoranthene
Concen: 0.39 ng
RT: 18.933 min Scan# 1763
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:202 Resp: 2911
Ion Ratio Lower Upper
202 100
101 5.9 5.8 8.6
203 16.8 13.7 20.5





#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.067 min Scan# 2079
Delta R.T. -0.032 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01





Tgt Ion:240 Resp: 2506
Ion Ratio Lower Upper
240 100
120 8.5 5.3 7.9#
236 28.3 21.8 32.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


11/12/2020 11:38:24 AM




#30
Pyrene
Concen: 0.35 ng
RT: 19.300 min Scan# 1837
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:202 Resp: 2985
Ion Ratio Lower Upper
202 100
200 21.2 16.6 24.8
203 17.3 13.8 20.8






#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.519 min Scan# 1881
Delta R.T. -0.025 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:244 Resp: 2243
Ion Ratio Lower Upper
244 100
212 12.9 7.3 10.9#
122 11.2 7.3 10.9#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#32
Benzo(a)anthracene
Concen: 0.37 ng
RT: 21.057 min Scan# 2078
Delta R.T. -0.021 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:228 Resp: 2944
Ion Ratio Lower Upper
228 100
226 27.5 22.3 33.5
229 20.8 15.7 23.5







#33
Chrysene
Concen: 0.38 ng
RT: 21.110 min Scan# 2083
Delta R.T. -0.021 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	228	Resp:	3191
Ion Ratio	Lower	Upper	
228	100		
226	30.9	24.2	36.2
229	21.0	15.7	23.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







11/12/2020 11:38:24 AM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.38 ng
RT: 21.004 min Scan# 2073
Delta R.T. -0.021 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:	149	Resp:	2031
Ion Ratio	Lower	Upper	
149	100		
167	24.3	23.2	34.8
279	5.1	3.9	5.9





#35
Indeno(1,2,3-cd)pyrene
Concen: 0.41 ng
RT: 25.274 min Scan# 3043
Delta R.T. -0.048 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01





Tgt Ion: 276 Resp: 5114
Ion Ratio Lower Upper
276 100
138 10.1 13.3 19.9#
277 24.5 20.2 30.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


11/12/2020 11:38:24 AM



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.186 min Scan# 2433
Delta R.T. -0.038 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion: 264 Resp: 3074
Ion Ratio Lower Upper
264 100
260 26.4 19.8 29.6
265 52.8 51.4 77.0





#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.563 min Scan# 2251
Delta R.T. -0.031 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01





Tgt Ion:252 Resp: 3549
Ion Ratio Lower Upper
252 100
253 28.9 20.1 30.1
125 9.6 7.2 10.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


11/12/2020 11:38:24 AM




#38
Benzo(k)fluoranthene
Concen: 0.38 ng m
RT: 22.604 min Scan# 2263
Delta R.T. -0.031 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion:252 Resp: 3968
Ion Ratio Lower Upper
252 100
253 30.5 19.8 29.8#
125 8.4 7.1 10.7







#39
Benzo(a)pyrene
Concen: 0.37 ng
RT: 23.097 min Scan# 2407
Delta R.T. -0.034 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:252 Resp: 3566
Ion Ratio Lower Upper
252 100
253 33.5 21.3 31.9#
125 18.2 8.3 12.5#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.39 ng
RT: 25.287 min Scan# 3047
Delta R.T. -0.048 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01



Tgt Ion:278 Resp: 4263
Ion Ratio Lower Upper
278 100
139 10.3 9.3 13.9
279 30.8 21.6 32.4








#41
Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 25.907 min Scan# 3228
Delta R.T. -0.055 min
Lab File: BN012548.D
Acq: 12 Nov 2020 02:01

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.2	19.8	29.8
138	11.5	11.7	17.5#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad



11/12/2020 11:38:24 AM