

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111122\
 Data File : BN022596.D
 Acq On : 10 Nov 2022 11:53
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Nov 10 23:12:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 23:09:20 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.892	152	1956	0.400	ng	0.00
7) Naphthalene-d8	10.720	136	5610	0.400	ng	0.00
13) Acenaphthene-d10	14.568	164	3370	0.400	ng	0.00
19) Phenanthrene-d10	17.325	188	6774	0.400	ng	0.00
29) Chrysene-d12	21.537	240	5873	0.400	ng	0.00
35) Perylene-d12	23.976	264	5102	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.443	112	17271	3.127	ng	0.00
5) Phenol-d6	7.083	99	22410	3.219	ng	0.00
8) Nitrobenzene-d5	9.065	82	18214	3.593	ng	-0.01
11) 2-Methylnaphthalene-d10	12.315	152	30414	3.230	ng	0.00
14) 2,4,6-Tribromophenol	16.072	330	5618	3.175	ng	0.00
15) 2-Fluorobiphenyl	13.189	172	44118	3.142	ng	-0.01
27) Fluoranthene-d10	19.369	212	61931	3.346	ng	0.00
31) Terphenyl-d14	19.968	244	59349	3.243	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.240	88	8466	2.880	ng	98
3) n-Nitrosodimethylamine	3.580	42	9154	3.001	ng	# 99
6) bis(2-Chloroethyl)ether	7.321	93	19965	3.203	ng	99
9) Naphthalene	10.763	128	53667	3.158	ng	98
10) Hexachlorobutadiene	11.051	225	9843	3.357	ng	# 100
12) 2-Methylnaphthalene	12.051	142	12534	3.099	ng	# 80
16) Acenaphthylene	14.290	152	55117	3.222	ng	100
17) Acenaphthene	14.632	154	34967	3.172	ng	99
18) Fluorene	15.626	166	47505	3.075	ng	97
20) 4,6-Dinitro-2-methylph...	15.712	198	4009	3.621	ng	# 68
21) 4-Bromophenyl-phenylether	16.518	248	13157	3.145	ng	91
22) Hexachlorobenzene	16.630	284	15151	3.169	ng	99
23) Atrazine	16.791	200	11651	3.092	ng	93
24) Pentachlorophenol	16.990	266	6514	3.606	ng	95
25) Phenanthrene	17.375	178	71759	3.169	ng	99
26) Anthracene	17.462	178	64641	3.245	ng	100
28) Fluoranthene	19.403	202	78570	3.220	ng	99
30) Pyrene	19.765	202	79406	3.131	ng	100
32) Benzo(a)anthracene	21.519	228	72107	3.053	ng	98
33) Chrysene	21.573	228	73348	3.136	ng	98
34) Bis(2-ethylhexyl)phtha...	21.439	149	41536	2.088	ng	100
36) Indeno(1,2,3-cd)pyrene	26.528	276	81008	3.298	ng	99
37) Benzo(b)fluoranthene	23.225	252	68605	3.420	ng	# 89
38) Benzo(k)fluoranthene	23.277	252	69595	3.402	ng	# 90
39) Benzo(a)pyrene	23.865	252	55807	3.248	ng	# 81
40) Dibenzo(a,h)anthracene	26.546	278	65253	3.339	ng	92
41) Benzo(g,h,i)perylene	27.315	276	67206	3.237	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111122\
Data File : BN022596.D
Acq On : 10 Nov 2022 11:53
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Nov 10 23:12:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 10 23:09:20 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.892 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN022596.D
 Acq: 10 Nov 2022 11:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:152 Resp: 1956
 Ion Ratio Lower Upper
 152 100
 150 157.6 126.4 189.6
 115 71.5 57.8 86.6



#2
 1,4-Dioxane
 Concen: 2.880 ng
 RT: 3.240 min Scan# 36
 Delta R.T. -0.007 min
 Lab File: BN022596.D
 Acq: 10 Nov 2022 11:53

Tgt Ion: 88 Resp: 8466
 Ion Ratio Lower Upper
 88 100
 43 28.5 22.0 33.0
 58 74.0 60.5 90.7









#3
n-Nitrosodimethylamine
Concen: 3.001 ng
RT: 3.580 min Scan# 84
Delta R.T. -0.007 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 9154
Ion Ratio Lower Upper
42 100
74 131.4 104.5 156.7
44 6.2 7.4 11.2#



#4
2-Fluorophenol
Concen: 3.127 ng
RT: 5.443 min Scan# 341
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion: 112 Resp: 17271
Ion Ratio Lower Upper
112 100
64 60.8 48.6 73.0
63 32.7 25.5 38.3









#5
Phenol-d6
Concen: 3.219 ng
RT: 7.083 min Scan# 501
Delta R.T. -0.007 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 22410
Ion Ratio Lower Upper
99 100
42 20.1 16.0 24.0
71 37.3 30.1 45.1



#6
bis(2-Chloroethyl)ether
Concen: 3.203 ng
RT: 7.321 min Scan# 601
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion: 93 Resp: 19965
Ion Ratio Lower Upper
93 100
63 78.9 63.4 95.0
95 32.7 27.2 40.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.720 min Scan# 90
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	5610
Ion	Ratio	Lower	Upper
136	100		
137	12.2	9.4	14.2
54	8.6	7.1	10.7
68	7.1	5.8	8.6



#8
Nitrobenzene-d5
Concen: 3.593 ng
RT: 9.065 min Scan# 809
Delta R.T. -0.011 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:	82	Resp:	18214
Ion	Ratio	Lower	Upper
82	100		
128	32.3	32.0	48.0
54	54.6	42.3	63.5





#9

Naphthalene

Concen: 3.158 ng

RT: 10.763 min Scan# 9

Delta R.T. -0.011 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:128 Resp: 53667

Ion Ratio Lower Upper

128 100

129 10.9 9.5 14.3

127 13.5 11.2 16.8



#10

Hexachlorobutadiene

Concen: 3.357 ng

RT: 11.051 min Scan# 995

Delta R.T. -0.000 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53




Tgt Ion:225 Resp: 9843


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.3 50.8 76.2





#11

2-Methylnaphthalene-d10

Concen: 3.230 ng

RT: 12.315 min Scan# 1181

Delta R.T. -0.004 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53

Instrument :

BNA_N

ClientSampleId :





SSTDICC3.2

Tgt Ion:152 Resp: 30414

Ion Ratio Lower Upper

152 100

151 20.7 16.3 24.5



#12

2-Methylnaphthalene

Concen: 3.099 ng

RT: 12.051 min Scan# 1181

Delta R.T. -0.011 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53




Tgt Ion:142 Resp: 12534


Ion Ratio Lower Upper

142 100

141 11.4 14.0 21.0#

115 3.3 11.4 17.2#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.568 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 3370
Ion Ratio Lower Upper
164 100
162 105.5 84.3 126.5
160 50.7 40.6 60.8



#14
2,4,6-Tribromophenol
Concen: 3.175 ng
RT: 16.072 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:330 Resp: 5618
Ion Ratio Lower Upper
330 100
332 98.2 77.5 116.3
141 45.7 33.3 49.9









#15
2-Fluorobiphenyl
Concen: 3.142 ng
RT: 13.189 min Scan# 1365
Delta R.T. -0.011 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 44118
Ion Ratio Lower Upper
172 100
171 35.6 28.0 42.0
170 24.1 18.6 28.0



#16
Acenaphthylene
Concen: 3.222 ng
RT: 14.290 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:152 Resp: 55117
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.0 10.2 15.4









#17
Acenaphthene
Concen: 3.172 ng
RT: 14.632 min Scan# 1500
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 34967
Ion Ratio Lower Upper
154 100
153 114.2 92.2 138.2
152 55.9 45.4 68.0



#18
Fluorene
Concen: 3.075 ng
RT: 15.626 min Scan# 1593
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:166 Resp: 47505
Ion Ratio Lower Upper
166 100
165 94.9 78.6 117.8
167 12.8 10.2 15.4





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.325 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 6774

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.4 13.1 19.7



#20

4,6-Dinitro-2-methylphenol

Concen: 3.621 ng

RT: 15.712 min Scan# 1601

Delta R.T. -0.012 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53




Tgt Ion:198 Resp: 4009


Ion Ratio Lower Upper

198 100

51 61.7 78.7 118.1#

105 62.5 68.5 102.7#









#21
4-Bromophenyl-phenylether
Concen: 3.145 ng
RT: 16.518 min Scan# 1666
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 13157
Ion Ratio Lower Upper
248 100
250 101.0 78.5 117.7
141 57.3 57.0 85.4



#22
Hexachlorobenzene
Concen: 3.169 ng
RT: 16.630 min Scan# 1675
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:284 Resp: 15151
Ion Ratio Lower Upper
284 100
142 44.4 35.7 53.5
249 32.7 27.0 40.4









#23
Atrazine
Concen: 3.092 ng
RT: 16.791 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 11651
Ion Ratio Lower Upper
200 100
173 27.9 27.4 41.2
215 43.6 36.9 55.3



#24
Pentachlorophenol
Concen: 3.606 ng
RT: 16.990 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:266 Resp: 6514
Ion Ratio Lower Upper
266 100
264 63.2 47.3 70.9
268 64.0 48.0 72.0









#25
Phenanthrene
Concen: 3.169 ng
RT: 17.375 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 71759
Ion Ratio Lower Upper
178 100
176 19.4 15.7 23.5
179 15.2 12.4 18.6



#26
Anthracene
Concen: 3.245 ng
RT: 17.462 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:178 Resp: 64641
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.1 12.2 18.2









#27
Fluoranthene-d10
Concen: 3.346 ng
RT: 19.369 min Scan# 2107
Delta R.T. -0.004 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 61931
Ion Ratio Lower Upper
212 100
106 16.5 13.0 19.6
104 9.3 7.4 11.2



#28
Fluoranthene
Concen: 3.220 ng
RT: 19.403 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:202 Resp: 78570
Ion Ratio Lower Upper
202 100
101 13.2 10.5 15.7
203 17.3 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.537 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 5873
Ion Ratio Lower Upper
240 100
120 15.6 15.0 22.4
236 29.1 23.1 34.7



#30
Pyrene
Concen: 3.131 ng
RT: 19.765 min Scan# 2201
Delta R.T. -0.004 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:202 Resp: 79406
Ion Ratio Lower Upper
202 100
200 21.2 16.8 25.2
203 17.7 14.2 21.2









#31
Terphenyl-d14
Concen: 3.243 ng
RT: 19.968 min Scan# 21
Delta R.T. -0.004 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 59349
Ion Ratio Lower Upper
244 100
212 9.3 8.0 12.0
122 0.0 0.0 0.0



#32
Benzo(a)anthracene
Concen: 3.053 ng
RT: 21.519 min Scan# 2427
Delta R.T. -0.000 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:228 Resp: 72107
Ion Ratio Lower Upper
228 100
226 27.7 23.0 34.6
229 19.4 16.3 24.5





#33

Chrysene

Concen: 3.136 ng

RT: 21.573 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 73348

Ion Ratio Lower Upper

228 100

226 30.6 25.3 37.9

229 19.2 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.088 ng

RT: 21.439 min Scan# 2418

Delta R.T. -0.000 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53




Tgt Ion:149 Resp: 41536


Ion Ratio Lower Upper

149 100

167 26.2 20.8 31.2

279 3.0 2.4 3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.976 min Scan# 3063
Delta R.T. -0.006 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:264 Resp: 5102
Ion Ratio Lower Upper
264 100
260 26.7 21.2 31.8
265 77.8 57.5 86.3



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.298 ng
RT: 26.528 min Scan# 3936
Delta R.T. -0.003 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:276 Resp: 81008
Ion Ratio Lower Upper
276 100
138 30.8 24.3 36.5
277 24.9 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 3.420 ng

RT: 23.225 min Scan# 2806

Delta R.T. -0.006 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 68605

Ion Ratio Lower Upper

252 100

253 22.4 21.5 32.3

125 13.2 15.5 23.3#



#38

Benzo(k)fluoranthene

Concen: 3.402 ng

RT: 23.277 min Scan# 2824

Delta R.T. -0.003 min

Lab File: BN022596.D

Acq: 10 Nov 2022 11:53




Tgt Ion:252 Resp: 69595


Ion Ratio Lower Upper

252 100

253 22.4 21.7 32.5

125 13.0 14.5 21.7#









#39
Benzo(a)pyrene
Concen: 3.248 ng
RT: 23.865 min Scan# 3027
Delta R.T. -0.006 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 55807
Ion Ratio Lower Upper
252 100
253 22.5 24.7 37.1#
125 14.9 21.3 31.9#



#40
Dibenzo(a,h)anthracene
Concen: 3.339 ng
RT: 26.546 min Scan# 3942
Delta R.T. -0.003 min
Lab File: BN022596.D
Acq: 10 Nov 2022 11:53

Tgt Ion:278 Resp: 65253
Ion Ratio Lower Upper
278 100
139 21.2 19.1 28.7
279 23.8 23.7 35.5





#41
 Benzo(g,h,i)perylene
 Concen: 3.237 ng
 RT: 27.315 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: BN022596.D
 Acq: 10 Nov 2022 11:53

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion: 276 Resp: 67206

Ion	Ratio	Lower	Upper
276	100		
277	23.9	20.6	31.0
138	26.5	23.0	34.4

