

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111122\
 Data File : BN022598.D
 Acq On : 10 Nov 2022 13:29
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBVN111122

Quant Time: Nov 10 23:29:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 10 23:28:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.371	7.3	90	0.00
3	n-Nitrosodimethylamine	0.400	0.366	8.5	90	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	86	0.00
5 S	Phenol-d6	0.400	0.358	10.5	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.352	12.0	84	0.00
9	Naphthalene	0.400	0.359	10.3	89	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.330	17.5	82	0.00
12	2-Methylnaphthalene	0.400	0.341	14.8	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.01
14 S	2,4,6-Tribromophenol	0.400	0.356	11.0	86	0.01
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	84	0.00
16	Acenaphthylene	0.400	0.349	12.8	89	0.01
17	Acenaphthene	0.400	0.365	8.8	88	0.01
18	Fluorene	0.400	0.365	8.8	88	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	91	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.446	-11.5	72	0.01
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	89	0.00
22	Hexachlorobenzene	0.400	0.383	4.3	91	0.00
23	Atrazine	0.400	0.351	12.3	86	0.01
24	Pentachlorophenol	0.400	0.311	22.3	82	0.01
25	Phenanthrene	0.400	0.368	8.0	88	0.00
26	Anthracene	0.400	0.363	9.3	90	0.01
27 SURR	Fluoranthene-d10	0.400	0.333	16.8	80	0.00
28	Fluoranthene	0.400	0.356	11.0	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
30	Pyrene	0.400	0.379	5.3	86	0.00
31 S	Terphenyl-d14	0.400	0.434	-8.5	92	0.00
32	Benzo(a)anthracene	0.400	0.362	9.5	83	0.00
33	Chrysene	0.400	0.373	6.8	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.363	9.3	86	0.00
35 I	Perylene-d12	0.400	0.400	0.0	86	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	81	0.02
37	Benzo(b)fluoranthene	0.400	0.392	2.0	87	0.00
38	Benzo(k)fluoranthene	0.400	0.378	5.5	87	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	88	0.00
40	Dibenzo(a,h)anthracene	0.400	0.377	5.8	86	0.01
41	Benzo(g,h,i)perylene	0.400	0.385	3.8	88	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN111122