

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111124\
 Data File : BN035055.D
 Acq On : 13 Nov 2024 03:10
 Operator : RC/JU
 Sample : P4714-06
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AG2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/13/2024
 Supervised By :mohammad ahmed 11/14/2024

Quant Time: Nov 13 03:32:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 00:41:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.558	152	3841	0.400	ng/ul	0.00
4) Naphthalene-d8	10.322	136	9346	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.194	164	6255	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.941	188	13414m	0.400	ng/ul	0.00
17) Chrysene-d12	21.137	240	12278	0.400	ng/ul	# 0.00
23) Perylene-d12	23.299	264	12254m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.149	96	15077	4.125	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.917	152	4838	0.338	ng/ul	0.00
18) Fluoranthene-d10	18.972	212	16322	0.518	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.249	166	518	0.021	ng/ul#	91
14) Pentachlorophenol	16.595	266	189	0.061	ng/ul	98
15) Phenanthrene	16.983	178	1074	0.030	ng/ul#	85
16) Anthracene	17.076	178	1004	0.030	ng/ul#	81
19) Fluoranthene	19.005	202	1414	0.034	ng/ul#	82
21) Benzo(a)anthracene	21.122	228	1375	0.031	ng/ul#	93
22) Chrysene	21.172	228	1349	0.030	ng/ul#	92
24) Benzo(b)fluoranthene	22.656	252	1574m	0.037	ng/ul	
25) Benzo(k)fluoranthene	22.700	252	1563m	0.035	ng/ul	
26) Benzo(a)pyrene	23.209	252	1413m	0.037	ng/ul	
27) Indeno(1,2,3-cd)pyrene	25.440	276	1804m	0.038	ng/ul	
28) Dibenzo(a,h)anthracene	25.463	278	1390m	0.038	ng/ul	
29) Benzo(g,h,i)perylene	26.097	276	1432m	0.040	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111124\
Data File : BN035055.D
Acq On : 13 Nov 2024 03:10
Operator : RC/JU
Sample : P4714-06
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AG2

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/13/2024
Supervised By :mohammad ahmed 11/14/2024

Quant Time: Nov 13 03:32:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 00:41:17 2024
Response via : Initial Calibration

