

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111321\
 Data File : BN017427.D
 Acq On : 14 Nov 2021 14:11
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4292

Quant Time: Nov 15 04:55:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN111321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 02:31:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.868	152	7258	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.662	136	25640	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.498	164	15067	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.243	188	31604	0.400	ng/ul	0.00
17) Chrysene-d12	21.427	240	23497	0.400	ng/ul	-0.01
23) Perylene-d12	23.800	264	15981	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.332	96	2981	0.374	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.257	152	15094	0.381	ng/ul	0.00
18) Fluoranthene-d10	19.272	212	32920	0.390	ng/ul	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.366	88	4140	0.403	ng/ul#	41
5) Naphthalene	10.711	128	28534	0.403	ng/ul	96
7) 2-Methylnaphthalene	12.328	142	19714	0.388	ng/ul	99
8) 1-Methylnaphthalene	12.548	142	19580	0.390	ng/ul	100
10) Acenaphthylene	14.220	152	26003	0.359	ng/ul	98
11) Acenaphthene	14.563	153	20943	0.405	ng/ul	100
12) Fluorene	15.548	166	24146	0.394	ng/ul	100
14) Pentachlorophenol	16.893	266	2439	0.352	ng/ul	99
15) Phenanthrene	17.285	178	39461	0.415	ng/ul	99
16) Anthracene	17.374	178	35468	0.371	ng/ul	99
19) Fluoranthene	19.300	202	43747	0.401	ng/ul	99
20) Pyrene	19.662	202	44113	0.396	ng/ul	98
21) Benzo(a)anthracene	21.409	228	34098	0.390	ng/ul	100
22) Chrysene	21.462	228	35440	0.405	ng/ul	100
24) Benzo(b)fluoranthene	23.078	252	28791	0.441	ng/ul	92
25) Benzo(k)fluoranthene	23.125	252	30424	0.452	ng/ul	93
26) Benzo(a)pyrene	23.695	252	22856	0.419	ng/ul#	91
27) Indeno(1,2,3-cd)pyrene	26.252	276	20195	0.439	ng/ul#	87
28) Dibenzo(a,h)anthracene	26.276	278	15025	0.439	ng/ul	93
29) Benzo(g,h,i)perylene	27.000	276	19069	0.453	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111321\
Data File : BN017427.D
Acq On : 14 Nov 2021 14:11
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4292

Quant Time: Nov 15 04:55:22 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN111321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 02:31:43 2021
Response via : Initial Calibration

