




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111322\
Data File : BN022624.D
Acq On : 12 Nov 2022 11:56
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 08:30:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 08:30:11 2022
Response via : Initial Calibration



TIC: BN022624.D\data.ms

(70) Pentachlorophenol (C)

16.981min (0.000) 19417.00 ng

response **192608**

Ion	Exp%	Act%
-----	------	------

265.70	100.00	100.00
--------	--------	--------




268.00	62.20	71.89
--------	-------	-------

264.00	61.60	69.32
--------	-------	-------

0.00 0.00 0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111322\
 Data File : BN022624.D
 Acq On : 12 Nov 2022 11:56
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 08:30:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 14 08:30:11 2022
 Response via : Initial Calibration



TIC: BN022624.D\data.ms

(77) Benzidine

19.598min (0.000) 0.00 ng

response 1222595

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.77
183.00	13.20	11.78
0.00	0.00	0.00