

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111322\
 Data File : BN022629.D
 Acq On : 12 Nov 2022 15:08
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Nov 14 01:38:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 14 01:35:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.891	152	1839	0.400	ng	0.00
7) Naphthalene-d8	10.709	136	5586	0.400	ng	#-0.01
13) Acenaphthene-d10	14.571	164	3443	0.400	ng	0.00
19) Phenanthrene-d10	17.329	188	7297	0.400	ng	# 0.00
29) Chrysene-d12	21.537	240	6552	0.400	ng	0.00
35) Perylene-d12	23.979	264	5815	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.429	112	8956	1.684	ng	0.00
5) Phenol-d6	7.068	99	11288	1.698	ng	0.00
8) Nitrobenzene-d5	9.065	82	9309	1.753	ng	-0.01
11) 2-Methylnaphthalene-d10	12.311	152	15090	1.657	ng	0.00
14) 2,4,6-Tribromophenol	16.075	330	3112	1.869	ng	0.00
15) 2-Fluorobiphenyl	13.192	172	22469	1.603	ng	0.00
27) Fluoranthene-d10	19.369	212	34445	1.714	ng	0.00
31) Terphenyl-d14	19.967	244	32460	1.598	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.233	88	4252	1.639	ng	# 92
3) n-Nitrosodimethylamine	3.580	42	4749	1.791	ng	# 96
6) bis(2-Chloroethyl)ether	7.314	93	9943	1.702	ng	99
9) Naphthalene	10.763	128	27583	1.673	ng	97
10) Hexachlorobutadiene	11.040	225	4856	1.589	ng	# 100
12) 2-Methylnaphthalene	12.039	142	6491	1.927	ng	# 82
16) Acenaphthylene	14.283	152	27488	1.734	ng	100
17) Acenaphthene	14.625	154	17798	1.638	ng	100
18) Fluorene	15.619	166	24850	1.689	ng	99
20) 4,6-Dinitro-2-methylph...	15.715	198	2455	2.242	ng	# 1
21) 4-Bromophenyl-phenylether	16.510	248	7090	1.648	ng	# 77
22) Hexachlorobenzene	16.621	284	7862	1.568	ng	98
23) Atrazine	16.795	200	6737	1.841	ng	# 92
24) Pentachlorophenol	16.981	266	3958	2.066	ng	99
25) Phenanthrene	17.366	178	39225	1.668	ng	99
26) Anthracene	17.465	178	34428	1.714	ng	99
28) Fluoranthene	19.398	202	43410	1.702	ng	100
30) Pyrene	19.765	202	44243	1.658	ng	100
32) Benzo(a)anthracene	21.519	228	40441	1.679	ng	98
33) Chrysene	21.573	228	40177	1.618	ng	99
34) Bis(2-ethylhexyl)phtha...	21.438	149	28823	2.073	ng	99
36) Indeno(1,2,3-cd)pyrene	26.525	276	41655	1.516	ng	99
37) Benzo(b)fluoranthene	23.227	252	37965	1.613	ng	# 89
38) Benzo(k)fluoranthene	23.277	252	38481	1.638	ng	# 89
39) Benzo(a)pyrene	23.864	252	31077	1.624	ng	# 86
40) Dibenzo(a,h)anthracene	26.548	278	33591	1.520	ng	91
41) Benzo(g,h,i)perylene	27.314	276	34407	1.480	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111322\
Data File : BN022629.D
Acq On : 12 Nov 2022 15:08
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Quant Time: Nov 14 01:38:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 01:35:37 2022
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.891 min Scan# 68
 Delta R.T. -0.001 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6





Tgt Ion:152 Resp: 1839
 Ion Ratio Lower Upper
 152 100
 150 152.9 122.2 183.4
 115 68.0 56.3 84.5




#2
 1,4-Dioxane
 Concen: 1.639 ng
 RT: 3.233 min Scan# 35
 Delta R.T. -0.007 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08



Tgt Ion: 88 Resp: 4252
 Ion Ratio Lower Upper
 88 100
 43 31.3 34.9 52.3#
 58 75.0 58.8 88.2









#3
 n-Nitrosodimethylamine
 Concen: 1.791 ng
 RT: 3.580 min Scan# 84
 Delta R.T. -0.007 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion: 42 Resp: 4749
 Ion Ratio Lower Upper
 42 100
 74 131.0 101.3 151.9
 44 6.6 7.4 11.0#



#4
 2-Fluorophenol
 Concen: 1.684 ng
 RT: 5.429 min Scan# 339
 Delta R.T. -0.007 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08

Tgt Ion: 112 Resp: 8956
 Ion Ratio Lower Upper
 112 100
 64 60.9 49.8 74.6
 63 33.1 26.0 39.0









#5
Phenol-d6
Concen: 1.698 ng
RT: 7.068 min Scan# 566
Delta R.T. -0.007 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	11288
Ion	Ratio	Lower	Upper
99	100		
42	20.4	16.9	25.3
71	37.9	30.2	45.4



#6
bis(2-Chloroethyl)ether
Concen: 1.702 ng
RT: 7.314 min Scan# 600
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:	93	Resp:	9943
Ion	Ratio	Lower	Upper
93	100		
63	79.3	62.5	93.7
95	32.6	26.1	39.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.709 min Scan# 90
Delta R.T. -0.011 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	5586
Ion Ratio	Lower	Upper	
136	100		
137	12.4	10.0	15.0
54	11.9	7.0	10.4#
68	9.2	5.3	7.9#



#8
Nitrobenzene-d5
Concen: 1.753 ng
RT: 9.065 min Scan# 809
Delta R.T. -0.011 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:	82	Resp:	9309
Ion Ratio	Lower	Upper	
82	100		
128	35.7	34.6	51.8
54	51.9	40.4	60.6









#9
Naphthalene
Concen: 1.673 ng
RT: 10.763 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	128	Resp:	27583
Ion	Ratio	Lower	Upper
128	100		
129	11.1	9.9	14.9
127	13.4	11.4	17.2



#10
Hexachlorobutadiene
Concen: 1.589 ng
RT: 11.040 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:	225	Resp:	4856
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.2	51.3	76.9









#11
2-Methylnaphthalene-d10
Concen: 1.657 ng
RT: 12.311 min Scan# 1178
Delta R.T. -0.004 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 15090
Ion Ratio Lower Upper
152 100
151 20.7 16.3 24.5



#12
2-Methylnaphthalene
Concen: 1.927 ng
RT: 12.039 min Scan# 1178
Delta R.T. -0.008 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:142 Resp: 6491
Ion Ratio Lower Upper
142 100
141 12.3 14.9 22.3#
115 4.4 11.3 16.9#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.571 min Scan# 14
Delta R.T. 0.004 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 3443
Ion Ratio Lower Upper
164 100
162 102.3 84.1 126.1
160 47.0 40.2 60.2



#14
2,4,6-Tribromophenol
Concen: 1.869 ng
RT: 16.075 min Scan# 1631
Delta R.T. 0.004 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:330 Resp: 3112
Ion Ratio Lower Upper
330 100
332 96.9 77.2 115.8
141 47.8 46.2 69.2









#15
2-Fluorobiphenyl
Concen: 1.603 ng
RT: 13.192 min Scan# 1366
Delta R.T. 0.004 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 22469
Ion Ratio Lower Upper
172 100
171 34.5 29.0 43.4
170 22.7 20.1 30.1



#16
Acenaphthylene
Concen: 1.734 ng
RT: 14.283 min Scan# 1468
Delta R.T. -0.007 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:152 Resp: 27488
Ion Ratio Lower Upper
152 100
151 19.7 15.7 23.5
153 12.9 10.3 15.5









#17
Acenaphthene
Concen: 1.638 ng
RT: 14.625 min Scan# 1500
Delta R.T. -0.007 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 17798
Ion Ratio Lower Upper
154 100
153 114.4 91.6 137.4
152 56.5 45.6 68.4



#18
Fluorene
Concen: 1.689 ng
RT: 15.619 min Scan# 1593
Delta R.T. -0.007 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:166 Resp: 24850
Ion Ratio Lower Upper
166 100
165 98.2 77.8 116.8
167 12.9 10.2 15.4





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.329 min Scan# 1732

Delta R.T. 0.004 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 7297

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 12.6 19.0#



#20

4,6-Dinitro-2-methylphenol

Concen: 2.242 ng

RT: 15.715 min Scan# 1602

Delta R.T. -0.009 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08




Tgt Ion:198 Resp: 2455


Ion Ratio Lower Upper

198 100

51 73.3 266.5 399.7#

105 52.0 64.5 96.7#





#21

4-Bromophenyl-phenylether

Concen: 1.648 ng

RT: 16.510 min Scan# 1666

Delta R.T. -0.009 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:248 Resp: 7090

Ion Ratio Lower Upper

248 100

250 92.2 79.8 119.6

141 98.7 49.8 74.8#



#22

Hexachlorobenzene

Concen: 1.568 ng

RT: 16.621 min Scan# 1675

Delta R.T. -0.009 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08




Tgt Ion:284 Resp: 7862


Ion Ratio Lower Upper

284 100

142 46.0 36.2 54.2

249 33.4 27.6 41.4









#23
 Atrazine
 Concen: 1.841 ng
 RT: 16.795 min Scan# 1688
 Delta R.T. 0.004 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 6737
 Ion Ratio Lower Upper
 200 100
 173 22.6 26.2 39.4#
 215 44.8 36.4 54.6



#24
 Pentachlorophenol
 Concen: 2.066 ng
 RT: 16.981 min Scan# 1704
 Delta R.T. -0.009 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08

Tgt Ion:266 Resp: 3958
 Ion Ratio Lower Upper
 266 100
 264 62.9 49.6 74.4
 268 64.4 50.9 76.3









#25
Phenanthrene
Concen: 1.668 ng
RT: 17.366 min Scan# 1735
Delta R.T. -0.009 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 39225
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.2 12.6 19.0



#26
Anthracene
Concen: 1.714 ng
RT: 17.465 min Scan# 1743
Delta R.T. 0.004 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:178 Resp: 34428
Ion Ratio Lower Upper
178 100
176 18.8 15.1 22.7
179 15.2 11.7 17.5









#27
Fluoranthene-d10
Concen: 1.714 ng
RT: 19.369 min Scan# 2108
Delta R.T. -0.005 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 34445
Ion Ratio Lower Upper
212 100
106 16.2 13.3 19.9
104 9.4 7.8 11.8



#28
Fluoranthene
Concen: 1.702 ng
RT: 19.398 min Scan# 2116
Delta R.T. -0.005 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:202 Resp: 43410
Ion Ratio Lower Upper
202 100
101 13.1 10.4 15.6
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.537 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240	Resp:	6552
Ion	Ratio	Lower Upper
240	100	
120	14.9	14.0 21.0
236	28.6	23.4 35.2



#30
Pyrene
Concen: 1.658 ng
RT: 19.765 min Scan# 2203
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:202	Resp:	44243
Ion	Ratio	Lower Upper
202	100	
200	21.2	17.0 25.4
203	17.9	14.6 22.0









#31
Terphenyl-d14
Concen: 1.598 ng
RT: 19.967 min Scan# 21
Delta R.T. -0.005 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 32460
Ion Ratio Lower Upper
244 100
212 9.4 8.3 12.5
122 0.0 0.0 0.0



#32
Benzo(a)anthracene
Concen: 1.679 ng
RT: 21.519 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:228 Resp: 40441
Ion Ratio Lower Upper
228 100
226 27.9 23.3 34.9
229 19.5 16.5 24.7





#33

Chrysene

Concen: 1.618 ng

RT: 21.573 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 40177

Ion Ratio Lower Upper

228 100

226 30.8 24.9 37.3

229 19.7 16.2 24.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.073 ng

RT: 21.438 min Scan# 2420

Delta R.T. -0.000 min

Lab File: BN022629.D

Acq: 12 Nov 2022 15:08




Tgt Ion:149 Resp: 28823


Ion Ratio Lower Upper

149 100

167 26.2 21.1 31.7

279 2.9 2.8 4.2









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.979 min Scan# 3066
Delta R.T. -0.003 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:264 Resp: 5815




Ion	Ratio	Lower	Upper
264	100		
260	27.7	22.0	33.0
265	71.9	60.6	90.8




#36
Indeno(1,2,3-cd)pyrene
Concen: 1.516 ng
RT: 26.525 min Scan# 3937
Delta R.T. -0.006 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:276 Resp: 41655

Ion	Ratio	Lower	Upper
276	100		
138	30.4	23.8	35.8
277	25.0	19.9	29.9









#37
Benzo(b)fluoranthene
Concen: 1.613 ng
RT: 23.227 min Scan# 2809
Delta R.T. -0.003 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 37965
Ion Ratio Lower Upper
252 100
253 23.1 22.7 34.1
125 13.5 15.2 22.8#



#38
Benzo(k)fluoranthene
Concen: 1.638 ng
RT: 23.277 min Scan# 2826
Delta R.T. -0.003 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:252 Resp: 38481
Ion Ratio Lower Upper
252 100
253 22.7 22.6 33.8
125 13.4 14.9 22.3#









#39
Benzo(a)pyrene
Concen: 1.624 ng
RT: 23.864 min Scan# 3026
Delta R.T. -0.003 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 31077
Ion Ratio Lower Upper
252 100
253 24.0 24.6 36.8#
125 16.5 19.4 29.0#



#40
Dibenzo(a,h)anthracene
Concen: 1.520 ng
RT: 26.548 min Scan# 3945
Delta R.T. -0.006 min
Lab File: BN022629.D
Acq: 12 Nov 2022 15:08

Tgt Ion:278 Resp: 33591
Ion Ratio Lower Upper
278 100
139 21.0 19.1 28.7
279 25.1 24.8 37.2





#41
 Benzo(g,h,i)perylene
 Concen: 1.480 ng
 RT: 27.314 min Scan# 41
 Delta R.T. -0.009 min
 Lab File: BN022629.D
 Acq: 12 Nov 2022 15:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	276	Resp:	34407
Ion Ratio	Lower	Upper	
276	100		
277	24.0	21.2	31.8
138	27.1	23.8	35.8

