

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111422\
 Data File : BN022641.D
 Acq On : 14 Nov 2022 18:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 15 00:40:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 15 00:38:02 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.562	0.584	-3.9	111	0.00
3	n-Nitrosodimethylamine	0.645	0.608	5.7	99	0.00
4 S	2-Fluorophenol	1.232	1.217	1.2	103	0.00
5 S	Phenol-d6	1.530	1.511	1.2	104	0.00
6	bis(2-Chloroethyl)ether	1.315	1.276	3.0	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.409	0.397	2.9	105	0.00
9	Naphthalene	1.648	1.441	12.6	102	0.00
10	Hexachlorobutadiene	0.216	0.228	-5.6	112	0.01
11 SURR	2-Methylnaphthalene-d10	0.664	0.680	-2.4	110	0.00
12	2-Methylnaphthalene	0.280	0.261	6.8	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.01
14 S	2,4,6-Tribromophenol	0.222	0.216	2.7	104	0.00
15 S	2-Fluorobiphenyl	1.606	1.668	-3.9	110	0.00
16	Acenaphthylene	1.898	1.859	2.1	110	0.01
17	Acenaphthene	1.273	1.309	-2.8	111	0.01
18	Fluorene	1.888	1.785	5.5	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.025	67.9#	37#	0.01
21	4-Bromophenyl-phenylether	0.238	0.242	-1.7	106	0.01
22	Hexachlorobenzene	0.264	0.284	-7.6	111	0.01
23	Atrazine	0.227	0.217	4.4	100	0.00
24	Pentachlorophenol	0.121	0.091	24.8	84	0.01
25	Phenanthrene	1.388	1.320	4.9	99	0.00
26	Anthracene	1.154	1.130	2.1	103	0.00
27 SURR	Fluoranthene-d10	1.162	1.140	1.9	101	0.00
28	Fluoranthene	1.459	1.449	0.7	103	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
30	Pyrene	1.675	1.741	-3.9	99	0.00
31 S	Terphenyl-d14	1.219	1.287	-5.6	101	0.00
32	Benzo(a)anthracene	1.533	1.502	2.0	94	0.00
33	Chrysene	1.518	1.600	-5.4	101	0.00
34	Bis(2-ethylhexyl)phthalate	1.121	1.022	8.8	87	0.00
35 I	Perylene-d12	1.000	1.000	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	1.706	1.776	-4.1	98	0.00
37	Benzo(b)fluoranthene	1.590	1.678	-5.5	99	0.00
38	Benzo(k)fluoranthene	1.615	1.731	-7.2	100	0.00
39 C	Benzo(a)pyrene	1.302	1.385	-6.4	100	0.00
40	Dibenzo(a,h)anthracene	1.366	1.387	-1.5	96	0.00
41	Benzo(g,h,i)perylene	1.415	1.462	-3.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4