

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022649.D
 Acq On : 15 Nov 2022 14:07
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV209

Quant Time: Nov 15 22:57:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	83446	20.000	ng/u1	0.00
20) Naphthalene-d8	10.705	136	378931	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	246260	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	508386	20.000	ng/u1	0.00
79) Chrysene-d12	21.528	240	501148	20.000	ng/u1	0.00
88) Perylene-d12	23.963	264	524361	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	16694	6.830	ng/uL	0.00
4) Pyridine-d5	3.628	84	121544	17.584	ng/u1	0.00
7) Phenol-d5	7.052	99	151733	19.252	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	101293	21.436	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	115130	19.947	ng/u1	0.00
15) 4-Methylphenol-d8	8.611	113	123077	19.781	ng/u1	0.00
21) Nitrobenzene-d5	9.058	128	59870	20.412	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	66661	20.369	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	113575	20.356	ng/u1	0.00
31) 4-Chloroaniline-d4	10.852	131	170098	20.015	ng/u1	0.00
46) Dimethylphthalate-d6	13.981	166	363141	20.274	ng/u1	0.00
49) Acenaphthylene-d8	14.251	160	403209	20.430	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	70375	19.709	ng/u1	0.00
60) Fluorene-d10	15.557	176	296673	20.195	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.693	200	62531	19.351	ng/u1	0.00
73) Anthracene-d10	17.422	188	446186	20.043	ng/u1	0.00
81) Pyrene-d10	19.728	212	500381	19.267	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.804	264	531172	20.151	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.228	88	17229	6.624	ng/uL	94
5) Pyridine	3.652	79	111778	16.310	ng/u1	98
6) Benzaldehyde	7.011	77	105444	26.444	ng/u1	97
8) Phenol	7.081	94	152686	18.601	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.299	93	118500	18.527	ng/u1	99
12) 2-Chlorophenol	7.434	128	116246	18.862	ng/u1	97
13) 2-Methylphenol	8.340	108	115523	18.908	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.411	45	97283	11.522	ng/u1	99
16) Acetophenone	8.716	105	197609	19.899	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.705	70	104543	19.752	ng/u1	99
18) 4-Methylphenol	8.675	108	127008	19.144	ng/u1	100
19) Hexachloroethane	8.958	117	49001	18.779	ng/u1	98
22) Nitrobenzene	9.099	77	152138	19.145	ng/u1	97
23) Isophorone	9.634	82	289356	19.036	ng/u1	99
25) 2-Nitrophenol	9.816	139	70678	19.643	ng/u1	98
26) 2,4-Dimethylphenol	9.887	107	148327	19.218	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.116	93	169570	19.345	ng/u1	98
29) 2,4-Dichlorophenol	10.358	162	111050	19.583	ng/u1	97
30) Naphthalene	10.752	128	383333	19.059	ng/u1	100
32) 4-Chloroaniline	10.875	127	172650	19.453	ng/u1	97
33) Hexachlorobutadiene	11.028	225	66508	19.054	ng/u1	96
34) Caprolactam	11.675	113	24862	11.557	ng/u1	99
35) 4-Chloro-3-methylphenol	12.016	107	147184	19.901	ng/u1	99
36) 2-Methylnaphthalene	12.375	142	270742	19.760	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022649.D
 Acq On : 15 Nov 2022 14:07
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SICV209


Quant Time: Nov 15 22:57:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.593	142	262604	19.119	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.746	216	124155	18.783	ng/ul	98
40) Hexachlorocyclopentadiene	12.716	237	70796	18.345	ng/ul	97
41) 2,4,6-Trichlorophenol	12.993	196	88547	18.868	ng/ul	99
42) 2,4,5-Trichlorophenol	13.069	196	97081	19.534	ng/ul	97
43) 1,1'-Biphenyl	13.393	154	352202	19.485	ng/ul	100
44) 2-Chloronaphthalene	13.434	162	268099	19.034	ng/ul	99
45) 2-Nitroaniline	13.651	65	98979	19.350	ng/ul	95
47) Dimethylphthalate	14.028	163	354685	18.840	ng/ul	98
48) 2,6-Dinitrotoluene	14.151	165	77320	20.289	ng/ul	98
50) Acenaphthylene	14.281	152	437967	19.865	ng/ul	99
51) 3-Nitroaniline	14.481	138	84684	20.638	ng/ul	99
52) Acenaphthene	14.622	153	297052	18.990	ng/ul	96
53) 2,4-Dinitrophenol	14.693	184	46204	17.348	ng/ul	99
55) 4-Nitrophenol	14.810	109	65963	18.636	ng/ul	98
56) Dibenzofuran	14.963	168	414803	20.069	ng/ul	100
57) 2,4-Dinitrotoluene	14.940	165	107271	19.329	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.193	232	82086	19.285	ng/ul	100
59) Diethylphthalate	15.387	149	376648	18.875	ng/ul	98
61) Fluorene	15.616	166	338041	19.585	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.604	204	156389	19.221	ng/ul	95
63) 4-Nitroaniline	15.651	138	86594	23.664	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.704	198	62586	18.738	ng/ul	98
67) N-Nitrosodiphenylamine	15.828	169	293921	19.310	ng/ul	99
68) 4-Bromophenyl-phenylether	16.504	248	96941	19.364	ng/ul	99
69) Hexachlorobenzene	16.616	284	111953	19.130	ng/ul	92
70) Atrazine	16.787	200	101369	18.683	ng/ul	95
71) Pentachlorophenol	16.969	266	67033	18.028	ng/ul	96
72) Phenanthrene	17.363	178	521662	19.038	ng/ul	98
74) Anthracene	17.451	178	527749	19.014	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.351	216	121943	18.189	ng/uL	96
76) Pentachlorobenzene	14.875	250	123156	17.021	ng/uL	100
77) Carbazole	17.734	167	507565	19.681	ng/ul	99
78) Di-n-butylphthalate	18.292	149	655467	17.859	ng/ul	99
80) Fluoranthene	19.392	202	602997	18.603	ng/ul	99
82) Pyrene	19.751	202	608505	18.326	ng/ul	100
83) Butylbenzylphthalate	20.657	149	303619	17.964	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.451	252	223619	20.223	ng/ul	87
85) Benzo(a)anthracene	21.510	228	615044	18.585	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.433	149	455164	18.246	ng/ul	98
87) Chrysene	21.569	228	565857	18.056	ng/ul	98
89) Di-n-octyl phthalate	22.369	149	771528	16.730	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	648199	18.241	ng/ul	98
91) Benzo(k)fluoranthene	23.263	252	608429	18.038	ng/ul	96
93) Benzo(a)pyrene	23.851	252	550333	18.271	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.510	276	671733	19.609	ng/ul	98
95) Dibenzo(a,h)anthracene	26.527	278	598378	19.489	ng/ul	99
96) Benzo(g,h,i)perylene	27.292	276	581807	19.479	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 15 22:57:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN11522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 15 22:53:43 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.624 ng/uL
RT: 3.228 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion: 88 Resp: 17229
Ion Ratio Lower Upper
88 100
43 30.9 30.8 46.2
58 74.9 61.7 92.5



#5
Pyridine
Concen: 16.310 ng/uL
RT: 3.652 min Scan# 113
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion: 79 Resp: 111778
Ion Ratio Lower Upper
79 100
52 71.9 58.6 88.0
51 33.7 28.8 43.2









#6
Benzaldehyde
Concen: 26.444 ng/ul
RT: 7.011 min Scan# 684
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion: 77 Resp: 105444
Ion Ratio Lower Upper
77 100
105 94.9 75.2 112.8
106 90.5 69.0 103.6



#8
Phenol
Concen: 18.601 ng/ul
RT: 7.081 min Scan# 696
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion: 94 Resp: 152686
Ion Ratio Lower Upper
94 100
65 30.3 24.3 36.5
66 41.4 33.5 50.3









#10
Bis(2-Chloroethyl)ether
Concen: 18.527 ng/ul
RT: 7.299 min Scan# 71
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion: 93 Resp: 118500
Ion Ratio Lower Upper
93 100
63 76.6 61.8 92.6
95 33.6 26.3 39.5



#12
2-Chlorophenol
Concen: 18.862 ng/ul
RT: 7.434 min Scan# 756
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:128 Resp: 116246
Ion Ratio Lower Upper
128 100
64 52.1 40.0 60.0
130 31.8 26.3 39.5





#13

2-Methylphenol

Concen: 18.908 ng/ul

RT: 8.340 min Scan# 910

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :





SICV209

Tgt Ion:108 Resp: 115523

Ion Ratio Lower Upper

108 100

107 90.7 74.1 111.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.522 ng/ul

RT: 8.411 min Scan# 922

Delta R.T. -0.006 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion: 45 Resp: 97283


Ion Ratio Lower Upper

45 100

77 16.7 13.0 19.4

79 12.2 9.5 14.3









#16
Acetophenone
Concen: 19.899 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:105 Resp: 197609
Ion Ratio Lower Upper
105 100
77 82.0 64.0 96.0
51 27.1 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 19.752 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion: 70 Resp: 104543
Ion Ratio Lower Upper
70 100
42 53.7 43.4 65.2
101 10.2 8.0 12.0
130 21.3 16.2 24.4









#18
4-Methylphenol
Concen: 19.144 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:108 Resp: 127008
Ion Ratio Lower Upper
108 100
107 112.3 89.4 134.2



#19
Hexachloroethane
Concen: 18.779 ng/ul
RT: 8.958 min Scan# 1015
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:117 Resp: 49001
Ion Ratio Lower Upper
117 100
201 85.7 69.5 104.3
199 52.5 43.9 65.9





#22

Nitrobenzene

Concen: 19.145 ng/ul

RT: 9.099 min Scan# 1039

Delta R.T. -0.006 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion: 77 Resp: 152138

Ion Ratio Lower Upper

77 100

123 42.8 32.6 49.0

65 13.6 10.7 16.1



#23

Isophorone

Concen: 19.036 ng/ul

RT: 9.634 min Scan# 1130

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion: 82 Resp: 289356


Ion Ratio Lower Upper

82 100

95 7.4 6.3 9.5

138 16.5 12.8 19.2





#25

2-Nitrophenol

Concen: 19.643 ng/ul

RT: 9.816 min Scan# 1161

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:139 Resp: 70678

Ion Ratio Lower Upper

139 100

65 57.1 45.4 68.2

109 37.3 27.9 41.9



#26

2,4-Dimethylphenol

Concen: 19.218 ng/ul

RT: 9.887 min Scan# 1173

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:107 Resp: 148327


Ion Ratio Lower Upper

107 100

121 47.2 38.5 57.7

122 80.1 64.1 96.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.345 ng/ul
 RT: 10.116 min Scan# 1212
 Delta R.T. -0.006 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07




Instrument :
 BNA_N
 ClientSampleId :
 SICV209


Tgt Ion: 93 Resp: 169570
 Ion Ratio Lower Upper
 93 100
 95 30.5 25.8 38.6
 123 10.8 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 19.583 ng/ul
 RT: 10.358 min Scan# 1253
 Delta R.T. 0.000 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07

Tgt Ion:162 Resp: 111050
 Ion Ratio Lower Upper
 162 100
 164 67.7 51.9 77.9
 98 43.3 33.1 49.7









#30
Naphthalene
Concen: 19.059 ng/ul
RT: 10.752 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:128 Resp: 383333
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 19.453 ng/ul
RT: 10.875 min Scan# 1341
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:127 Resp: 172650
Ion Ratio Lower Upper
127 100
129 33.2 25.4 38.0









#33
Hexachlorobutadiene
Concen: 19.054 ng/ul
RT: 11.028 min Scan# 1367
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:225 Resp: 66508
Ion Ratio Lower Upper
225 100
223 63.0 48.6 73.0
227 63.3 47.8 71.8



#34
Caprolactam
Concen: 11.557 ng/ul
RT: 11.675 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:113 Resp: 24862
Ion Ratio Lower Upper
113 100
55 166.0 133.8 200.8
56 128.6 103.8 155.6









#35
4-Chloro-3-methylphenol
Concen: 19.901 ng/ul
RT: 12.016 min Scan# 1535
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:107 Resp: 147184
Ion Ratio Lower Upper
107 100
144 22.6 19.0 28.4
142 71.2 57.0 85.6



#36
2-Methylnaphthalene
Concen: 19.760 ng/ul
RT: 12.375 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:142 Resp: 270742
Ion Ratio Lower Upper
142 100
141 87.6 71.2 106.8









#37
1-Methylnaphthalene
Concen: 19.119 ng/ul
RT: 12.593 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:142 Resp: 262604
Ion Ratio Lower Upper
142 100
141 95.1 77.2 115.8
116 4.6 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.783 ng/ul
RT: 12.746 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:216 Resp: 124155
Ion Ratio Lower Upper
216 100
214 80.4 62.9 94.3
179 24.9 19.0 28.4
108 21.6 17.4 26.2





#40

Hexachlorocyclopentadiene

Concen: 18.345 ng/ul

RT: 12.716 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:237 Resp: 70796

Ion Ratio Lower Upper

237 100

235 61.0 50.3 75.5

272 13.3 9.4 14.2



#41

2,4,6-Trichlorophenol

Concen: 18.868 ng/ul

RT: 12.993 min Scan# 1701

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:196 Resp: 88547


Ion Ratio Lower Upper

196 100

198 98.4 77.3 115.9

200 31.8 25.4 38.2









#42
2,4,5-Trichlorophenol
Concen: 19.534 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:196 Resp: 97081
Ion Ratio Lower Upper
196 100
198 96.2 78.8 118.2
200 29.4 26.0 39.0



#43
1,1'-Biphenyl
Concen: 19.485 ng/ul
RT: 13.393 min Scan# 1769
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:154 Resp: 352202
Ion Ratio Lower Upper
154 100
153 40.1 32.2 48.2
76 13.8 11.2 16.8









#44
2-Chloronaphthalene
Concen: 19.034 ng/ul
RT: 13.434 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:162 Resp: 268099
Ion Ratio Lower Upper
162 100
164 33.0 26.0 39.0
127 46.2 37.8 56.6



#45
2-Nitroaniline
Concen: 19.350 ng/ul
RT: 13.651 min Scan# 1813
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion: 65 Resp: 98979
Ion Ratio Lower Upper
65 100
92 65.0 55.2 82.8
138 97.6 74.1 111.1









#47
Dimethylphthalate
Concen: 18.840 ng/ul
RT: 14.028 min Scan# 1876
Delta R.T. 0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:163 Resp: 354685
Ion Ratio Lower Upper
163 100
194 3.8 3.2 4.8
164 10.0 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 20.289 ng/ul
RT: 14.151 min Scan# 1898
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:165 Resp: 77320
Ion Ratio Lower Upper
165 100
89 66.4 55.2 82.8
121 24.7 19.6 29.4









#50
Acenaphthylene
Concen: 19.865 ng/ul
RT: 14.281 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:152 Resp: 437967
Ion Ratio Lower Upper
152 100
151 20.2 16.4 24.6
153 13.2 10.9 16.3



#51
3-Nitroaniline
Concen: 20.638 ng/ul
RT: 14.481 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:138 Resp: 84684
Ion Ratio Lower Upper
138 100
108 12.7 9.7 14.5
92 131.2 103.7 155.5









#52
Acenaphthene
Concen: 18.990 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:153 Resp: 297052
Ion Ratio Lower Upper
153 100
152 47.2 38.2 57.4
154 84.3 71.9 107.9



#53
2,4-Dinitrophenol
Concen: 17.348 ng/ul
RT: 14.693 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:184 Resp: 46204
Ion Ratio Lower Upper
184 100
63 82.3 66.4 99.6
154 73.3 59.4 89.0









#55
4-Nitrophenol
Concen: 18.636 ng/ul
RT: 14.810 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:109 Resp: 65963
Ion Ratio Lower Upper
109 100
139 108.8 86.0 129.0
65 125.6 102.6 154.0



#56
Dibenzofuran
Concen: 20.069 ng/ul
RT: 14.963 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:168 Resp: 414803
Ion Ratio Lower Upper
168 100
139 43.1 34.6 52.0









#57
2,4-Dinitrotoluene
Concen: 19.329 ng/ul
RT: 14.940 min Scan# 2032
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:165 Resp: 107271
Ion Ratio Lower Upper
165 100
63 54.8 43.4 65.0
182 2.8 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.285 ng/ul
RT: 15.193 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:232 Resp: 82086
Ion Ratio Lower Upper
232 100
131 58.9 47.4 71.0
130 3.3 2.6 3.8
166 36.4 29.2 43.8









#59
Diethylphthalate
Concen: 18.875 ng/ul
RT: 15.387 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:149 Resp: 376648
Ion Ratio Lower Upper
149 100
177 21.1 16.2 24.4
150 12.7 9.7 14.5



#61
Fluorene
Concen: 19.585 ng/ul
RT: 15.616 min Scan# 2147
Delta R.T. 0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:166 Resp: 338041
Ion Ratio Lower Upper
166 100
165 96.5 74.0 111.0
167 13.6 10.2 15.2









#62
4-Chlorophenyl-phenylether
Concen: 19.221 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:204 Resp: 156389
Ion Ratio Lower Upper
204 100
206 33.9 25.6 38.4
141 76.2 57.2 85.8



#63
4-Nitroaniline
Concen: 23.664 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:138 Resp: 86594
Ion Ratio Lower Upper
138 100
92 57.3 45.5 68.3
108 97.7 79.1 118.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.738 ng/ul
 RT: 15.704 min Scan# 2162
 Delta R.T. 0.000 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07




Instrument :
 BNA_N
 ClientSampleId :
 SICV209


Tgt Ion:198 Resp: 62586
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.1 4.7
 77 33.1 27.3 40.9



#67
 N-Nitrosodiphenylamine
 Concen: 19.310 ng/ul
 RT: 15.828 min Scan# 2183
 Delta R.T. 0.000 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07

Tgt Ion:169 Resp: 293921
 Ion Ratio Lower Upper
 169 100
 168 65.1 52.6 79.0
 167 36.5 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 19.364 ng/ul
RT: 16.504 min Scan# 2298
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:248 Resp: 96941
Ion Ratio Lower Upper
248 100
250 95.9 77.0 115.4
141 93.9 77.0 115.4



#69
Hexachlorobenzene
Concen: 19.130 ng/ul
RT: 16.616 min Scan# 2317
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:284 Resp: 111953
Ion Ratio Lower Upper
284 100
142 41.8 37.9 56.9
249 31.1 28.6 43.0





#70

Atrazine

Concen: 18.683 ng/ul

RT: 16.787 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:200 Resp: 101369

Ion Ratio Lower Upper

200 100

173 27.3 23.2 34.8

215 43.5 38.1 57.1



#71

Pentachlorophenol

Concen: 18.028 ng/ul

RT: 16.969 min Scan# 2377

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:266 Resp: 67033


Ion Ratio Lower Upper

266 100

264 56.8 47.8 71.8

268 60.0 50.2 75.2





#72

Phenanthrene

Concen: 19.038 ng/u1

RT: 17.363 min Scan# 2444

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:178 Resp: 521662

Ion Ratio Lower Upper

178 100

179 15.1 13.0 19.6

176 19.5 16.1 24.1



#74

Anthracene

Concen: 19.014 ng/u1

RT: 17.451 min Scan# 2459

Delta R.T. -0.006 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:178 Resp: 527749


Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 17.7 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.189 ng/uL
RT: 13.351 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:216 Resp: 121943
Ion Ratio Lower Upper
216 100
214 78.0 60.3 90.5
218 45.3 39.2 58.8



#76
Pentachlorobenzene
Concen: 17.021 ng/uL
RT: 14.875 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:250 Resp: 123156
Ion Ratio Lower Upper
250 100
248 62.8 50.7 76.1
252 63.8 51.0 76.4





#77

Carbazole

Concen: 19.681 ng/ul

RT: 17.734 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:167 Resp: 507565

Ion Ratio Lower Upper

167 100

166 22.2 18.2 27.2

139 14.6 11.9 17.9



#78

Di-n-butylphthalate

Concen: 17.859 ng/ul

RT: 18.292 min Scan# 2602

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:149 Resp: 655467


Ion Ratio Lower Upper

149 100

150 8.8 7.3 10.9

104 5.7 4.6 7.0









#80
Fluoranthene
Concen: 18.603 ng/ul
RT: 19.392 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:202 Resp: 602997
Ion Ratio Lower Upper
202 100
101 11.7 9.2 13.8
100 8.8 7.3 10.9



#82
Pyrene
Concen: 18.326 ng/ul
RT: 19.751 min Scan# 2850
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:202 Resp: 608505
Ion Ratio Lower Upper
202 100
101 15.0 11.9 17.9
100 12.0 9.5 14.3





#83

Butylbenzylphthalate

Concen: 17.964 ng/ul

RT: 20.657 min Scan# 3004

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:149 Resp: 303619

Ion Ratio Lower Upper

149 100

91 77.2 63.4 95.2

206 20.0 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 20.223 ng/ul

RT: 21.451 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:252 Resp: 223619


Ion Ratio Lower Upper

252 100

254 73.4 49.0 73.6

126 14.1 11.4 17.0









#85
Benzo(a)anthracene
Concen: 18.585 ng/ul
RT: 21.510 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07




Instrument :
BNA_N
ClientSampleId :
SICV209


Tgt Ion:228 Resp: 615044
Ion Ratio Lower Upper
228 100
229 19.8 16.0 24.0
226 27.2 22.6 33.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.246 ng/ul
RT: 21.433 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BN022649.D
Acq: 15 Nov 2022 14:07

Tgt Ion:149 Resp: 455164
Ion Ratio Lower Upper
149 100
167 27.0 20.7 31.1
279 3.8 2.9 4.3





#87

Chrysene

Concen: 18.056 ng/ul

RT: 21.569 min Scan# 3159

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:228 Resp: 565857

Ion Ratio Lower Upper

228 100

226 29.2 22.6 34.0

229 19.5 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 16.730 ng/ul




RT: 22.369 min Scan# 3295


Delta R.T. -0.006 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Tgt Ion:149 Resp: 771528





#90

Benzo(b)fluoranthene

Concen: 18.241 ng/ul

RT: 23.216 min Scan# 3439

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:252 Resp: 648199

Ion Ratio Lower Upper

252 100

253 21.4 16.2 24.2

125 10.8 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 18.038 ng/ul

RT: 23.263 min Scan# 3447

Delta R.T. 0.000 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:252 Resp: 608429


Ion Ratio Lower Upper

252 100

253 22.8 16.2 24.4

125 11.4 9.0 13.6





#93

Benzo(a)pyrene

Concen: 18.271 ng/ul

RT: 23.851 min Scan# 31

Delta R.T. -0.006 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07

Instrument :

BNA_N

ClientSampleId :

SICV209





Tgt Ion:252 Resp: 550333

Ion Ratio Lower Upper

252 100

253 21.0 15.4 23.2

125 12.5 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.609 ng/ul

RT: 26.510 min Scan# 3999

Delta R.T. -0.012 min

Lab File: BN022649.D

Acq: 15 Nov 2022 14:07




Tgt Ion:276 Resp: 671733


Ion Ratio Lower Upper

276 100

138 25.6 21.3 31.9

277 23.0 19.2 28.8









#95
 Dibenzo(a,h)anthracene
 Concen: 19.489 ng/ul
 RT: 26.527 min Scan# 4003
 Delta R.T. -0.006 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07

Instrument :
 BNA_N
 ClientSampleId :
 SICV209

Tgt Ion:278 Resp: 598378
 Ion Ratio Lower Upper
 278 100
 139 17.8 14.4 21.6
 279 23.2 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 19.479 ng/ul
 RT: 27.292 min Scan# 4132
 Delta R.T. 0.000 min
 Lab File: BN022649.D
 Acq: 15 Nov 2022 14:07

Tgt Ion:276 Resp: 581807
 Ion Ratio Lower Upper
 276 100
 138 22.3 18.6 27.8
 277 22.8 18.0 27.0

