

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022652.D
 Acq On : 15 Nov 2022 17:00
 Operator : CG/JU
 Sample : PB149003BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Quant Time: Nov 15 23:16:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	74418	20.000	ng/ul	0.00
20) Naphthalene-d8	10.704	136	327209	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.563	164	211132	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	432020	20.000	ng/ul	0.00
79) Chrysene-d12	21.539	240	421823	20.000	ng/ul	0.00
88) Perylene-d12	23.974	264	436892	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	12315	5.650	ng/uL	0.00
4) Pyridine-d5	3.628	84	149756	24.294	ng/ul	0.00
7) Phenol-d5	7.052	99	201125	28.615	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	121210	28.763	ng/ul	0.00
11) 2-Chlorophenol-d4	7.405	132	150958	29.327	ng/ul	0.00
15) 4-Methylphenol-d8	8.616	113	159955	28.827	ng/ul	0.00
21) Nitrobenzene-d5	9.063	128	74530	29.427	ng/ul	0.00
24) 2-Nitrophenol-d4	9.787	143	83143	29.420	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	144549	30.003	ng/ul	0.00
31) 4-Chloroaniline-d4	10.857	131	177523	24.191	ng/ul	0.00
46) Dimethylphthalate-d6	13.981	166	464407	30.241	ng/ul	0.00
49) Acenaphthylene-d8	14.257	160	502158	29.677	ng/ul	0.00
54) 4-Nitrophenol-d4	14.798	143	93271	30.467	ng/ul	0.00
60) Fluorene-d10	15.563	176	384414	30.521	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	81536	29.693	ng/ul	0.00
73) Anthracene-d10	17.428	188	598976	31.663	ng/ul	0.00
81) Pyrene-d10	19.727	212	648312	29.657	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.810	264	638271	29.062	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.228	88	23125	9.969	ng/ul#	95
5) Pyridine	3.652	79	153818	25.167	ng/ul	96
6) Benzaldehyde	7.016	77	117261	32.975	ng/ul	98
8) Phenol	7.081	94	194922	26.627	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.305	93	150217	26.334	ng/ul	98
12) 2-Chlorophenol	7.440	128	150435	27.370	ng/ul	97
13) 2-Methylphenol	8.346	108	144078	26.442	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.422	45	198627	26.379	ng/ul	99
16) Acetophenone	8.722	105	236683	26.725	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.705	70	128133	27.146	ng/ul	97
18) 4-Methylphenol	8.675	108	158623	26.810	ng/ul	99
19) Hexachloroethane	8.963	117	61376	26.375	ng/ul	95
22) Nitrobenzene	9.105	77	187838	27.374	ng/ul	99
23) Isophorone	9.634	82	355803	27.108	ng/ul	100
25) 2-Nitrophenol	9.822	139	86636	27.884	ng/ul	99
26) 2,4-Dimethylphenol	9.887	107	183809	27.579	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.122	93	206612	27.297	ng/ul	99
29) 2,4-Dichlorophenol	10.363	162	138151	28.213	ng/ul	97
30) Naphthalene	10.757	128	476063	27.411	ng/ul	99
32) 4-Chloroaniline	10.881	127	201462	26.287	ng/ul	99
33) Hexachlorobutadiene	11.034	225	81111	26.911	ng/ul	96
34) Caprolactam	11.675	113	29819	16.052	ng/ul	98
35) 4-Chloro-3-methylphenol	12.022	107	177067	27.726	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022652.D
 Acq On : 15 Nov 2022 17:00
 Operator : CG/JU
 Sample : PB149003BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Quant Time: Nov 15 23:16:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.381	142	323107	27.310	ng/ul	99
37) 1-Methylnaphthalene	12.598	142	326432	27.523	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.746	216	152948	26.989	ng/ul	98
40) Hexachlorocyclopentadiene	12.716	237	82673	24.987	ng/ul	98
41) 2,4,6-Trichlorophenol	12.998	196	107394	26.691	ng/ul	95
42) 2,4,5-Trichlorophenol	13.075	196	118520	27.816	ng/ul	96
43) 1,1'-Biphenyl	13.393	154	415734	26.826	ng/ul	98
44) 2-Chloronaphthalene	13.440	162	326569	27.043	ng/ul	99
45) 2-Nitroaniline	13.657	65	122736	27.987	ng/ul	97
47) Dimethylphthalate	14.028	163	447864	27.747	ng/ul	98
48) 2,6-Dinitrotoluene	14.157	165	94098	28.800	ng/ul	96
50) Acenaphthylene	14.287	152	525801	27.816	ng/ul	99
51) 3-Nitroaniline	14.487	138	98396	27.969	ng/ul	97
52) Acenaphthene	14.628	153	366327	27.314	ng/ul	95
53) 2,4-Dinitrophenol	14.698	184	59084	25.876	ng/ul	99
55) 4-Nitrophenol	14.816	109	85653	28.225	ng/ul	96
56) Dibenzofuran	14.963	168	496187	28.001	ng/ul	99
57) 2,4-Dinitrotoluene	14.945	165	138828	29.177	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.198	232	100214	27.462	ng/ul	99
59) Diethylphthalate	15.392	149	479311	28.017	ng/ul	99
61) Fluorene	15.616	166	410566	27.744	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.610	204	189165	27.118	ng/ul	93
63) 4-Nitroaniline	15.657	138	99347	31.665	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.710	198	79165	27.891	ng/ul	98
67) N-Nitrosodiphenylamine	15.828	169	363263	28.084	ng/ul	99
68) 4-Bromophenyl-phenylether	16.510	248	121719	28.611	ng/ul	98
69) Hexachlorobenzene	16.622	284	141052	28.363	ng/ul	93
70) Atrazine	16.792	200	128946	27.967	ng/ul	96
71) Pentachlorophenol	16.975	266	84359	26.698	ng/ul	93
72) Phenanthrene	17.369	178	661572	28.411	ng/ul	100
74) Anthracene	17.457	178	679119	28.793	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.357	216	152515	26.770	ng/uL	94
76) Pentachlorobenzene	14.881	250	153602	24.981	ng/uL	98
77) Carbazole	17.739	167	645389	29.449	ng/ul	99
78) Di-n-butylphthalate	18.298	149	825057	26.453	ng/ul	99
80) Fluoranthene	19.392	202	764592	28.024	ng/ul	96
82) Pyrene	19.763	202	801910	28.693	ng/ul	98
83) Butylbenzylphthalate	20.663	149	390058	27.418	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.457	252	270626	29.077	ng/ul	96
85) Benzo(a)anthracene	21.522	228	771789	27.707	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.439	149	565502	26.932	ng/ul	98
87) Chrysene	21.574	228	739295	28.026	ng/ul	98
89) Di-n-octyl phthalate	22.374	149	991834	25.813	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	805319	27.200	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	770548	27.417	ng/ul	99
93) Benzo(a)pyrene	23.868	252	681896	27.171	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.527	276	802325	28.111	ng/ul	98
95) Dibenzo(a,h)anthracene	26.545	278	705881	27.594	ng/ul	99
96) Benzo(g,h,i)perylene	27.315	276	676084	27.168	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :

BNA_N

ClientSampleId :

SLCS003



1,4-Dioxane

Concen: 9.969 ng/uL

RT: 3.228 min Scan# 41

Delta R.T. 0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003





Tgt Ion: 88 Resp: 23125

Ion Ratio Lower Upper

88 100

43 30.4 30.8 46.2#

58 77.9 61.7 92.5



Pyridine

Concen: 25.167 ng/uL

RT: 3.652 min Scan# 113

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion: 79 Resp: 153818


Ion Ratio Lower Upper

79 100

52 70.2 58.6 88.0

51 34.2 28.8 43.2





Benzaldehyde

Concen: 32.975 ng/ul

RT: 7.016 min Scan# 685

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003





Tgt Ion: 77 Resp: 117261

Ion Ratio Lower Upper

77 100

105 91.6 75.2 112.8

106 85.5 69.0 103.6



Phenol

Concen: 26.627 ng/ul

RT: 7.081 min Scan# 696

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion: 94 Resp: 194922


Ion Ratio Lower Upper

94 100

65 31.4 24.3 36.5

66 42.4 33.5 50.3









Bis(2-Chloroethyl)ether
 Concen: 26.334 ng/ul
 RT: 7.305 min Scan# 734
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003


Tgt Ion:	93	Resp:	150217
Ion	Ratio	Lower	Upper
93	100		
63	75.8	61.8	92.6
95	32.1	26.3	39.5



2-Chlorophenol
 Concen: 27.370 ng/ul
 RT: 7.440 min Scan# 757
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:	128	Resp:	150435
Ion	Ratio	Lower	Upper
128	100		
64	48.3	40.0	60.0
130	31.3	26.3	39.5





2-Methylphenol

Concen: 26.442 ng/ul

RT: 8.346 min Scan# 910

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :





SLCS003

Tgt Ion:108 Resp: 144078

Ion Ratio Lower Upper

108 100

107 92.4 74.1 111.1



2,2'-oxybis(1-Chloropropane)

Concen: 26.379 ng/ul

RT: 8.422 min Scan# 924

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion: 45 Resp: 198627


Ion Ratio Lower Upper

45 100

77 15.6 13.0 19.4

79 11.5 9.5 14.3





Acetophenone

Concen: 26.725 ng/ul

RT: 8.722 min Scan# 975

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003





Tgt Ion:105 Resp: 236683

Ion Ratio Lower Upper

105 100

77 81.0 64.0 96.0

51 27.3 20.8 31.2



N-Nitroso-di-n-propylamine

Concen: 27.146 ng/ul

RT: 8.705 min Scan# 972

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion: 70 Resp: 128133




Ion Ratio Lower Upper


70 100

42 52.6 43.4 65.2

101 9.4 8.0 12.0

130 18.2 16.2 24.4





4-Methylphenol

Concen: 26.810 ng/ul

RT: 8.675 min Scan# 967

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :





SLCS003

Tgt Ion:108 Resp: 158623

Ion Ratio Lower Upper

108 100

107 113.4 89.4 134.2



Hexachloroethane

Concen: 26.375 ng/ul

RT: 8.963 min Scan# 1016

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion:117 Resp: 61376


Ion Ratio Lower Upper

117 100

201 89.9 69.5 104.3

199 59.7 43.9 65.9





Nitrobenzene

Concen: 27.374 ng/ul

RT: 9.105 min Scan#

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003





Tgt Ion: 77 Resp: 187838

Ion Ratio Lower Upper

77 100

123 41.4 32.6 49.0

65 13.7 10.7 16.1



Isophorone

Concen: 27.108 ng/ul

RT: 9.634 min Scan# 1130

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion: 82 Resp: 355803

Ion Ratio Lower Upper


82 100

95 7.9 6.3 9.5

138 15.8 12.8 19.2



Abundance Scan 1161 (9.816 min): BN022645.D\data.ms (-11525-)





2-Nitrophenol
Concen: 27.884 ng/ul
RT: 9.822 min Scan# 1162
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00


Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	139	Resp:	86636
Ion	Ratio	Lower	Upper
139	100		
65	57.0	45.4	68.2
109	36.9	27.9	41.9


Abundance Scan 1162 (9.822 min): BN022652.D\data.ms



Abundance Scan 1162 (9.822 min): BN022652.D\data.ms (-1127-)





Abundance Scan 1173 (9.887 min): BN022645.D\data.ms (-11626-)




2,4-Dimethylphenol
Concen: 27.579 ng/ul
RT: 9.887 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	107	Resp:	183809
Ion	Ratio	Lower	Upper
107	100		
121	47.4	38.5	57.7
122	74.3	64.1	96.1

Abundance Scan 1173 (9.887 min): BN022652.D\data.ms




Abundance Scan 1173 (9.887 min): BN022652.D\data.ms (-1139-)





Abundance Scan 1213 (10.122 min): BN022645.D\data.ms (-1207) (-)

Bis(2-Chloroethoxy)methane
Concen: 27.297 ng/ul
RT: 10.122 min Scan# 1213
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00


Instrument :
BNA_N
ClientSampleId :
SLCS003



Tgt Ion:	93	Resp:	206612
Ion	Ratio	Lower	Upper
93	100		
95	31.7	25.8	38.6
123	10.2	8.6	13.0




Abundance Scan 1213 (10.122 min): BN022652.D\data.ms (-1179) (-)





Abundance Scan 1253 (10.358 min): BN022645.D\data.ms (-1229) (-)


2,4-Dichlorophenol
Concen: 28.213 ng/ul
RT: 10.363 min Scan# 1254
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	162	Resp:	138151
Ion	Ratio	Lower	Upper
162	100		
164	67.6	51.9	77.9
98	43.2	33.1	49.7



Abundance Scan 1254 (10.363 min): BN022652.D\data.ms (-1219) (-)









Scan 1320 (10.752 min): BN022645.D\data.ms (-1330) (-)
 Naphthalene
 Concen: 27.411 ng/ul
 RT: 10.757 min Scan# 1320
 Delta R.T. 0.006 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003


Tgt Ion:	128	Resp:	476063
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.9	13.3
127	14.0	10.6	16.0



Scan 1342 (10.881 min): BN022645.D\data.ms (-1332) (-)
 4-Chloroaniline
 Concen: 26.287 ng/ul
 RT: 10.881 min Scan# 1342
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:	127	Resp:	201462
Ion	Ratio	Lower	Upper
127	100		
129	32.5	25.4	38.0









Hexachlorobutadiene
 Concen: 26.911 ng/ul
 RT: 11.034 min Scan# 1368
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Tgt Ion:	225	Resp:	81111
Ion	Ratio	Lower	Upper
225	100		
223	59.7	48.6	73.0
227	65.0	47.8	71.8




Caprolactam
 Concen: 16.052 ng/ul
 RT: 11.675 min Scan# 1477
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:	113	Resp:	29819
Ion	Ratio	Lower	Upper
113	100		
55	172.2	133.8	200.8
56	129.6	103.8	155.6






Abundance Scan 1536 (12.022 min): BN022645.D\data.ms (-1525) (-)




4-Chloro-3-methylphenol
Concen: 27.726 ng/ul
RT: 12.022 min Scan# 1536
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	107	Resp:	177067
Ion	Ratio	Lower	Upper
107	100		
144	23.7	19.0	28.4
142	72.6	57.0	85.6






Abundance Scan 1596 (12.375 min): BN022645.D\data.ms (-1536) (-)




2-Methylnaphthalene
Concen: 27.310 ng/ul
RT: 12.381 min Scan# 1597
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	142	Resp:	323107
Ion	Ratio	Lower	Upper
142	100		
141	89.8	71.2	106.8





Abundance Scan 1634 (12.598 min): BN022645.D\data.ms (-1625) (-)




1-Methylnaphthalene
Concen: 27.523 ng/ul
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003


Tgt Ion:	142	Resp:	326432
Ion	Ratio	Lower	Upper
142	100		
141	92.3	77.2	115.8
116	4.1	3.4	5.0



Abundance Scan 1634 (12.598 min): BN022652.D\data.ms (-1600) (-)





Abundance Scan 1659 (12.746 min): BN022645.D\data.ms (-1632) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 26.989 ng/ul
RT: 12.746 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	216	Resp:	152948
Ion	Ratio	Lower	Upper
216	100		
214	80.9	62.9	94.3
179	24.3	19.0	28.4
108	21.6	17.4	26.2



Abundance Scan 1659 (12.746 min): BN022652.D\data.ms (-1625) (-)



Abundance Scan 1654 (12.716 min): BN022645.D\data.ms (-1640) (-)

Hexachlorocyclopentadiene

Concen: 24.987 ng/ul

RT: 12.716 min Scan# 1640

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003

Tgt Ion:237 Resp: 82673



Ion Ratio Lower Upper

237 100


235 64.1 50.3 75.5

272 13.8 9.4 14.2

Abundance Scan 1654 (12.716 min): BN022652.D\data.ms



Abundance Scan 1654 (12.716 min): BN022652.D\data.ms (-1620) (-)



Abundance Scan 1701 (12.993 min): BN022645.D\data.ms (-1695) (-)

2,4,6-Trichlorophenol

Concen: 26.691 ng/ul

RT: 12.998 min Scan# 1702

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion:196 Resp: 107394



Ion Ratio Lower Upper

196 100


198 101.6 77.3 115.9

200 29.6 25.4 38.2

Abundance Scan 1702 (12.998 min): BN022652.D\data.ms




Abundance Scan 1702 (12.998 min): BN022652.D\data.ms (-1667) (-)





Abundance Scan 1715 (13.075 min): BN022645.D\data.ms (-1702) (-)

2,4,5-Trichlorophenol
Concen: 27.816 ng/ul
RT: 13.075 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00


Instrument :
BNA_N
ClientSampleId :
SLCS003



Tgt Ion:	196	Resp:	118520
Ion	Ratio	Lower	Upper
196	100		
198	93.4	78.8	118.2
200	31.6	26.0	39.0




Abundance Scan 1715 (13.075 min): BN022652.D\data.ms (-1698) (-)





Abundance Scan 1769 (13.393 min): BN022645.D\data.ms (-1753) (-)


1,1'-Biphenyl
Concen: 26.826 ng/ul
RT: 13.393 min Scan# 1769
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00



Tgt Ion:	154	Resp:	415734
Ion	Ratio	Lower	Upper
154	100		
153	41.8	32.2	48.2
76	14.5	11.2	16.8



Abundance Scan 1769 (13.393 min): BN022652.D\data.ms (-1735) (-)



Abundance Scan 1776 (13.434 min): BN022645.D\data.ms (-1769) (-)

2-Chloronaphthalene

Concen: 27.043 ng/ul

RT: 13.440 min Scan# 1777

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003

Tgt Ion:162 Resp: 326569

Ion Ratio Lower Upper

162 100

164 32.2 26.0 39.0

127 45.9 37.8 56.6

Abundance Scan 1777 (13.440 min): BN022652.D\data.ms

Ref

50

162.0

127.1

63.0

101.1

207.0

m/z-->

Abundance Scan 1777 (13.440 min): BN022652.D\data.ms (-1742) (-)

Ref

50

162.0

127.1

63.0

101.1

207.0

m/z-->

Abundance

200000

150000

100000

50000

0

13.40

13.440

13.50

Time-->

Abundance Scan 1813 (13.651 min): BN022645.D\data.ms (-1805) (-)

2-Nitroaniline

Concen: 27.987 ng/ul

RT: 13.657 min Scan# 1814

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion: 65 Resp: 122736

Ion Ratio Lower Upper

65 100

92 66.5 55.2 82.8

138 95.6 74.1 111.1

Abundance Scan 1814 (13.657 min): BN022652.D\data.ms

Ref

50

138.0

92.1

65.1

39.1

207.0

m/z-->

Abundance Scan 1814 (13.657 min): BN022652.D\data.ms (-1779) (-)

Ref

50

138.0

92.1

65.0

39.1

207.0

m/z-->

Abundance

80000

60000

40000

20000


0

13.60

13.657

13.70





Time-->



Dimethylphthalate
Concen: 27.747 ng/ul
RT: 14.028 min Scan# 1877
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00




Instrument :
BNA_N
ClientSampleId :
SLCS003


Tgt Ion:	163	Resp:	447864
Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.2	4.8
164	9.9	8.6	12.8



2,6-Dinitrotoluene
Concen: 28.800 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	165	Resp:	94098
Ion	Ratio	Lower	Upper
165	100		
89	66.0	55.2	82.8
121	22.8	19.6	29.4









Acenaphthylene
Concen: 27.816 ng/ul
RT: 14.287 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00




Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	152	Resp:	525801
Ion	Ratio	Lower	Upper
152	100		
151	21.1	16.4	24.6
153	13.4	10.9	16.3




3-Nitroaniline
Concen: 27.969 ng/ul
RT: 14.487 min Scan# 1955
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	138	Resp:	98396
Ion	Ratio	Lower	Upper
138	100		
108	12.0	9.7	14.5
92	125.6	103.7	155.5






Abundance Scan 1979 (14.628 min): BN022645.D\data.ms (-1952) (-)




Acenaphthene
Concen: 27.314 ng/ul
RT: 14.628 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	153	Resp:	366327
Ion	Ratio	Lower	Upper
153	100		
152	47.2	38.2	57.4
154	83.7	71.9	107.9







Abundance Scan 1990 (14.693 min): BN022645.D\data.ms (-1953) (-)



2,4-Dinitrophenol
Concen: 25.876 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	184	Resp:	59084
Ion	Ratio	Lower	Upper
184	100		
63	81.7	66.4	99.6
154	75.1	59.4	89.0









4-Nitrophenol
 Concen: 28.225 ng/ul
 RT: 14.816 min Scan# 2011
 Delta R.T. 0.006 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Tgt Ion:109 Resp: 85653
 Ion Ratio Lower Upper
 109 100
 139 112.0 86.0 129.0
 65 124.0 102.6 154.0



Dibenzofuran
 Concen: 28.001 ng/ul
 RT: 14.963 min Scan# 2036
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:168 Resp: 496187
 Ion Ratio Lower Upper
 168 100
 139 43.7 34.6 52.0



Abundance Scan 2032 (14.940 min): BN022645.D\data.ms (-2050) (-)

2,4-Dinitrotoluene

Concen: 29.177 ng/ul

RT: 14.945 min Scan# 2032

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003

Tgt Ion:165 Resp: 138828



Ion Ratio Lower Upper

165 100


63 54.8 43.4 65.0

182 2.8 2.3 3.5

Abundance Scan 2033 (14.945 min): BN022652.D\data.ms



Abundance Scan 2033 (14.945 min): BN022652.D\data.ms (-1998) (-)



Abundance Scan 2075 (15.193 min): BN022645.D\data.ms (-2050) (-)

2,3,4,6-Tetrachlorophenol

Concen: 27.462 ng/ul

RT: 15.198 min Scan# 2076

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion:232 Resp: 100214

Ion Ratio Lower Upper



232 100

131 58.1 47.4 71.0


130 3.4 2.6 3.8

166 36.4 29.2 43.8


Abundance Scan 2076 (15.198 min): BN022652.D\data.ms



Abundance Scan 2076 (15.198 min): BN022652.D\data.ms (-2041) (-)






Abundance Scan 2108 (15.387 min): BN022645.D\data.ms (-2159) (-)




Diethylphthalate
Concen: 28.017 ng/ul
RT: 15.392 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:149 Resp: 479311
Ion Ratio Lower Upper
149 100
177 20.3 16.2 24.4
150 13.1 9.7 14.5






Abundance Scan 2146 (15.610 min): BN022645.D\data.ms (-2160) (-)



Fluorene
Concen: 27.744 ng/ul
RT: 15.616 min Scan# 2147
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:166 Resp: 410566
Ion Ratio Lower Upper
166 100
165 96.3 74.0 111.0
167 13.1 10.2 15.2



Abundance Scan 2146 (15.610 min): BN022645.D\data.ms (-2160) (-)

4-Chlorophenyl-phenylether

Concen: 27.118 ng/ul

RT: 15.610 min Scan# 2146

Delta R.T. -0.000 min

Lab File: BN022652.D


Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003





Tgt Ion:204 Resp: 189165

Ion Ratio Lower Upper


204 100

206 35.9 25.6 38.4

141 77.7 57.2 85.8



Abundance Scan 2146 (15.610 min): BN022652.D\data.ms (-2112) (-)



Abundance Scan 2153 (15.651 min): BN022645.D\data.ms (-2168) (-)

4-Nitroaniline

Concen: 31.665 ng/ul

RT: 15.657 min Scan# 2154

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00



Tgt Ion:138 Resp: 99347

Ion Ratio Lower Upper


138 100

92 57.7 45.5 68.3

108 94.9 79.1 118.7



Abundance Scan 2154 (15.657 min): BN022652.D\data.ms (-2119) (-)



Abundance Scan 2162 (15.704 min): BN022645.D\data.ms (-2156) (-)

4,6-Dinitro-2-methylphenol

Concen: 27.891 ng/ul

RT: 15.710 min Scan# 2163

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003

Tgt Ion:198 Resp: 79165



Ion Ratio Lower Upper

198 100


182 4.0 3.1 4.7

77 32.9 27.3 40.9

Abundance Scan 2163 (15.710 min): BN022652.D\data.ms



Abundance Scan 2163 (15.710 min): BN022652.D\data.ms (-2128) (-)



Abundance Scan 2183 (15.828 min): BN022645.D\data.ms (-2167) (-)

N-Nitrosodiphenylamine

Concen: 28.084 ng/ul

RT: 15.828 min Scan# 2183

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion:169 Resp: 363263



Ion Ratio Lower Upper

169 100


168 65.5 52.6 79.0

167 35.6 30.0 45.0

Abundance Scan 2183 (15.828 min): BN022652.D\data.ms



Abundance Scan 2183 (15.828 min): BN022652.D\data.ms (-2149) (-)



Abundance Scan 2298 (16.504 min): BN022645.D\data.ms (-2298) (-)

4-Bromophenyl-phenylether

Concen: 28.611 ng/ul

RT: 16.510 min Scan# 2298

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003

Tgt Ion:248 Resp: 121719



Ion Ratio Lower Upper

248 100


250 99.3 77.0 115.4

141 95.6 77.0 115.4

Abundance Scan 2299 (16.510 min): BN022652.D\data.ms



Abundance Scan 2299 (16.510 min): BN022652.D\data.ms (-2264) (-)



Abundance Scan 2317 (16.616 min): BN022645.D\data.ms (-2317) (-)

Hexachlorobenzene

Concen: 28.363 ng/ul

RT: 16.622 min Scan# 2318

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Tgt Ion:284 Resp: 141052



Ion Ratio Lower Upper

284 100


142 41.7 37.9 56.9

249 33.4 28.6 43.0


Abundance Scan 2318 (16.622 min): BN022652.D\data.ms



Abundance Scan 2318 (16.622 min): BN022652.D\data.ms (-2283) (-)



Abundance Scan 2346 (16.787 min): BN022645.D\data.ms (-2370) (-)





Atrazine
Concen: 27.967 ng/ul
RT: 16.792 min Scan# 2370
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00


Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.8	23.2	34.8
215	45.2	38.1	57.1


Abundance Scan 2347 (16.792 min): BN022652.D\data.ms



Abundance Scan 2347 (16.792 min): BN022652.D\data.ms (-2312) (-)





Abundance Scan 2377 (16.969 min): BN022645.D\data.ms (-2371) (-)




Pentachlorophenol
Concen: 26.698 ng/ul
RT: 16.975 min Scan# 2378
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00


Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.8	47.8	71.8
268	68.3	50.2	75.2

Abundance Scan 2378 (16.975 min): BN022652.D\data.ms



Abundance Scan 2378 (16.975 min): BN022652.D\data.ms (-2343) (-)









Scan 2444 (17.363 min): BN022645.D\data.ms (-2472) (-)
 Phenanthrene
 Concen: 28.411 ng/ul
 RT: 17.369 min Scan# 2444
 Delta R.T. 0.006 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Tgt Ion:	178	Resp:	661572
Ion	Ratio	Lower	Upper
178	100		
179	16.3	13.0	19.6
176	20.2	16.1	24.1



Scan 2460 (17.457 min): BN022645.D\data.ms (-2473) (-)
 Anthracene
 Concen: 28.793 ng/ul
 RT: 17.457 min Scan# 2460
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:	178	Resp:	679119
Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.5	18.7
176	19.1	15.1	22.7



Abundance Scan 1763 (13.357 min): BN022645.D\data.ms (-1756) (-)

1,2,3,4-Tetrachlorobenzene

Concen: 26.770 ng/uL

RT: 13.357 min Scan#

Delta R.T. -0.000 min

Lab File: BN022652.D


Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003






Tgt Ion:216 Resp: 152515

Ion Ratio Lower Upper

216 100

214 82.9 60.3 90.5

218 48.0 39.2 58.8



Abundance Scan 2021 (14.875 min): BN022645.D\data.ms (-2076) (-)

Pentachlorobenzene


Concen: 24.981 ng/uL

RT: 14.881 min Scan# 2022

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00






Tgt Ion:250 Resp: 153602

Ion Ratio Lower Upper


250 100

248 61.1 50.7 76.1

252 63.1 51.0 76.4






Abundance Scan 2507 (17.734 min): BN022645.D\data.ms (-2507) (-)




Carbazole
Concen: 29.449 ng/ul
RT: 17.739 min Scan# 2507
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	167	Resp:	645389
Ion	Ratio	Lower	Upper
167	100		
166	21.9	18.2	27.2
139	14.8	11.9	17.9






Abundance Scan 2602 (18.292 min): BN022645.D\data.ms (-2576) (-)




Di-n-butylphthalate
Concen: 26.453 ng/ul
RT: 18.298 min Scan# 2603
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	149	Resp:	825057
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.3	10.9
104	6.0	4.6	7.0



Abundance Scan 2789 (19.392 min): BN022645.D\data.ms (-2789) (-)



Fluoranthene

Concen: 28.024 ng/ul

RT: 19.392 min Scan# 2789

Delta R.T. -0.000 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00

Instrument :

BNA_N

ClientSampleId :

SLCS003




Tgt Ion:202 Resp: 764592

Ion Ratio Lower Upper


202 100

101 13.1 9.2 13.8

100 10.2 7.3 10.9



Abundance Scan 2851 (19.757 min): BN022645.D\data.ms (-2851) (-)



Pyrene

Concen: 28.693 ng/ul

RT: 19.763 min Scan# 2852

Delta R.T. 0.006 min

Lab File: BN022652.D

Acq: 15 Nov 2022 17:00




Tgt Ion:202 Resp: 801910


Ion Ratio Lower Upper

202 100

101 13.5 11.9 17.9

100 11.4 9.5 14.3









Butylbenzylphthalate
Concen: 27.418 ng/ul
RT: 20.663 min Scan# 3100
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00




Instrument :
BNA_N
ClientSampleId :
SLCS003


Tgt Ion:	149	Resp:	390058
Ion	Ratio	Lower	Upper
149	100		
91	77.4	63.4	95.2
206	17.6	14.3	21.5



3,3'-Dichlorobenzidine
Concen: 29.077 ng/ul
RT: 21.457 min Scan# 3140
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	252	Resp:	270626
Ion	Ratio	Lower	Upper
252	100		
254	57.9	49.0	73.6
126	13.1	11.4	17.0









Benzo(a)anthracene
Concen: 27.707 ng/ul
RT: 21.522 min Scan# 3151
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00




Instrument :
BNA_N
ClientSampleId :
SLCS003


Tgt Ion:	228	Resp:	771789
Ion	Ratio	Lower	Upper
228	100		
229	19.8	16.0	24.0
226	27.9	22.6	33.8



Bis(2-ethylhexyl)phthalate
Concen: 26.932 ng/ul
RT: 21.439 min Scan# 3137
Delta R.T. 0.006 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	149	Resp:	565502
Ion	Ratio	Lower	Upper
149	100		
167	26.9	20.7	31.1
279	3.8	2.9	4.3









Scan 3159 (21.569 min): BN022645.D\data.ms (-3157) (-)
Chrysene
 Concen: 28.026 ng/ul
 RT: 21.574 min Scan# 3157
 Delta R.T. 0.006 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
ClientSampleId :
 SLCS003


Tgt Ion:	228	Resp:	739295
Ion	Ratio	Lower	Upper
228	100		
226	29.1	22.6	34.0
229	19.6	14.8	22.2



Scan 3296 (22.375 min): BN022645.D\data.ms (-3289) (-)
Di-n-octyl phthalate
 Concen: 25.813 ng/ul
 RT: 22.374 min Scan# 3296
 Delta R.T. -0.000 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:149 Resp: 991834









Benzo(b)fluoranthene
Concen: 27.200 ng/ul
RT: 23.227 min Scan# 3441
Delta R.T. 0.012 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00




Instrument :
BNA_N
ClientSampleId :
SLCS003


Tgt Ion:	252	Resp:	805319
Ion Ratio	Lower	Upper	
252	100		
253	19.5	16.2	24.2
125	10.8	8.7	13.1



Benzo(k)fluoranthene
Concen: 27.417 ng/ul
RT: 23.274 min Scan# 3449
Delta R.T. 0.012 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	252	Resp:	770548
Ion Ratio	Lower	Upper	
252	100		
253	20.9	16.2	24.4
125	10.6	9.0	13.6









Benzo(a)pyrene
 Concen: 27.171 ng/ul
 RT: 23.868 min Scan# 4001
 Delta R.T. 0.012 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00




Instrument :
 BNA_N
 ClientSampleId :
 SLCS003

Tgt Ion:	252	Resp:	681896
Ion	Ratio	Lower	Upper
252	100		
253	21.2	15.4	23.2
125	11.6	9.5	14.3




Indeno(1,2,3-cd)pyrene
 Concen: 28.111 ng/ul
 RT: 26.527 min Scan# 4002
 Delta R.T. 0.006 min
 Lab File: BN022652.D
 Acq: 15 Nov 2022 17:00

Tgt Ion:	276	Resp:	802325
Ion	Ratio	Lower	Upper
276	100		
138	26.1	21.3	31.9
277	25.6	19.2	28.8






Abundance Scan 4003 (26.533 min): BN022645.D\data.ms (-3995)




Dibenzo(a,h)anthracene
Concen: 27.594 ng/ul
RT: 26.545 min Scan# 4135
Delta R.T. 0.012 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Instrument :
BNA_N
ClientSampleId :
SLCS003

Tgt Ion:	278	Resp:	705881
Ion	Ratio	Lower	Upper
278	100		
139	17.1	14.4	21.6
279	23.3	19.0	28.6



Abundance Scan 4132 (27.292 min): BN022645.D\data.ms (-4198)



Benzo(g,h,i)perylene
Concen: 27.168 ng/ul
RT: 27.315 min Scan# 4136
Delta R.T. 0.023 min
Lab File: BN022652.D
Acq: 15 Nov 2022 17:00

Tgt Ion:	276	Resp:	676084
Ion	Ratio	Lower	Upper
276	100		
138	22.4	18.6	27.8
277	23.4	18.0	27.0

