

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022666.D
 Acq On : 16 Nov 2022 01:33
 Operator : CG/JU
 Sample : N5577-10
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C9834

Quant Time: Nov 16 03:13:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	78428	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	348053	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.557	164	223199	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	465371	20.000	ng/u1	0.00
79) Chrysene-d12	21.527	240	432934	20.000	ng/u1	0.00
88) Perylene-d12	23.957	264	396829	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	13583	5.913	ng/uL	0.00
4) Pyridine-d5	3.640	84	35082	5.400	ng/u1	0.01
7) Phenol-d5	7.052	99	232194	31.347	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	146584	33.005	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	179039	33.004	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	179970	30.776	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	91419	33.933	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	102764	34.186	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	168311	32.843	ng/u1	0.00
31) 4-Chloroaniline-d4	10.851	131	218617	28.006	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	584156	35.982	ng/u1	0.00
49) Acenaphthylene-d8	14.251	160	630540	35.250	ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	98569	30.456	ng/u1	0.00
60) Fluorene-d10	15.557	176	480354	36.076	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	78098	26.402	ng/u1	0.00
73) Anthracene-d10	17.422	188	728806	35.765	ng/u1	0.00
81) Pyrene-d10	19.727	212	820776	36.583	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.798	264	731481	36.669	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
Data File : BN022666.D
Acq On : 16 Nov 2022 01:33
Operator : CG/JU
Sample : N5577-10
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C9834

Quant Time: Nov 16 03:13:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 15 22:53:43 2022
Response via : Initial Calibration

