

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
 Data File : BN022667.D
 Acq On : 16 Nov 2022 02:10
 Operator : CG/JU
 Sample : N5577-14
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C9838

Quant Time: Nov 16 03:13:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 15 22:53:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	71618	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	316248	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.557	164	206787	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	423455	20.000	ng/u1	0.00
79) Chrysene-d12	21.527	240	403256	20.000	ng/u1	0.00
88) Perylene-d12	23.957	264	371563	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	12606	6.009	ng/uL	0.00
4) Pyridine-d5	3.640	84	26891	4.533	ng/u1	0.01
7) Phenol-d5	7.052	99	222633	32.914	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	140816	34.721	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	169214	34.159	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	180898	33.876	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	87173	35.611	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	99853	36.558	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	160770	34.526	ng/u1	0.00
31) 4-Chloroaniline-d4	10.851	131	226527	31.938	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	563748	37.481	ng/u1	0.00
49) Acenaphthylene-d8	14.251	160	602054	36.328	ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	78886	26.309	ng/u1	0.00
60) Fluorene-d10	15.557	176	472395	38.294	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.686	200	52649	19.561	ng/u1	0.00
73) Anthracene-d10	17.422	188	724163	39.055	ng/u1	0.00
81) Pyrene-d10	19.727	212	796276	38.103	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.798	264	720022	38.549	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111522\
Data File : BN022667.D
Acq On : 16 Nov 2022 02:10
Operator : CG/JU
Sample : N5577-14
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C9838

Quant Time: Nov 16 03:13:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 15 22:53:43 2022
Response via : Initial Calibration

