

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022714.D
 Acq On : 17 Nov 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359

Quant Time: Nov 17 22:10:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 17 22:08:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	74848	20.000	ng/u1	0.01
20) Naphthalene-d8	10.698	136	336251	20.000	ng/u1	0.02
38) Acenaphthene-d10	14.557	164	214884	20.000	ng/u1	0.03
64) Phenanthrene-d10	17.316	188	443076	20.000	ng/u1	0.03
79) Chrysene-d12	21.527	240	416624	20.000	ng/u1	0.04
88) Perylene-d12	23.956	264	414993	20.000	ng/u1	0.06
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.187	96	16765	7.647	ng/uL	0.00
4) Pyridine-d5	3.622	84	121480	19.594	ng/u1	0.00
7) Phenol-d5	7.040	99	134760	19.063	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.199	67	85549	20.184	ng/u1	0.00
11) 2-Chlorophenol-d4	7.399	132	102199	19.741	ng/u1	0.01
15) 4-Methylphenol-d8	8.598	113	109866	19.686	ng/u1	0.01
21) Nitrobenzene-d5	9.051	128	51529	19.798	ng/u1	0.01
24) 2-Nitrophenol-d4	9.781	143	56566	19.478	ng/u1	0.02
28) 2,4-Dichlorophenol-d3	10.322	165	99373	20.072	ng/u1	0.02
31) 4-Chloroaniline-d4	10.845	131	145284	19.265	ng/u1	0.02
46) Dimethylphthalate-d6	13.969	166	303445	19.415	ng/u1	0.02
49) Acenaphthylene-d8	14.245	160	346784	20.137	ng/u1	0.02
54) 4-Nitrophenol-d4	14.781	143	55414	17.785	ng/u1	0.03
60) Fluorene-d10	15.551	176	249725	19.481	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.686	200	45532	16.167	ng/u1	0.03
73) Anthracene-d10	17.416	188	394991	20.359	ng/u1	0.03
81) Pyrene-d10	19.721	212	435155	20.155	ng/u1	0.04
92) Benzo(a)pyrene-d12	23.798	264	406407	19.481	ng/u1	0.06
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.222	88	18232	7.815	ng/uL#	88
5) Pyridine	3.646	79	121555	19.774	ng/u1	100
6) Benzaldehyde	7.004	77	77638	21.707	ng/u1	99
8) Phenol	7.069	94	145446	19.755	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.293	93	111963	19.515	ng/u1	100
12) 2-Chlorophenol	7.428	128	111621	20.192	ng/u1	99
13) 2-Methylphenol	8.328	108	107606	19.635	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.404	45	96765	12.777	ng/u1	98
16) Acetophenone	8.710	105	179506	20.152	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.693	70	94624	19.932	ng/u1	98
18) 4-Methylphenol	8.663	108	117735	19.785	ng/u1	92
19) Hexachloroethane	8.951	117	46579	19.901	ng/u1	98
22) Nitrobenzene	9.093	77	141409	20.054	ng/u1	98
23) Isophorone	9.622	82	257931	19.123	ng/u1	100
25) 2-Nitrophenol	9.810	139	63843	19.996	ng/u1	97
26) 2,4-Dimethylphenol	9.875	107	139371	20.349	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.110	93	154808	19.903	ng/u1	99
29) 2,4-Dichlorophenol	10.351	162	103502	20.568	ng/u1	97
30) Naphthalene	10.745	128	362473	20.310	ng/u1	99
32) 4-Chloroaniline	10.869	127	157622	20.014	ng/u1	98
33) Hexachlorobutadiene	11.022	225	63995	20.661	ng/u1	95
34) Caprolactam	11.663	113	34709	18.182	ng/u1	96
35) 4-Chloro-3-methylphenol	12.010	107	132644	20.211	ng/u1	99
36) 2-Methylnaphthalene	12.369	142	247641	20.368	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022714.D
 Acq On : 17 Nov 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359

Quant Time: Nov 17 22:10:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 17 22:08:43 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.587	142	246656	20.238	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.739	216	116672	20.228	ng/ul	99
40) Hexachlorocyclopentadiene	12.710	237	52668	15.640	ng/ul	95
41) 2,4,6-Trichlorophenol	12.986	196	79498	19.413	ng/ul	99
42) 2,4,5-Trichlorophenol	13.063	196	86099	19.854	ng/ul	92
43) 1,1'-Biphenyl	13.386	154	312472	19.811	ng/ul	98
44) 2-Chloronaphthalene	13.428	162	247851	20.166	ng/ul	99
45) 2-Nitroaniline	13.645	65	87949	19.704	ng/ul	97
47) Dimethylphthalate	14.016	163	316302	19.254	ng/ul	99
48) 2,6-Dinitrotoluene	14.145	165	64293	19.334	ng/ul	98
50) Acenaphthylene	14.275	152	388015	20.169	ng/ul	99
51) 3-Nitroaniline	14.475	138	68929	19.251	ng/ul	99
52) Acenaphthene	14.622	153	269734	19.761	ng/ul	99
53) 2,4-Dinitrophenol	14.686	184	38323	16.490	ng/ul	97
55) 4-Nitrophenol	14.798	109	58302	18.877	ng/ul	95
56) Dibenzofuran	14.957	168	367558	20.380	ng/ul	99
57) 2,4-Dinitrotoluene	14.933	165	96410	19.909	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.186	232	71616	19.282	ng/ul	98
59) Diethylphthalate	15.380	149	336843	19.345	ng/ul	99
61) Fluorene	15.610	166	304910	20.245	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.604	204	141377	19.913	ng/ul	97
63) 4-Nitroaniline	15.645	138	67853	21.250	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.698	198	47088	16.176	ng/ul#	93
67) N-Nitrosodiphenylamine	15.822	169	263259	19.845	ng/ul	98
68) 4-Bromophenyl-phenylether	16.498	248	86618	19.852	ng/ul	97
69) Hexachlorobenzene	16.610	284	101200	19.842	ng/ul	95
70) Atrazine	16.780	200	90207	19.077	ng/ul	97
71) Pentachlorophenol	16.969	266	57855	17.853	ng/ul	98
72) Phenanthrene	17.357	178	478282	20.027	ng/ul	98
74) Anthracene	17.451	178	491924	20.336	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.345	216	115881	19.833	ng/uL	97
76) Pentachlorobenzene	14.875	250	124846	19.798	ng/uL	98
77) Carbazole	17.727	167	449259	19.988	ng/ul	98
78) Di-n-butylphthalate	18.286	149	606330	18.955	ng/ul	100
80) Fluoranthene	19.386	202	547291	20.310	ng/ul	98
82) Pyrene	19.751	202	579987	21.011	ng/ul	99
83) Butylbenzylphthalate	20.657	149	269589	19.186	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.451	252	181224	19.714	ng/ul	96
85) Benzo(a)anthracene	21.510	228	544876	19.805	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.433	149	393645	18.981	ng/ul	99
87) Chrysene	21.563	228	501851	19.262	ng/ul	95
89) Di-n-octyl phthalate	22.368	149	657647	18.019	ng/ul	100
90) Benzo(b)fluoranthene	23.209	252	538041	19.131	ng/ul	97
91) Benzo(k)fluoranthene	23.257	252	543028	20.342	ng/ul	99
93) Benzo(a)pyrene	23.845	252	453131	19.009	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.503	276	500878	18.475	ng/ul	97
95) Dibenzo(a,h)anthracene	26.521	278	458180	18.856	ng/ul	99
96) Benzo(g,h,i)perylene	27.292	276	431978	18.274	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022714.D
 Acq On : 17 Nov 2022 09:48
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359

Quant Time: Nov 17 22:10:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 17 22:08:43 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.815 ng/uL
RT: 3.222 min Scan# 40
Delta R.T. 0.006 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion: 88 Resp: 18232
Ion Ratio Lower Upper
88 100
43 28.6 30.8 46.2#
58 68.3 61.7 92.5



#5
Pyridine
Concen: 19.774 ng/uL
RT: 3.646 min Scan# 112
Delta R.T. 0.006 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion: 79 Resp: 121555
Ion Ratio Lower Upper
79 100
52 73.4 58.6 88.0
51 35.4 28.8 43.2









#6
Benzaldehyde
Concen: 21.707 ng/ul
RT: 7.004 min Scan# 683
Delta R.T. 0.012 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion: 77 Resp: 77638
Ion Ratio Lower Upper
77 100
105 94.9 75.2 112.8
106 87.5 69.0 103.6



#8
Phenol
Concen: 19.755 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. 0.012 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion: 94 Resp: 145446
Ion Ratio Lower Upper
94 100
65 29.7 24.3 36.5
66 41.9 33.5 50.3









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.515 ng/ul
 RT: 7.293 min Scan# 732
 Delta R.T. 0.006 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359

Tgt Ion: 93 Resp: 111963




Ion	Ratio	Lower	Upper
93	100		
63	77.4	61.8	92.6
95	32.8	26.3	39.5




#12
 2-Chlorophenol
 Concen: 20.192 ng/ul
 RT: 7.428 min Scan# 755
 Delta R.T. 0.012 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion:128 Resp: 111621

Ion	Ratio	Lower	Upper
128	100		
64	50.6	40.0	60.0
130	31.8	26.3	39.5





#13

2-Methylphenol

Concen: 19.635 ng/ul

RT: 8.328 min Scan# 908

Delta R.T. 0.012 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :





SSTD020359

Tgt Ion:108 Resp: 107606

Ion Ratio Lower Upper

108 100

107 88.2 74.1 111.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.777 ng/ul

RT: 8.404 min Scan# 921

Delta R.T. 0.006 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion: 45 Resp: 96765


Ion Ratio Lower Upper

45 100

77 15.8 13.0 19.4

79 10.8 9.5 14.3









#16
Acetophenone
Concen: 20.152 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. 0.012 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:105 Resp: 179506
Ion Ratio Lower Upper
105 100
77 81.6 64.0 96.0
51 27.9 20.8 31.2



#17
N-Nitroso-di-n-propylamine
Concen: 19.932 ng/ul
RT: 8.693 min Scan# 970
Delta R.T. 0.012 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion: 70 Resp: 94624
Ion Ratio Lower Upper
70 100
42 55.4 43.4 65.2
101 9.2 8.0 12.0
130 19.2 16.2 24.4









#18
4-Methylphenol
Concen: 19.785 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.012 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:108 Resp: 117735
Ion Ratio Lower Upper
108 100
107 119.8 89.4 134.2



#19
Hexachloroethane
Concen: 19.901 ng/ul
RT: 8.951 min Scan# 1014
Delta R.T. 0.018 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:117 Resp: 46579
Ion Ratio Lower Upper
117 100
201 85.2 69.5 104.3
199 53.3 43.9 65.9





#22

Nitrobenzene

Concen: 20.054 ng/ul

RT: 9.093 min Scan# 1038

Delta R.T. 0.012 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion: 77 Resp: 141409

Ion Ratio Lower Upper

77 100

123 41.8 32.6 49.0

65 14.1 10.7 16.1



#23

Isophorone

Concen: 19.123 ng/ul

RT: 9.622 min Scan# 1128

Delta R.T. 0.018 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion: 82 Resp: 257931


Ion Ratio Lower Upper

82 100

95 7.7 6.3 9.5

138 15.8 12.8 19.2









#25
2-Nitrophenol
Concen: 19.996 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. 0.018 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:139 Resp: 63843
Ion Ratio Lower Upper
139 100
65 59.7 45.4 68.2
109 35.9 27.9 41.9



#26
2,4-Dimethylphenol
Concen: 20.349 ng/ul
RT: 9.875 min Scan# 1171
Delta R.T. 0.018 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:107 Resp: 139371
Ion Ratio Lower Upper
107 100
121 47.4 38.5 57.7
122 78.2 64.1 96.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.903 ng/ul
 RT: 10.110 min Scan# 1211
 Delta R.T. 0.018 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359


Tgt Ion: 93 Resp: 154808
 Ion Ratio Lower Upper
 93 100
 95 32.0 25.8 38.6
 123 10.2 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 20.568 ng/ul
 RT: 10.351 min Scan# 1252
 Delta R.T. 0.023 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion: 162 Resp: 103502
 Ion Ratio Lower Upper
 162 100
 164 62.2 51.9 77.9
 98 39.2 33.1 49.7





#30

Naphthalene

Concen: 20.310 ng/ul

RT: 10.745 min Scan# 1319

Delta R.T. 0.018 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:128 Resp: 362473

Ion Ratio Lower Upper

128 100

129 10.6 8.9 13.3

127 13.2 10.6 16.0



#32

4-Chloroaniline

Concen: 20.014 ng/ul

RT: 10.869 min Scan# 1340

Delta R.T. 0.018 min

Lab File: BN022714.D




Acq: 17 Nov 2022 09:48


Tgt Ion:127 Resp: 157622

Ion Ratio Lower Upper

127 100

129 32.7 25.4 38.0





#33

Hexachlorobutadiene

Concen: 20.661 ng/ul

RT: 11.022 min Scan# 1366

Delta R.T. 0.018 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:225 Resp: 63995

Ion Ratio Lower Upper

225 100

223 64.0 48.6 73.0

227 64.2 47.8 71.8



#34

Caprolactam

Concen: 18.182 ng/ul

RT: 11.663 min Scan# 1475

Delta R.T. 0.023 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:113 Resp: 34709


Ion Ratio Lower Upper

113 100

55 158.5 133.8 200.8

56 127.9 103.8 155.6









#35
4-Chloro-3-methylphenol
Concen: 20.211 ng/ul
RT: 12.010 min Scan# 1534
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:107 Resp: 132644
Ion Ratio Lower Upper
107 100
144 24.6 19.0 28.4
142 71.5 57.0 85.6



#36
2-Methylnaphthalene
Concen: 20.368 ng/ul
RT: 12.369 min Scan# 1595
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:142 Resp: 247641
Ion Ratio Lower Upper
142 100
141 89.8 71.2 106.8









#37
1-Methylnaphthalene
Concen: 20.238 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:142 Resp: 246656
Ion Ratio Lower Upper
142 100
141 92.1 77.2 115.8
116 4.8 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.228 ng/ul
RT: 12.739 min Scan# 1658
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:216 Resp: 116672
Ion Ratio Lower Upper
216 100
214 78.8 62.9 94.3
179 23.1 19.0 28.4
108 21.0 17.4 26.2









#40
Hexachlorocyclopentadiene
Concen: 15.640 ng/ul
RT: 12.710 min Scan# 1653
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:237 Resp: 52668
Ion Ratio Lower Upper
237 100
235 66.8 50.3 75.5
272 12.8 9.4 14.2



#41
2,4,6-Trichlorophenol
Concen: 19.413 ng/ul
RT: 12.986 min Scan# 1700
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:196 Resp: 79498
Ion Ratio Lower Upper
196 100
198 95.7 77.3 115.9
200 30.5 25.4 38.2





#42

2,4,5-Trichlorophenol

Concen: 19.854 ng/ul

RT: 13.063 min Scan# 1713

Delta R.T. 0.023 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:196 Resp: 86099

Ion Ratio Lower Upper

196 100

198 90.3 78.8 118.2

200 29.5 26.0 39.0



#43

1,1'-Biphenyl

Concen: 19.811 ng/ul

RT: 13.386 min Scan# 1768

Delta R.T. 0.023 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:154 Resp: 312472


Ion Ratio Lower Upper

154 100

153 41.6 32.2 48.2

76 14.4 11.2 16.8









#44
2-Chloronaphthalene
Concen: 20.166 ng/ul
RT: 13.428 min Scan# 1775
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:162 Resp: 247851
Ion Ratio Lower Upper
162 100
164 33.4 26.0 39.0
127 46.3 37.8 56.6



#45
2-Nitroaniline
Concen: 19.704 ng/ul
RT: 13.645 min Scan# 1812
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion: 65 Resp: 87949
Ion Ratio Lower Upper
65 100
92 66.3 55.2 82.8
138 90.0 74.1 111.1









#47
Dimethylphthalate
Concen: 19.254 ng/ul
RT: 14.016 min Scan# 1875
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:163 Resp: 316302
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.5 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 19.334 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:165 Resp: 64293
Ion Ratio Lower Upper
165 100
89 70.8 55.2 82.8
121 24.1 19.6 29.4









#50
Acenaphthylene
Concen: 20.169 ng/ul
RT: 14.275 min Scan# 1919
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:152 Resp: 388015
Ion Ratio Lower Upper
152 100
151 19.8 16.4 24.6
153 13.0 10.9 16.3



#51
3-Nitroaniline
Concen: 19.251 ng/ul
RT: 14.475 min Scan# 1953
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:138 Resp: 68929
Ion Ratio Lower Upper
138 100
108 12.4 9.7 14.5
92 128.1 103.7 155.5









#52
 Acenaphthene
 Concen: 19.761 ng/ul
 RT: 14.622 min Scan# 1978
 Delta R.T. 0.029 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359

Tgt Ion:153 Resp: 269734




Ion	Ratio	Lower	Upper
153	100		
152	48.5	38.2	57.4
154	91.1	71.9	107.9




#53
 2,4-Dinitrophenol
 Concen: 16.490 ng/ul
 RT: 14.686 min Scan# 1989
 Delta R.T. 0.023 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion:184 Resp: 38323

Ion	Ratio	Lower	Upper
184	100		
63	84.6	66.4	99.6
154	71.3	59.4	89.0









#55
4-Nitrophenol
Concen: 18.877 ng/ul
RT: 14.798 min Scan# 2008
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:109 Resp: 58302
Ion Ratio Lower Upper
109 100
139 104.9 86.0 129.0
65 119.4 102.6 154.0



#56
Dibenzofuran
Concen: 20.380 ng/ul
RT: 14.957 min Scan# 2035
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:168 Resp: 367558
Ion Ratio Lower Upper
168 100
139 42.4 34.6 52.0





#57

2,4-Dinitrotoluene

Concen: 19.909 ng/ul

RT: 14.933 min Scan# 2031

Delta R.T. 0.023 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:165 Resp: 96410

Ion Ratio Lower Upper

165 100

63 54.1 43.4 65.0

182 2.8 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.282 ng/ul

RT: 15.186 min Scan# 2074

Delta R.T. 0.029 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Tgt Ion:232 Resp: 71616




Ion Ratio Lower Upper


232 100

131 60.5 47.4 71.0

130 3.2 2.6 3.8

166 35.7 29.2 43.8





#59

Diethylphthalate

Concen: 19.345 ng/ul

RT: 15.380 min Scan# 2107

Delta R.T. 0.023 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:149 Resp: 336843

Ion Ratio Lower Upper

149 100

177 19.9 16.2 24.4

150 12.7 9.7 14.5



#61

Fluorene

Concen: 20.245 ng/ul

RT: 15.610 min Scan# 2146

Delta R.T. 0.029 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:166 Resp: 304910


Ion Ratio Lower Upper

166 100

165 97.0 74.0 111.0

167 13.2 10.2 15.2









#62
4-Chlorophenyl-phenylether
Concen: 19.913 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:204 Resp: 141377
Ion Ratio Lower Upper
204 100
206 32.7 25.6 38.4
141 74.3 57.2 85.8



#63
4-Nitroaniline
Concen: 21.250 ng/ul
RT: 15.645 min Scan# 2152
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:138 Resp: 67853
Ion Ratio Lower Upper
138 100
92 57.7 45.5 68.3
108 98.5 79.1 118.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.176 ng/ul
 RT: 15.698 min Scan# 2161
 Delta R.T. 0.023 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359


Tgt Ion:198 Resp: 47088
 Ion Ratio Lower Upper
 198 100
 182 5.0 3.1 4.7#
 77 38.1 27.3 40.9



#67
 N-Nitrosodiphenylamine
 Concen: 19.845 ng/ul
 RT: 15.822 min Scan# 2182
 Delta R.T. 0.029 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion:169 Resp: 263259
 Ion Ratio Lower Upper
 169 100
 168 67.3 52.6 79.0
 167 36.5 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 19.852 ng/ul
RT: 16.498 min Scan# 2297
Delta R.T. 0.023 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:248 Resp: 86618
Ion Ratio Lower Upper
248 100
250 95.0 77.0 115.4
141 100.1 77.0 115.4



#69
Hexachlorobenzene
Concen: 19.842 ng/ul
RT: 16.610 min Scan# 2316
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:284 Resp: 101200
Ion Ratio Lower Upper
284 100
142 43.2 37.9 56.9
249 33.4 28.6 43.0









#70
 Atrazine
 Concen: 19.077 ng/ul
 RT: 16.780 min Scan# 21
 Delta R.T. 0.029 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020359


Tgt Ion:200 Resp: 90207
 Ion Ratio Lower Upper
 200 100
 173 26.9 23.2 34.8
 215 46.3 38.1 57.1



#71
 Pentachlorophenol
 Concen: 17.853 ng/ul
 RT: 16.969 min Scan# 2377
 Delta R.T. 0.035 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion:266 Resp: 57855
 Ion Ratio Lower Upper
 266 100
 264 57.0 47.8 71.8
 268 62.5 50.2 75.2









#72
Phenanthrene
Concen: 20.027 ng/ul
RT: 17.357 min Scan# 2443
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:178 Resp: 478282
Ion Ratio Lower Upper
178 100
179 15.6 13.0 19.6
176 19.0 16.1 24.1



#74
Anthracene
Concen: 20.336 ng/ul
RT: 17.451 min Scan# 2459
Delta R.T. 0.029 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:178 Resp: 491924
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 18.3 15.1 22.7









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 19.833 ng/uL
 RT: 13.345 min Scan# 1761
 Delta R.T. 0.018 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Instrument :
 BNA_N
 ClientSampleId :
 SST020359

Tgt Ion: 216 Resp: 115881




Ion	Ratio	Lower	Upper
216	100		
214	73.3	60.3	90.5
218	51.4	39.2	58.8




#76
 Pentachlorobenzene
 Concen: 19.798 ng/uL
 RT: 14.875 min Scan# 2021
 Delta R.T. 0.029 min
 Lab File: BN022714.D
 Acq: 17 Nov 2022 09:48

Tgt Ion: 250 Resp: 124846

Ion	Ratio	Lower	Upper
250	100		
248	61.6	50.7	76.1
252	64.9	51.0	76.4





#77

Carbazole

Concen: 19.988 ng/ul

RT: 17.727 min Scan# 2506

Delta R.T. 0.029 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:167 Resp: 449259

Ion Ratio Lower Upper

167 100

166 21.3 18.2 27.2

139 15.1 11.9 17.9



#78

Di-n-butylphthalate

Concen: 18.955 ng/ul

RT: 18.286 min Scan# 2601

Delta R.T. 0.029 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:149 Resp: 606330


Ion Ratio Lower Upper

149 100

150 8.9 7.3 10.9

104 5.8 4.6 7.0









#80
Fluoranthene
Concen: 20.310 ng/ul
RT: 19.386 min Scan# 2788
Delta R.T. 0.035 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48




Instrument :
BNA_N
ClientSampleId :
SSTD020359


Tgt Ion:202 Resp: 547291
Ion Ratio Lower Upper
202 100
101 12.3 9.2 13.8
100 9.7 7.3 10.9



#82
Pyrene
Concen: 21.011 ng/ul
RT: 19.751 min Scan# 2850
Delta R.T. 0.035 min
Lab File: BN022714.D
Acq: 17 Nov 2022 09:48

Tgt Ion:202 Resp: 579987
Ion Ratio Lower Upper
202 100
101 13.9 11.9 17.9
100 12.0 9.5 14.3





#83

Butylbenzylphthalate

Concen: 19.186 ng/ul

RT: 20.657 min Scan# 3004

Delta R.T. 0.035 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:149 Resp: 269589

Ion Ratio Lower Upper

149 100

91 73.7 63.4 95.2

206 19.1 14.3 21.5



#84

3,3'-Dichlorobenzidine

Concen: 19.714 ng/ul

RT: 21.451 min Scan# 3139

Delta R.T. 0.041 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:252 Resp: 181224


Ion Ratio Lower Upper

252 100

254 63.7 49.0 73.6

126 11.7 11.4 17.0





#85

Benzo(a)anthracene

Concen: 19.805 ng/ul

RT: 21.510 min Scan# 3149

Delta R.T. 0.035 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:228 Resp: 544876

Ion Ratio Lower Upper

228 100

229 19.2 16.0 24.0

226 27.1 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.981 ng/ul

RT: 21.433 min Scan# 3136

Delta R.T. 0.035 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:149 Resp: 393645


Ion Ratio Lower Upper

149 100

167 26.2 20.7 31.1

279 3.9 2.9 4.3





#87

Chrysene

Concen: 19.262 ng/ul

RT: 21.563 min Scan# 3158

Delta R.T. 0.035 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:228 Resp: 501851

Ion Ratio Lower Upper

228 100

226 31.1 22.6 34.0

229 20.0 14.8 22.2



#89

Di-n-octyl phthalate

Concen: 18.019 ng/ul




RT: 22.368 min Scan# 3295


Delta R.T. 0.047 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Tgt Ion:149 Resp: 657647





#90

Benzo(b)fluoranthene

Concen: 19.131 ng/ul

RT: 23.209 min Scan# 3438

Delta R.T. 0.053 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:252 Resp: 538041

Ion Ratio Lower Upper

252 100

253 21.1 16.2 24.2

125 12.3 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 20.342 ng/ul

RT: 23.257 min Scan# 3446

Delta R.T. 0.047 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:252 Resp: 543028


Ion Ratio Lower Upper

252 100

253 20.8 16.2 24.4

125 11.0 9.0 13.6





#93

Benzo(a)pyrene

Concen: 19.009 ng/ul

RT: 23.845 min Scan# 3546

Delta R.T. 0.053 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:252 Resp: 453131

Ion Ratio Lower Upper

252 100

253 19.4 15.4 23.2

125 12.8 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.475 ng/ul

RT: 26.503 min Scan# 3998

Delta R.T. 0.088 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48




Tgt Ion:276 Resp: 500878


Ion Ratio Lower Upper

276 100

138 25.1 21.3 31.9

277 25.5 19.2 28.8





#95

Dibenzo(a,h)anthracene

Concen: 18.856 ng/ul

RT: 26.521 min Scan# 4001

Delta R.T. 0.094 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Instrument :

BNA_N

ClientSampleId :

SSTD020359





Tgt Ion:278 Resp: 458180

Ion Ratio Lower Upper

278 100

139 18.1 14.4 21.6

279 23.3 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 18.274 ng/ul

RT: 27.292 min Scan# 4132

Delta R.T. 0.112 min

Lab File: BN022714.D

Acq: 17 Nov 2022 09:48

Tgt Ion:276 Resp: 431978

Ion Ratio Lower Upper

276 100

138 22.5 18.6 27.8

277 21.6 18.0 27.0

