

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022724.D
 Acq On : 17 Nov 2022 17:41
 Operator : CG/JU
 Sample : PB149132BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK132

Quant Time: Nov 17 22:12:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 17 22:08:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	45946	20.000	ng/u1	0.00
20) Naphthalene-d8	10.675	136	222296	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	156326	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	336278	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	358651	20.000	ng/u1	0.00
88) Perylene-d12	23.892	264	380838	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	7669	5.699	ng/uL	0.00
4) Pyridine-d5	3.622	84	58634	15.406	ng/u1	0.00
7) Phenol-d5	7.028	99	123728	28.512	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	85804	32.978	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	95657	30.100	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	103737	30.280	ng/u1	0.00
21) Nitrobenzene-d5	9.040	128	51006	29.643	ng/u1	0.00
24) 2-Nitrophenol-d4	9.763	143	55453	28.883	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.304	165	90834	27.752	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	116129	23.293	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	372829	32.789	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	379098	30.259	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	65049	28.697	ng/u1	0.00
60) Fluorene-d10	15.522	176	305940	32.806	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	53573	25.064	ng/u1	0.00
73) Anthracene-d10	17.386	188	487796	33.127	ng/u1	0.00
81) Pyrene-d10	19.686	212	558003	30.022	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	622355	32.509	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
Data File : BN022724.D
Acq On : 17 Nov 2022 17:41
Operator : CG/JU
Sample : PB149132BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK132

Quant Time: Nov 17 22:12:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 17 22:08:43 2022
Response via : Initial Calibration

