

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022772.D
 Acq On : 19 Nov 2022 01:04
 Operator : CG/JU
 Sample : N5689-17
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGCF9

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/21/2022
 Supervised By :mohammad ahmed 11/21/2022

Quant Time: Nov 19 01:44:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 18 05:17:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	56926	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	276818	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.522	164	189897	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.275	188	406625	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	409951	20.000	ng/u1	0.00
88) Perylene-d12	23.880	264	404587	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	4526m	2.714	ng/uL	0.00
4) Pyridine-d5	3.628	84	23408	4.964	ng/u1	0.01
7) Phenol-d5	7.022	99	38913	7.238	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	120856	37.491	ng/u1	0.00
11) 2-Chlorophenol-d4	7.375	132	104714	26.594	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	74809	17.625	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	74724	34.874	ng/u1	0.00
24) 2-Nitrophenol-d4	9.752	143	75638	31.637	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	120360	29.530	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	165028	26.581	ng/u1	0.00
46) Dimethylphthalate-d6	13.940	166	523857	37.927	ng/u1	0.00
49) Acenaphthylene-d8	14.216	160	547270	35.960	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	17048	6.191	ng/u1	0.00
60) Fluorene-d10	15.516	176	433739	38.288	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	79683	30.830	ng/u1	0.00
73) Anthracene-d10	17.375	188	758189	42.582	ng/u1	0.00
81) Pyrene-d10	19.680	212	855676	40.277	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.721	264	898816	44.194	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
Data File : BN022772.D
Acq On : 19 Nov 2022 01:04
Operator : CG/JU
Sample : N5689-17
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGCF9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/21/2022
Supervised By :mohammad ahmed 11/21/2022

Quant Time: Nov 19 01:44:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 18 05:17:17 2022
Response via : Initial Calibration

