

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
 Data File : BN022773.D
 Acq On : 19 Nov 2022 01:40
 Operator : CG/JU
 Sample : N5689-18
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BGCG0

Quant Time: Nov 19 02:18:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 18 05:17:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	60755	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	286330	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.522	164	195474	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.275	188	416373	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	408164	20.000	ng/u1	0.00
88) Perylene-d12	23.874	264	432515	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	4110	2.310	ng/uL	0.00
4) Pyridine-d5	3.628	84	29529	5.868	ng/u1	0.01
7) Phenol-d5	7.022	99	30153	5.255	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	121917	35.436	ng/u1	0.00
11) 2-Chlorophenol-d4	7.375	132	76562	18.219	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	56204	12.407	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	74103	33.435	ng/u1	0.00
24) 2-Nitrophenol-d4	9.757	143	52171	21.096	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.293	165	83074	19.705	ng/u1	0.00
31) 4-Chloroaniline-d4	10.816	131	180006	28.031	ng/u1	0.00
46) Dimethylphthalate-d6	13.939	166	528414	37.165	ng/u1	0.00
49) Acenaphthylene-d8	14.216	160	562619	35.914	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	10944	3.861	ng/u1	0.00
60) Fluorene-d10	15.516	176	450217	38.609	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	44455	16.797	ng/u1	0.00
73) Anthracene-d10	17.375	188	734057	40.262	ng/u1	0.00
81) Pyrene-d10	19.680	212	834943	39.473	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.721	264	828292	38.096	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111722\
Data File : BN022773.D
Acq On : 19 Nov 2022 01:40
Operator : CG/JU
Sample : N5689-18
Misc :
ALS Vial : 63 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BGCG0

Quant Time: Nov 19 02:18:04 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN111522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 18 05:17:17 2022
Response via : Initial Calibration

