

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111821\
 Data File : BN017513.D
 Acq On : 18 Nov 2021 18:56
 Operator : CG/JU
 Sample : M4740-11
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB-01-B-111721

Quant Time: Nov 19 03:36:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 16 14:06:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.874	152	27475	0.400	ng	0.00
7) Naphthalene-d8	10.661	136	124789	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	65644	0.400	ng	-0.01
19) Phenanthrene-d10	17.237	188	118692	0.400	ng	0.00
29) Chrysene-d12	21.419	240	88883	0.400	ng	# 0.00
35) Perylene-d12	23.793	264	66545	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.467	112	8337	0.127	ng	0.02
5) Phenol-d6	7.035	99	5597	0.080	ng	0.00
8) Nitrobenzene-d5	9.013	82	23991	0.310	ng	0.00
11) 2-Methylnaphthalene-d10	12.248	152	51963	0.253	ng	0.00
14) 2,4,6-Tribromophenol	15.982	330	4529	0.272	ng	0.00
15) 2-Fluorobiphenyl	13.115	172	68979	0.279	ng	-0.01
27) Fluoranthene-d10	19.262	212	116654	0.329	ng	0.00
31) Terphenyl-d14	19.863	244	91620	0.420	ng	0.00
Target Compounds						
9) Naphthalene	10.712	128	42005	0.134	ng	Qvalue 100
25) Phenanthrene	17.274	178	15123	0.046	ng	99
28) Fluoranthene	19.297	202	8024	0.023	ng	98
34) Bis(2-ethylhexyl)phtha...	21.345	149	29664	0.206	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111821\
Data File : BN017513.D
Acq On : 18 Nov 2021 18:56
Operator : CG/JU
Sample : M4740-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB-01-B-111721

Quant Time: Nov 19 03:36:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 16 14:06:59 2021
Response via : Initial Calibration

