

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111821\
 Data File : BN017503.D
 Acq On : 18 Nov 2021 09:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 18 10:44:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 16 14:06:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.367	8.3	105	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	108	0.00
4 S	2-Fluorophenol	0.400	0.412	-3.0	99	0.00
5 S	Phenol-d6	0.400	0.410	-2.5	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.422	-5.5	109	0.00
9	Naphthalene	0.400	0.404	-1.0	98	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	93	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.427	-6.7	95	-0.01
16	Acenaphthylene	0.400	0.367	8.3	93	0.00
17	Acenaphthene	0.400	0.395	1.3	93	-0.01
18	Fluorene	0.400	0.390	2.5	92	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.341	14.8	98	0.00
21	4-Bromophenyl-phenylether	0.400	0.380	5.0	94	-0.01
22	Hexachlorobenzene	0.400	0.395	1.3	95	-0.01
23	Atrazine	0.400	0.395	1.3	102	0.00
24	Pentachlorophenol	0.400	0.328	18.0	90	-0.01
25	Phenanthrene	0.400	0.395	1.3	95	0.00
26	Anthracene	0.400	0.376	6.0	96	-0.01
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	96	0.00
28	Fluoranthene	0.400	0.405	-1.3	97	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.408	-2.0	97	0.00
31 S	Terphenyl-d14	0.400	0.410	-2.5	94	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	96	0.00
33	Chrysene	0.400	0.403	-0.8	97	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.381	4.8	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.335	16.3	95	-0.01
37	Benzo(b)fluoranthene	0.400	0.398	0.5	101	0.00
38	Benzo(k)fluoranthene	0.400	0.411	-2.7	104	0.00
39 C	Benzo(a)pyrene	0.400	0.380	5.0	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.323	19.3	93	-0.01
41	Benzo(g,h,i)perylene	0.400	0.351	12.3	95	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4