

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111918\
 Data File : BN003613.D
 Acq On : 20 Nov 2018 02:16
 Operator : MJ/SJ
 Sample : J5928-20
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A40S2

Manual Integrations
 APPROVED

Sohil
 11/20/2018 5:13:43 PM

Quant Time: Nov 20 03:32:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 20 03:27:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	6202	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	25981	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	18473	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	34160m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	34637m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	32610m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.23	152	15092	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	38008m	0.38	ng/ul	0.00
Target Compounds						
8) Acenaphthene	14.55	153	7253	0.100	ng/ul#	Ovalue 23

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111918\
Data File : BN003613.D
Acq On : 20 Nov 2018 02:16
Operator : MJ/SJ
Sample : J5928-20
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A40S2

Quant Time: Nov 20 03:32:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 20 03:27:39 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
11/20/2018 5:13:43 PM

