

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111918\
 Data File : BN003594.D
 Acq On : 19 Nov 2018 15:10
 Operator : MJ/SJ
 Sample : SSTD3.266
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.266

Quant Time: Nov 19 15:43:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 19 13:56:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	5386	0.40	ng/ul	0.00
2) Naphthalene-d8	10.64	136	21160	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.48	164	11732	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	26515	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	32057	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	24214	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.86	266	20858	3.928	ng/ul	Ovalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111918\
Data File : BN003594.D
Acq On : 19 Nov 2018 15:10
Operator : MJ/SJ
Sample : SSTD3.266
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.266

Quant Time: Nov 19 15:43:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 19 13:56:58 2018
Response via : Initial Calibration

