

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111924\
 Data File : BN035178.D
 Acq On : 20 Nov 2024 00:52
 Operator : RC/JU
 Sample : PB165074BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS074

Quant Time: Nov 20 01:19:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 18 00:44:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.545	152	4141	0.400	ng/ul	0.00
4) Naphthalene-d8	10.306	136	10178	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.185	164	7249	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.937	188	15802	0.400	ng/ul	0.00
17) Chrysene-d12	21.148	240	11037	0.400	ng/ul	0.00
23) Perylene-d12	23.323	264	10962	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.136	96	2263	0.598	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.906	152	5591	0.375	ng/ul	0.00
18) Fluoranthene-d10	18.977	212	13212	0.434	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.170	88	6004	1.461	ng/ul#	94
5) Naphthalene	10.355	128	8292	0.323	ng/ul	98
7) 2-Methylnaphthalene	11.978	142	5944	0.334	ng/ul	99
8) 1-Methylnaphthalene	12.203	142	6053	0.337	ng/ul	99
10) Acenaphthylene	13.898	152	9182	0.319	ng/ul	99
11) Acenaphthene	14.245	153	7104	0.323	ng/ul	99
12) Fluorene	15.239	166	8451	0.327	ng/ul	99
14) Pentachlorophenol	16.595	266	2157	0.586	ng/ul	98
15) Phenanthrene	16.979	178	13522	0.324	ng/ul	99
16) Anthracene	17.072	178	12344	0.318	ng/ul	100
19) Fluoranthene	19.004	202	15124	0.394	ng/ul	100
20) Pyrene	19.372	202	15257	0.382	ng/ul	100
21) Benzo(a)anthracene	21.134	228	12974	0.339	ng/ul	100
22) Chrysene	21.183	228	13421	0.348	ng/ul	99
24) Benzo(b)fluoranthene	22.673	252	12586	0.347	ng/ul	99
25) Benzo(k)fluoranthene	22.717	252	13684	0.366	ng/ul	99
26) Benzo(a)pyrene	23.226	252	11472	0.346	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.477	276	15469	0.346	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.490	278	12095	0.345	ng/ul	98
29) Benzo(g,h,i)perylene	26.131	276	12043	0.342	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111924\
Data File : BN035178.D
Acq On : 20 Nov 2024 00:52
Operator : RC/JU
Sample : PB165074BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS074

Quant Time: Nov 20 01:19:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 18 00:44:37 2024
Response via : Initial Calibration

